

Chemical Analysis Report

TLH-ROC-2017-10-10-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SMP 2017 Marine Kit B**
Request ID: **RQ-2017-07-17-37**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 26-OCT-2017 12:13

A handwritten signature in black ink, appearing to read "Liang Lin", on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Direct Runoff to Gulf

Collection Date/Time: 10/10/2017 13:40

Field ID: G1TLHR0013-2017-05

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1935516	EPA 180.1 Rev. 2.0	Turbidity	5.6	A	NTU	P333191	
	SM 2320 B-97	Total Alkalinity	140		mg CaCO3/L	P333538	
	SM 2540 D-97	TSS	11	A	mg/L	P333286	
	SM 4500 F-C-97	Fluoride	0.75		mg F/L	P333541	
1935517	EPA 350.1 Rev. 2.0	Ammonia-N	0.12		mg N/L	P333390	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P333644	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.018		mg N/L	P333420	
	EPA 365.1 Rev. 2.0	Total-P	0.081		mg P/L	P333271	
	SM 5310 B-00	Organic Carbon	20		mg C/L	P333821	
1935518	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.038		mg P/L	P333218	

Ref. Method and Comment:

EPA 350.1 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P333191

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P333390

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P333644

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P333420

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P333271

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P333218

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P333538

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 D-97
Batch ID: P333286

Component	Result	Code	Units
TSS	3	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P333541

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P333821

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P333390

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	98.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P333644

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P333420

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P333271

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P333218

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P333218

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P333538

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P333541

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.9		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P333821

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.2		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P333644

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1936768	Kjeldahl Nitrogen	103	98.7	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P333420

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1934451	NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P333271

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1935701	Total-P	97.1	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P333218

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1935840	O-Phosphate-P	97.5	98.3	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P333541

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P333541

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1936092	Fluoride	98.3	96.9	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P333821

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1935517	Organic Carbon	90.9	90.1	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P333191

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1935516	Turbidity	0.178	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P333390

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1934985	Ammonia-N	1.26	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P333644

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1936768	Kjeldahl Nitrogen	3.60	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P333420

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1934451	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P333271

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1935701	Total-P	4.82	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P333218

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P333218

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1935840	O-Phosphate-P	0.729	Spike	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P333538

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1936092	Total Alkalinity	0.292	Sample	P	0 - 4.8

Reference Method: SM 2540 D-97
Batch ID: P333286

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1935516	TSS	3.77	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P333541

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1936092	Fluoride	0.986	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P333821

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1935517	Organic Carbon	0.568	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A79495
Included Lab Sample IDs: 1935516

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A79502
Included Lab Sample IDs: 1935518

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.5	98.4	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A79555

Included Lab Sample IDs: 1935517

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	101	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A79568

Included Lab Sample IDs: 1935517

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A79578

Included Lab Sample IDs: 1935517

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	102	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A79625

Included Lab Sample IDs: 1935516

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	103	100	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A79625

Included Lab Sample IDs: 1935516

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	99.5	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A79713

Included Lab Sample IDs: 1935517

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.2	90.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A79737

Included Lab Sample IDs: 1935517

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	104	102	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements. Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					0.178	
EPA 350.1 Rev. 2.0	Ammonia-N	98.0					1.26
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	103	98.7			3.60
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	104	102				0.0
EPA 365.1 Rev. 2.0	Total-P	104	97.1	102			4.82
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102	97.5	98.3			0.729
SM 2320 B-97	Total Alkalinity	102				0.292	
SM 2540 D-97	TSS					3.77	
SM 4500 F-C-97	Fluoride	98.9	98.3	96.9			0.986
SM 5310 B-00	Organic Carbon	95.2	90.9	90.1			0.568

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1935516
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1935517
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1935517
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1935517
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1935517
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1935518
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1935516
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1935516
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1935516
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1935517

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	10/10/2017			10/11/2017 16:38	DeAsia Armster	1935516
EPA 350.1 Rev. 2.0	10/10/2017			10/13/2017	Sofia Elnemr	1935517
EPA 351.2 Rev. 2.0	10/10/2017	10/19/2017	Adrian Shelby	10/20/2017	Joseph Suarez	1935517
EPA 353.2 Rev. 2.0	10/10/2017			10/16/2017	Beverly Y. Sanders	1935517
EPA 365.1 Rev. 2.0	10/10/2017	10/12/2017	Sima Feizi	10/16/2017	Lipi Saha	1935517
EPA 365.1 Rev. 2.0 dissolved	10/10/2017			10/11/2017 15:30	Megan Treadwell	1935518
SM 2320 B-97	10/10/2017			10/18/2017	Dale L. Simmons	1935516
SM 2540 D-97	10/10/2017			10/11/2017	Yijie Li	1935516
SM 4500 F-C-97	10/10/2017			10/17/2017	Dale L. Simmons	1935516
SM 5310 B-00	10/10/2017			10/21/2017	Ping Hua	1935517