

Chemical Analysis Report

TLH-ROC-2017-11-08-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **SMP Marine Kit B**
Request ID: **RQ-2017-10-23-34**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 04-DEC-2017 13:13

A handwritten signature in black ink, appearing to read "Colin Wright", is written over a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: FENHOLLWY RIVER @ CR 356 BOAT RAMP

Collection Date/Time: 11/08/2017 11:20

Field ID: G1TLHR0080-2017-01

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1944816	EPA 180.1 Rev. 2.0	Turbidity	7.6		NTU	P335010	
	SM 2320 B-97	Total Alkalinity	228		mg CaCO3/L	P335377	
	SM 2540 D-97	TSS	6	I	mg/L	P335585	
	SM 4500 F-C-97	Fluoride	0.19		mg F/L	P335381	
1944818	EPA 350.1 Rev. 2.0	Ammonia-N	0.33		mg N/L	P335891	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.2		mg N/L	P335874	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.77		mg N/L	P335936	
	EPA 365.1 Rev. 2.0	Total-P	0.54		mg P/L	P335296	
	SM 5310 B-00	Organic Carbon	31		mg C/L	P335512	
1944820	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.044		mg P/L	P335085	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: GULF OF MEXICO (ECONFINA)

Collection Date/Time: 11/08/2017 13:50

Field ID: G1TLHR0043-2017-04

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1944817	EPA 180.1 Rev. 2.0	Turbidity	6.2	A	NTU	P335010	
	SM 2320 B-97	Total Alkalinity	140		mg CaCO3/L	P335402	
	SM 2540 D-97	TSS	21		mg/L	P335588	
	SM 4500 F-C-97	Fluoride	0.94		mg F/L	P335854	
1944819	EPA 350.1 Rev. 2.0	Ammonia-N	0.007		mg N/L	P335760	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.72		mg N/L	P335807	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P335725	
	EPA 365.1 Rev. 2.0	Total-P	0.042		mg P/L	P335355	
	SM 5310 B-00	Organic Carbon	7.2		mg C/L	P335304	
1944821	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P335079	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P335010

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P335760

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P335891

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P335807

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P335874

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P335725

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P335936

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P335296

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P335355

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P335079

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P335085

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2320 B-97
Batch ID: P335377

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P335402

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 D-97
Batch ID: P335585

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P335588

Component	Result	Code	Units
TSS	3	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P335381

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 4500 F-C-97
Batch ID: P335854

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P335304

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: SM 5310 B-00
Batch ID: P335512

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P335760

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	104		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P335891

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P335891

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	105		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P335807

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P335874

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P335725

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P335936

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P335296

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P335355

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P335079

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P335085

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	104		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 2320 B-97
Batch ID: P335377

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 2320 B-97
Batch ID: P335402

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P335381

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.0		P	92 - 103

Reference Method: SM 4500 F-C-97
Batch ID: P335854

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.0		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P335304

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.0		P	85 - 115

Reference Method: SM 5310 B-00
Batch ID: P335512

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P335760

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1944525	Ammonia-N	100	99.7	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P335891

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1943022	Ammonia-N	101	96.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P335807

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P335807

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1945044	Kjeldahl Nitrogen	102	94.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P335874

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1945776	Kjeldahl Nitrogen	99.2	104	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P335725

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1944819	NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P335936

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1946589	NO2NO3-N	105	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P335296

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1944859	Total-P	99.5	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P335355

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1945044	Total-P	95.0	94.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P335079

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1944592	O-Phosphate-P	99.5	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P335085

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1945041	O-Phosphate-P	105	99.7	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P335381

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1944887	Fluoride	100	100	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P335854

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1946323	Fluoride	97.0	97.0	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P335304

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1943768	Organic Carbon	93.2	92.6	P/P	85 - 115

Reference Method: SM 5310 B-00
Batch ID: P335512

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1944219	Organic Carbon	95.6	96.6	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P335010

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944817	Turbidity	1.13	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P335760

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944525	Ammonia-N	0.389	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P335891

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1943022	Ammonia-N	2.92	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P335807

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1945044	Kjeldahl Nitrogen	5.60	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P335874

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1945776	Kjeldahl Nitrogen	2.22	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P335725

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944819	NO2NO3-N	0.491	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P335936

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1946589	NO2NO3-N	0.475	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P335296

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944859	Total-P	0.614	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P335355

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1945044	Total-P	0.423	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P335079

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944592	O-Phosphate-P	1.00	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P335085

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1945041	O-Phosphate-P	4.63	Spike	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P335377

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944887	Total Alkalinity	0.331	Sample	P	0 - 4.8

Reference Method: SM 2320 B-97
Batch ID: P335402

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944313	Total Alkalinity	0.316	Sample	P	0 - 4.8

Quality Assurance Report Precision

Reference Method: SM 2540 D-97
Batch ID: P335588

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944626	TSS	7.59	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P335381

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944887	Fluoride	0.0	Spike	P	0 - 3.0

Reference Method: SM 4500 F-C-97
Batch ID: P335854

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1946323	Fluoride	0.0	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P335304

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1943768	Organic Carbon	0.326	Spike	P	0 - 20

Reference Method: SM 5310 B-00
Batch ID: P335512

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1944219	Organic Carbon	0.549	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A80151
Included Lab Sample IDs: 1944816, 1944817

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A80191
Included Lab Sample IDs: 1944821

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.4	97.5	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A80192

Included Lab Sample IDs: 1944820

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	103	108	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A80269

Included Lab Sample IDs: 1944819

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.0	97.4	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A80294

Included Lab Sample IDs: 1944816

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	104	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A80294

Included Lab Sample IDs: 1944816

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	102	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A80300

Included Lab Sample IDs: 1944817

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.9	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A80313

Included Lab Sample IDs: 1944818

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	102	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A80341

Included Lab Sample IDs: 1944818

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A80354

Included Lab Sample IDs: 1944819

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	105	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A80432

Included Lab Sample IDs: 1944819

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A80450

Included Lab Sample IDs: 1944819

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	103	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A80493

Included Lab Sample IDs: 1944817

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A80513

Included Lab Sample IDs: 1944818

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	102	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A80537

Included Lab Sample IDs: 1944818

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A80551

Included Lab Sample IDs: 1944819

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	103	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A80585

Included Lab Sample IDs: 1944818

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.3	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements. Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity					1.13	
EPA 350.1 Rev. 2.0	Ammonia-N	104	100	99.7			0.389
	Ammonia-N	105	101	96.8			2.92
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105	102	94.2			5.60
	Kjeldahl Nitrogen	104	99.2	104			2.22
EPA 353.2 Rev. 2.0	NO2NO3-N	104	102	102			0.491
	NO2NO3-N	102	105	106			0.475
EPA 365.1 Rev. 2.0	Total-P	104	99.5	100			0.614
	Total-P	97.0	95.0	94.6			0.423
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.5	99.5	100			1.00
	O-Phosphate-P	104	105	99.7			4.63
SM 2320 B-97	Total Alkalinity	103				0.331	
	Total Alkalinity	103				0.316	
SM 2540 D-97	TSS					7.59	
SM 4500 F-C-97	Fluoride	99.0	100	100			0.0
	Fluoride	98.0	97.0	97.0			0.0
SM 5310 B-00	Organic Carbon	99.0	93.2	92.6			0.326
	Organic Carbon	98.8	95.6	96.6			0.549

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1944816, 1944817
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1944818, 1944819
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1944818, 1944819
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1944818, 1944819
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1944818, 1944819
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1944820, 1944821
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1944816, 1944817
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1944816, 1944817
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1944816, 1944817
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1944818, 1944819

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/08/2017			11/08/2017 17:35	DeAsia Armster	1944816, 1944817
EPA 350.1 Rev. 2.0	11/08/2017			11/21/2017	Sofia Elnemr	1944819
	11/08/2017			11/22/2017	Sofia Elnemr	1944818
EPA 351.2 Rev. 2.0	11/08/2017	11/22/2017	Adrian Shelby	11/27/2017	Joseph Suarez	1944819
	11/08/2017	11/27/2017	Adrian Shelby	11/28/2017	Joseph Suarez	1944818
EPA 353.2 Rev. 2.0	11/08/2017			11/21/2017	Beverly Y. Sanders	1944819
	11/08/2017			11/28/2017	Beverly Y. Sanders	1944818
EPA 365.1 Rev. 2.0	11/08/2017	11/15/2017	Sima Feizi	11/16/2017	Lipi Saha	1944818
	11/08/2017	11/15/2017	Sima Feizi	11/17/2017	Lipi Saha	1944819
EPA 365.1 Rev. 2.0 dissolved	11/08/2017			11/09/2017 11:50	Megan Treadwell	1944821
	11/08/2017			11/09/2017 15:45	Megan Treadwell	1944820
SM 2320 B-97	11/08/2017			11/15/2017	Dale L. Simmons	1944816
	11/08/2017			11/16/2017	Dale L. Simmons	1944817
SM 2540 D-97	11/08/2017			11/09/2017	Yijie Li	1944816, 1944817
SM 4500 F-C-97	11/08/2017			11/15/2017	Dale L. Simmons	1944816
	11/08/2017			11/23/2017	Dale L. Simmons	1944817
SM 5310 B-00	11/08/2017			11/15/2017	Ping Hua	1944819
	11/08/2017			11/17/2017	Ping Hua	1944818