

Chemical Analysis Report

TLH-ROC-2017-11-30-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **SMP Freshwater Kit A**
Request ID: **RQ-2017-10-30-54**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 18-DEC-2017 11:26

A handwritten signature in black ink, appearing to read "Colin Wright", is positioned on a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Little Aucilla River @ US 90

Collection Date/Time: 11/30/2017 11:15

Field ID: G1TLHR0029-2017-06

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1949892	EPA 180.1 Rev. 2.0	Turbidity	0.75		NTU	P336257	
	EPA 300.0 Rev. 2.1	Chloride	7.0		mg Cl/L	P336481	
	SM 2120 B	Color (true)	340	A	PCU	P336158	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P336617	
	SM 2540 C-97	TDS	98	A	mg/L	P336444	
	SM 2540 D-97	TSS	2	U	mg/L	P336304	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P336620	
1949893	EPA 350.1 Rev. 2.0	Ammonia-N	0.005		mg N/L	P336794	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.3		mg N/L	P336569	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P336385	
	EPA 365.1 Rev. 2.0	Total-P	0.017		mg P/L	P336530	
	SM 5310 B-00	Organic Carbon	42		mg C/L	P336445	
1949894	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P336217	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P336257

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P336481

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P336794

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P336569

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P336385

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336530

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336217

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P336158

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P336617

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P336444

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P336304

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P336620

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P336445

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336481

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P336481

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.0		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P336794

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	104		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P336569

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.8		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P336385

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P336530

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P336217

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.4		P	90 - 110

Reference Method: SM 2320 B-97
 Batch ID: P336617

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
 Batch ID: P336620

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.2		P	92 - 103

Reference Method: SM 5310 B-00
 Batch ID: P336445

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336481

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949769	Chloride	95.2		P	90 - 110
1949897	Chloride	95.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336794

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949773	Ammonia-N	102	98.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336569

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949918	Kjeldahl Nitrogen	96.5	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336385

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949724	NO2NO3-N	106	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336530

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949909	Total-P	97.8	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336217

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949964	O-Phosphate-P	98.8	98.9	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P336620

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949722	Fluoride	98.2	97.6	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P336445

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949724	Organic Carbon	95.1	95.0	P/P	85 - 115

Quality Assurance Report Precision

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P336257

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949887	Turbidity	1.49	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336481

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949769	Chloride	0.468	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336794

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949773	Ammonia-N	2.26	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336569

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949918	Kjeldahl Nitrogen	4.05	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336385

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949724	NO2NO3-N	1.91	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336530

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949909	Total-P	4.27	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336217

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949964	O-Phosphate-P	0.0881	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P336158

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949892	Color (true)	1.38	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2320 B-97
Batch ID: P336617

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949722	Total Alkalinity	1.19	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P336444

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949892	TDS	1.02	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P336620

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949722	Fluoride	0.474	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P336445

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949724	Organic Carbon	0.0825	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A80620
Included Lab Sample IDs: 1949892

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	103	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A80641
Included Lab Sample IDs: 1949894

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	101	100	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A80658
Included Lab Sample IDs: 1949892

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	100	P/P	95 - 105

Quality Assurance Report Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A80702

Included Lab Sample IDs: 1949893

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	102	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A80723

Included Lab Sample IDs: 1949893

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.6	97.2	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A80735

Included Lab Sample IDs: 1949892

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	104	102	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A80770

Included Lab Sample IDs: 1949892

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	103	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A80770

Included Lab Sample IDs: 1949892

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	103	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A80797

Included Lab Sample IDs: 1949893

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.0	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A80809

Included Lab Sample IDs: 1949893

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.8	103	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A80825

Included Lab Sample IDs: 1949893

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	104	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	SMP
EPA 180.1 Rev. 2.0	Turbidity					1.49
EPA 300.0 Rev. 2.1	Chloride	97.0	95.2	95.8		0.468
EPA 350.1 Rev. 2.0	Ammonia-N	104	102	98.2		
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.8	96.5	102		2.26
EPA 353.2 Rev. 2.0	NO2NO3-N	100	106	104		4.05
EPA 365.1 Rev. 2.0	Total-P	103	97.8	102		1.91
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.4	98.8	98.9		4.27
SM 2120 B	Color (true)					0.0881
SM 2320 B-97	Total Alkalinity	103			1.38	
SM 2540 C-97	TDS				1.19	
SM 4500 F-C-97	Fluoride	98.2	98.2	97.6	1.02	
SM 5310 B-00	Organic Carbon	99.4	95.1	95.0		0.474
						0.0825

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1949892
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1949892
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1949893
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1949893
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1949893
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1949893
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1949894
SM 2120 B / E31780	True Color measured at 450 nm	1949892
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1949892
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1949892
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1949892
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1949892
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1949893

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/30/2017			12/01/2017 15:40	Tanya Welborn	1949892
EPA 300.0 Rev. 2.1	11/30/2017			12/05/2017	Julia Storbeck	1949892
EPA 350.1 Rev. 2.0	11/30/2017			12/08/2017	Sofia Elnemr	1949893
EPA 351.2 Rev. 2.0	11/30/2017	12/07/2017	Adrian Shelby	12/08/2017	Joseph Suarez	1949893
EPA 353.2 Rev. 2.0	11/30/2017			12/05/2017	Lauren Bottorf	1949893
EPA 365.1 Rev. 2.0	11/30/2017	12/06/2017	Sima Feizi	12/11/2017	Beverly Y. Sanders	1949893
EPA 365.1 Rev. 2.0 dissolved	11/30/2017			12/01/2017 14:51	Megan Treadwell	1949894
SM 2120 B	11/30/2017			12/01/2017 09:48	Tanya Welborn	1949892
SM 2320 B-97	11/30/2017			12/07/2017	Dale L. Simmons	1949892
SM 2540 C-97	11/30/2017			12/01/2017	DeAsia Armster	1949892
SM 2540 D-97	11/30/2017			12/01/2017	DeAsia Armster	1949892
SM 4500 F-C-97	11/30/2017			12/07/2017	Dale L. Simmons	1949892
SM 5310 B-00	11/30/2017			12/05/2017	Ping Hua	1949893