

Chemical Analysis Report

TLH-ROC-2017-11-30-03

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **kit C_pesticides**
Request ID: **RQ-2017-11-06-40**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 22-DEC-2017 13:13

A handwritten signature in black ink, appearing to read "Colin Wright", is written over a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals, Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1949901	EPA 8321B	Glyphosate**	0.10	U	ug/L	P335934	
		Pyraclostrobin**	0.0020	U	ug/L	P335933	
		Sucralose**	0.022	I	ug/L	P336107	
	USGS O-2060-01	2,4-D	0.033		ug/L	P336034	
		Diuron	0.064		ug/L	P336034	
		Fenuron	0.0080	U	ug/L	P336034	
		Imidacloprid	8.0E-04	U	ug/L	P336034	MS
		Bentazon	8.0E-04	U	ug/L	P336034	
		Linuron	0.0040	U	ug/L	P336034	

Field ID: G1TLHR0048-2017-05

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1949901	USGS O-2060-01	Acetaminophen**	0.0080	U	ug/L	P336034	
		MCP	0.0020	U	ug/L	P336034	
		Carbamazepine**	4.0E-04	U	ug/L	P336034	
		Triclopyr**	0.0040	U	ug/L	P336034	
		Primidone**	0.0040	U	ug/L	P336034	
		Fluridone**	4.0E-04	U	ug/L	P336034	
1949902	EPA 8270D	Aldrin	4.0	UJ	ng/L	P336154	RPD
		Alpha-BHC	2.0	U	ng/L	P336154	
		Beta-BHC	4.0	UJ	ng/L	P336154	LCS
		Delta-BHC	4.0	U	ng/L	P336154	
		Gamma-BHC	4.0	U	ng/L	P336154	
		Carbophenothion	4.0	U	ng/L	P336154	
		Chlorothalonil	2.0	U	ng/L	P336154	MS
		Cypermethrin	20	U	ng/L	P336154	
		DDD-p,p'	3.0	U	ng/L	P336154	
		DDE-p,p'	3.0	U	ng/L	P336154	
		DDT-p,p'	3.0	UJ	ng/L	P336154	LCS
		Dieldrin	4.0	U	ng/L	P336154	
		Endosulfan I	4.0	U	ng/L	P336154	
		Endosulfan II	4.0	U	ng/L	P336154	
		Endosulfan Sulfate	2.0	U	ng/L	P336154	
		Endrin	2.0	U	ng/L	P336154	
		Endrin Aldehyde	2.0	U	ng/L	P336154	
		Heptachlor	1.5	U	ng/L	P336154	
		Heptachlor Epoxide	2.0	U	ng/L	P336154	
		Methoxychlor	4.0	U	ng/L	P336154	
		Mirex	2.0	U	ng/L	P336154	
		Permethrin	10	U	ng/L	P336154	
		Trifluralin	1.0	U	ng/L	P336154	
		Alpha-Chlordane	1.0	U	ng/L	P336154	
		Gamma-Chlordane	1.0	U	ng/L	P336154	
		Endrin Ketone	2.0	UJ	ng/L	P336154	LCS
		Dicofol	30	U	ng/L	P336154	
		Chlordane**	2.5	U	ng/L	P336154	
		Toxaphene**	12	U	ng/L	P336154	
1949903	EPA 8270D	Alachlor	0.25	U	ng/L	P336157	
		Ametryn	0.32	U	ng/L	P336157	
		Atrazine	4.6		ng/L	P336157	
		Atrazine Desethyl	4.2	I	ng/L	P336157	
		Azinphos Methyl	2.2	U	ng/L	P336157	
		Bromacil	0.50	U	ng/L	P336157	
		Butylate	0.40	U	ng/L	P336157	
		Chlorpyrifos Ethyl	0.25	U	ng/L	P336157	
		Chlorpyrifos Methyl	0.10	U	ng/L	P336157	
		Demeton	2.0	UJ	ng/L	P336157	RPD

Field ID: G1TLHR0048-2017-05

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1949903	EPA 8270D	Diazinon	0.12	U	ng/L	P336157	
		Disulfoton	0.50	U	ng/L	P336157	
		EPTC	1.2	U	ng/L	P336157	
		Ethion	0.15	U	ng/L	P336157	MS, RPD
		Ethoprop	0.10	U	ng/L	P336157	
		Fenamiphos	0.25	U	ng/L	P336157	
		Fipronil	0.25	U	ng/L	P336157	
		Fipronil Sulfide	0.15	U	ng/L	P336157	
		Fipronil Sulfone	0.15	U	ng/L	P336157	
		Hexazinone	0.50	U	ng/L	P336157	
		Malathion	0.35	U	ng/L	P336157	
		Metribuzin	0.20	U	ng/L	P336157	MS
		Mevinphos	0.50	U	ng/L	P336157	
		Molinate	0.30	U	ng/L	P336157	
		Norflurazon	0.50	U	ng/L	P336157	
		Pendimethalin	1.0	U	ng/L	P336157	
		Phorate	0.25	U	ng/L	P336157	
		Prometon	0.90	U	ng/L	P336157	
		Prometryn	0.20	U	ng/L	P336157	
		Simazine	0.50	U	ng/L	P336157	
		Terbufos	0.10	U	ng/L	P336157	
		Terbutylazine	0.42	U	ng/L	P336157	
		Fonofos	0.20	U	ng/L	P336157	
		Metaxyl	0.25	U	ng/L	P336157	
		Metolachlor	1.0		ng/L	P336157	
		Acetochlor	0.25	U	ng/L	P336157	
		Cyanazine	0.50	UJ	ng/L	P336157	CCV
		Parathion Ethyl	0.25	U	ng/L	P336157	
		Parathion Methyl	0.25	U	ng/L	P336157	

Ref. Method and Comment:

EPA 8270D: Insufficient sample available to perform duplicate matrix spikes. Refer to the narrative for an explanation of QC Codes.

EPA 8270D: Refer to the narrative for an explanation of QC Codes.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P336257

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4

Batch ID: P336291

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P336291

Component	Result	Code	Units
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P336291

Component	Result	Code	Units
Arsenic	0.050	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.20	U	ug/L
Lead	0.050	U	ug/L
Nickel	0.20	U	ug/L
Silver	0.010	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336481

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336484

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336794

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336569

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336385

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336530

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336217

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8270D
Batch ID: P336154

Component	Result	Code	Units
Aldrin	4.0	U	ng/L
Alpha-BHC	2.0	U	ng/L
Alpha-Chlordane	1.0	U	ng/L
Beta-BHC	4.0	U	ng/L
Carbophenothion	4.0	U	ng/L
Chlorothalonil	2.0	U	ng/L
Cypermethrin	20	U	ng/L
DDD-p,p'	3.0	U	ng/L
DDE-p,p'	3.0	U	ng/L
DDT-p,p'	3.0	U	ng/L
Delta-BHC	4.0	U	ng/L
Dicofol	30	U	ng/L
Dieldrin	4.0	U	ng/L
Endosulfan I	4.0	U	ng/L
Endosulfan II	4.0	U	ng/L
Endosulfan Sulfate	2.0	U	ng/L
Endrin	2.0	U	ng/L
Endrin Aldehyde	2.0	U	ng/L
Endrin Ketone	2.0	U	ng/L
Gamma-BHC	4.0	U	ng/L
Gamma-Chlordane	1.0	U	ng/L
Heptachlor	1.5	U	ng/L
Heptachlor Epoxide	2.0	U	ng/L
Methoxychlor	4.0	U	ng/L
Mirex	2.0	U	ng/L
Permethrin	10	U	ng/L
Trifluralin	1.0	U	ng/L

Reference Method: EPA 8270D
Batch ID: P336157

Component	Result	Code	Units
Acetochlor	0.25	U	ng/L
Alachlor	0.25	U	ng/L
Ametryn	0.32	U	ng/L
Atrazine	0.25	U	ng/L
Atrazine Desethyl	1.5	U	ng/L
Azinphos Methyl	2.2	U	ng/L
Bromacil	0.50	U	ng/L
Butylate	0.40	U	ng/L
Chlorpyrifos Ethyl	0.25	U	ng/L
Chlorpyrifos Methyl	0.10	U	ng/L
Cyanazine	0.50	U	ng/L
Demeton	2.0	U	ng/L
Diazinon	0.12	U	ng/L
Disulfoton	0.50	U	ng/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8270D
Batch ID: P336157

Component	Result	Code	Units
EPTC	1.2	U	ng/L
Ethion	0.15	U	ng/L
Ethoprop	0.10	U	ng/L
Fenamiphos	0.25	U	ng/L
Fipronil	0.25	U	ng/L
Fipronil Sulfide	0.15	U	ng/L
Fipronil Sulfone	0.15	U	ng/L
Fonofos	0.20	U	ng/L
Hexazinone	0.50	U	ng/L
Malathion	0.35	U	ng/L
Metaxyl	0.25	U	ng/L
Metolachlor	0.25	U	ng/L
Metribuzin	0.20	U	ng/L
Mevinphos	0.50	U	ng/L
Molinate	0.30	U	ng/L
Norflurazon	0.50	U	ng/L
Parathion Ethyl	0.25	U	ng/L
Parathion Methyl	0.25	U	ng/L
Pendimethalin	1.0	U	ng/L
Phorate	0.25	U	ng/L
Prometon	0.90	U	ng/L
Prometryn	0.20	U	ng/L
Simazine	0.50	U	ng/L
Terbufos	0.10	U	ng/L
Terbuthylazine	0.42	U	ng/L

Reference Method: EPA 8276
Batch ID: P336154

Component	Result	Code	Units
Chlordane	2.5	U	ng/L
Toxaphene	12	U	ng/L

Reference Method: EPA 8321B
Batch ID: P335933

Component	Result	Code	Units
Pyraclostrobin	0.0020	U	ug/L

Reference Method: EPA 8321B
Batch ID: P335934

Component	Result	Code	Units
Glyphosate	0.10	U	ug/L

Reference Method: EPA 8321B
Batch ID: P336107

Component	Result	Code	Units
Sucralose	0.010	U	ug/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2120 B
 Batch ID: P336158

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
 Batch ID: P336617

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
 Batch ID: P336444

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
 Batch ID: P336305

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
 Batch ID: P336620

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
 Batch ID: P336635

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: USGS O-2060-01
 Batch ID: P336034

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
Acetaminophen	0.0080	U	ug/L
Bentazon	8.0E-04	U	ug/L
Carbamazepine	4.0E-04	U	ug/L
Diuron	8.0E-04	U	ug/L
Fenuron	0.0080	U	ug/L
Fluridone	4.0E-04	U	ug/L
Imidacloprid	8.0E-04	U	ug/L
Linuron	0.0040	U	ug/L
MCP	0.0020	U	ug/L
Primidone	0.0040	U	ug/L
Triclopyr	0.0040	U	ug/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P336291

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	101		P	85 - 115
Iron	98.9		P	85 - 115
Magnesium	105		P	85 - 115
Potassium	106		P	85 - 115
Sodium	102		P	85 - 115
Zinc	98.9		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P336291

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Arsenic	102		P	85 - 115
Cadmium	98.4		P	85 - 115
Chromium	97.0		P	85 - 115
Copper	99.1		P	85 - 115
Lead	97.6		P	85 - 115
Nickel	101		P	85 - 115
Silver	100		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336481

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336484

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.0		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336794

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	104		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336569

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.8		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336385

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P336530

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P336217

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.4		P	90 - 110

Reference Method: EPA 8270D
 Batch ID: P336154

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Aldrin	50.1	84.6	P/P	26 - 88.0
Alpha-BHC	110	108	P/P	63 - 114
Alpha-Chlordane	107	99.2	P/P	56 - 117
Beta-BHC	138	123	F/P	57 - 138
Carbophenothion	91.9	84.6	P/P	36 - 148
Chlorothalonil	111	91.0	P/P	26 - 173
Cypermethrin	125	107	P/P	42 - 159
DDD-p,p'	124	107	P/P	62 - 126
DDE-p,p'	110	104	P/P	53 - 115
DDT-p,p'	122	107	F/P	54 - 117
Delta-BHC	142	147	P/P	68 - 158
Dicofol	110	87.5	P/P	25 - 114
Dieldrin	112	103	P/P	59 - 126
Endosulfan I	115	109	P/P	62 - 138
Endosulfan II	145	120	P/P	61 - 150
Endosulfan Sulfate	125	95.7	P/P	63 - 132
Endrin	115	104	P/P	49 - 148
Endrin Aldehyde	126	98.9	P/P	50 - 144
Endrin Ketone	123	101	F/P	66 - 122
Gamma-BHC	113	124	P/P	60 - 137
Gamma-Chlordane	104	94.6	P/P	54 - 113
Heptachlor	95.4	88.2	P/P	43 - 104
Heptachlor Epoxide	110	100	P/P	64 - 118
Methoxychlor	129	127	P/P	63 - 129
Mirex	86.5	86.1	P/P	40 - 96.0
Permethrin	108	97.9	P/P	51 - 123
Trifluralin	97.1	94.5	P/P	35 - 187

Reference Method: EPA 8270D
 Batch ID: P336157

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Acetochlor	99.2	98.8	P/P	78 - 118
Alachlor	100	96.6	P/P	77 - 119
Ametryn	84.1	90.0	P/P	61 - 136
Atrazine	103	104	P/P	73 - 120
Atrazine Desethyl	82.5	83.2	P/P	16 - 122
Azinphos Methyl	113	109	P/P	69 - 186
Bromacil	100	99.2	P/P	40 - 125
Butylate	59.7	50.6	P/P	32 - 87.0

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8270D
Batch ID: P336157

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlorpyrifos Ethyl	93.7	91.6	P/P	53 - 110
Chlorpyrifos Methyl	100	99.0	P/P	31 - 119
Cyanazine	161	171	P/P	36 - 224
Demeton	45.3	78.5	P/P	39 - 122
Diazinon	95.9	91.8	P/P	75 - 119
Disulfoton	109	99.7	P/P	42 - 139
EPTC	35.9	42.9	P/P	33 - 96.0
Ethion	92.2	81.8	P/P	63 - 127
Ethoprop	102	96.8	P/P	65 - 107
Fenamiphos	109	103	P/P	64 - 137
Fipronil	102	98.4	P/P	60 - 126
Fipronil Sulfide	93.2	90.9	P/P	58 - 124
Fipronil Sulfone	89.5	85.6	P/P	57 - 126
Fonofos	93.8	96.9	P/P	65 - 111
Hexazinone	90.3	93.8	P/P	46 - 151
Malathion	99.9	98.8	P/P	68 - 132
Metaxyl	102	99.8	P/P	51 - 132
Metolachlor	98.2	98.0	P/P	78 - 119
Metribuzin	104	110	P/P	58 - 148
Mevinphos	56.9	69.7	P/P	34 - 128
Molinate	50.4	48.2	P/P	44 - 101
Norflurazon	93.3	91.9	P/P	63 - 129
Parathion Ethyl	94.2	96.2	P/P	80 - 109
Parathion Methyl	104	105	P/P	74 - 125
Pendimethalin	105	107	P/P	48 - 134
Phorate	87.8	84.0	P/P	52 - 119
Prometon	94.7	97.3	P/P	69 - 124
Prometryn	96.8	92.3	P/P	66 - 124
Simazine	97.9	104	P/P	61 - 134
Terbufos	97.1	95.8	P/P	58 - 115
Terbutylazine	98.6	99.4	P/P	77 - 113

Reference Method: EPA 8276
Batch ID: P336154

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chlordane	87.0	93.3	P/P	48 - 121
Toxaphene	113	111	P/P	43 - 135

Reference Method: EPA 8321B
Batch ID: P335933

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Pyraclostrobin	132		P	40 - 150

Reference Method: EPA 8321B
Batch ID: P335934

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Glyphosate	87.6		P	40 - 150

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P336107

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sucralose	67.4		P	40 - 150

Reference Method: SM 2320 B-97
Batch ID: P336617

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P336620

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.2		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P336635

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.1		P	85 - 115

Reference Method: USGS O-2060-01
Batch ID: P336034

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	97.6		P	40 - 150
Acetaminophen	115		P	30 - 150
Bentazon	75.8		P	30 - 150
Carbamazepine	109		P	40 - 150
Diuron	106		P	40 - 150
Fenuron	109		P	40 - 150
Fluridone	105		P	40 - 150
Imidacloprid	98.3		P	40 - 160
Linuron	100		P	40 - 150
MCPPP	92.1		P	40 - 150
Primidone	93.9		P	40 - 150
Triclopyr	89.1		P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P336291

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949568	Calcium	97.8		P	80 - 120
1949568	Iron	98.5		P	80 - 120
1949568	Magnesium	98.3		P	80 - 120
1949568	Sodium	101		P	80 - 120
1949568	Zinc	96.5		P	80 - 120
1949904	Calcium	96.5	95.5	P/P	80 - 120
1949904	Iron	94.8	95.9	P/P	80 - 120

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P336291

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949904	Magnesium	100	102	P/P	80 - 120
1949904	Potassium	104	101	P/P	80 - 120
1949904	Sodium	97.3		P	80 - 120
1949904	Zinc	98.2	97.6	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P336291

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949568	Arsenic	104		P	80 - 120
1949568	Cadmium	100		P	80 - 120
1949568	Chromium	94.4		P	80 - 120
1949568	Copper	96.4		P	80 - 120
1949568	Lead	98.5		P	80 - 120
1949568	Nickel	97.7		P	80 - 120
1949568	Silver	101		P	80 - 120
1949904	Arsenic	98.7	99.1	P/P	80 - 120
1949904	Cadmium	97.2	99.1	P/P	80 - 120
1949904	Chromium	93.9	94.3	P/P	80 - 120
1949904	Copper	96.6	97.2	P/P	80 - 120
1949904	Lead	97.2	97.6	P/P	80 - 120
1949904	Nickel	97.8	97.9	P/P	80 - 120
1949904	Silver	98.6	99.6	P/P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336481

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949769	Chloride	95.2		P	90 - 110
1949897	Chloride	95.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336484

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949769	Sulfate	97.2		P	90 - 110
1949897	Sulfate	97.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336794

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949773	Ammonia-N	102	98.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336569

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949918	Kjeldahl Nitrogen	96.5	102	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336385

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949724	NO2NO3-N	106	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336530

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949909	Total-P	97.8	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336217

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949964	O-Phosphate-P	98.8	98.9	P/P	90 - 110

Reference Method: EPA 8270D
Batch ID: P336154

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949273	Aldrin	57.8		P	22 - 97.0
1949273	Alpha-BHC	107		P	43 - 132
1949273	Alpha-Chlordane	91.5		P	36 - 126
1949273	Beta-BHC	88.8		P	41 - 150
1949273	Carbophenothion	126		P	13 - 196
1949273	Chlorothalonil	4.12		F	10 - 266
1949273	Cypermethrin	127		P	37 - 161
1949273	DDD-p,p'	78.4		P	44 - 131
1949273	DDE-p,p'	80.6		P	32 - 127
1949273	DDT-p,p'	74.9		P	45 - 116
1949273	Delta-BHC	110		P	50 - 201
1949273	Dicofol	83.6		P	10 - 100
1949273	Dieldrin	92.0		P	39 - 142
1949273	Endosulfan I	109		P	51 - 139
1949273	Endosulfan II	112		P	49 - 155
1949273	Endosulfan Sulfate	96.5		P	54 - 136
1949273	Endrin	93.5		P	38 - 141
1949273	Endrin Aldehyde	65.2		P	14 - 152
1949273	Endrin Ketone	100		P	60 - 128
1949273	Gamma-BHC	109		P	48 - 157
1949273	Gamma-Chlordane	90.4		P	35 - 123
1949273	Heptachlor	96.0		P	25 - 112
1949273	Heptachlor Epoxide	96.3		P	48 - 125
1949273	Methoxychlor	113		P	52 - 127
1949273	Mirex	94.3		P	25 - 108
1949273	Permethrin	108		P	46 - 133
1949273	Trifluralin	97.0		P	29 - 254

Reference Method: EPA 8270D
Batch ID: P336157

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1948578	Acetochlor	85.0	85.8	P/P	74 - 123
1948578	Alachlor	90.2	91.6	P/P	73 - 124

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8270D
Batch ID: P336157

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1948578	Ametryn	68.5	77.9	P/P	56 - 149
1948578	Atrazine	95.4	92.4	P/P	73 - 124
1948578	Atrazine Desethyl	98.4	83.9	P/P	12 - 103
1948578	Azinphos Methyl	104	79.5	P/P	60 - 202
1948578	Bromacil	76.0	86.7	P/P	42 - 133
1948578	Butylate	26.3	27.7	P/P	10 - 85.0
1948578	Chlorpyrifos Ethyl	74.1	78.0	P/P	53 - 113
1948578	Chlorpyrifos Methyl	80.4	79.5	P/P	40 - 115
1948578	Cyanazine	152	126	P/P	22 - 210
1948578	Demeton	80.1	70.6	P/P	17 - 149
1948578	Diazinon	80.7	80.1	P/P	72 - 124
1948578	Disulfoton	105	96.5	P/P	43 - 139
1948578	EPTC	22.6	26.3	P/P	10 - 86.0
1948578	Ethion	41.0	83.1	F/P	65 - 131
1948578	Ethoprop	93.3	90.6	P/P	59 - 110
1948578	Fenamiphos	107	98.9	P/P	57 - 138
1948578	Fipronil	76.9	85.3	P/P	40 - 130
1948578	Fipronil Sulfide	71.5	75.1	P/P	36 - 128
1948578	Fipronil Sulfone	67.0	68.4	P/P	16 - 145
1948578	Fonofos	88.1	81.2	P/P	60 - 112
1948578	Hexazinone	84.5	90.2	P/P	69 - 142
1948578	Malathion	80.2	75.4	P/P	68 - 132
1948578	Metalaxyl	79.6	78.7	P/P	66 - 118
1948578	Metolachlor	85.5	83.7	P/P	71 - 124
1948578	Metribuzin	172	133	F/P	61 - 140
1948578	Mevinphos	51.7	49.4	P/P	38 - 130
1948578	Molinate	32.8	31.2	P/P	23 - 99.0
1948578	Norflurazon	76.2	82.8	P/P	39 - 139
1948578	Parathion Ethyl	85.7	83.8	P/P	75 - 111
1948578	Parathion Methyl	97.7	94.1	P/P	64 - 130
1948578	Pendimethalin	99.3	95.6	P/P	26 - 173
1948578	Phorate	71.9	73.6	P/P	43 - 127
1948578	Prometon	98.3	95.5	P/P	50 - 144
1948578	Prometryn	74.0	80.6	P/P	69 - 126
1948578	Simazine	100	96.6	P/P	51 - 152
1948578	Terbufos	92.2	82.9	P/P	54 - 125
1948578	Terbutylazine	93.2	92.9	P/P	76 - 113

Reference Method: EPA 8276
Batch ID: P336154

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1948659	Chlordane	72.5	77.6	P/P	44 - 134
1948659	Toxaphene	93.5	95.5	P/P	36 - 147

Reference Method: EPA 8321B
Batch ID: P335933

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1948475	Pyraclostrobin	123	139	P/P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P335934

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1948475	Glyphosate	91.0	91.7	P/P	40 - 150

Reference Method: EPA 8321B
Batch ID: P336107

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949632	Sucralose	76.2	82.2	P/P	40 - 150

Reference Method: SM 4500 F-C-97
Batch ID: P336620

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1949722	Fluoride	98.2	97.6	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P336635

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1950126	Organic Carbon	103		P	85 - 115

Reference Method: USGS O-2060-01
Batch ID: P336034

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1948788	2,4-D	70.0	71.5	P/P	40 - 150
1948788	Acetaminophen	122	125	P/P	30 - 150
1948788	Bentazon	45.1	44.6	P/P	30 - 150
1948788	Carbamazepine	90.4	93.1	P/P	40 - 150
1948788	Diuron	85.6	88.0	P/P	40 - 150
1948788	Fenuron	66.5	74.1	P/P	40 - 150
1948788	Fluridone	89.7	94.3	P/P	40 - 150
1948788	Imidacloprid	164	174	F/F	40 - 160
1948788	Linuron	99.0	99.5	P/P	40 - 150
1948788	MCP	59.4	58.1	P/P	40 - 150
1948788	Primidone	82.9	82.9	P/P	40 - 150
1948788	Triclopyr	61.3	57.4	P/P	40 - 150

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P336257

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949887	Turbidity	1.49	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P336291

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949904	Calcium	0.567	Spike	P	0 - 20
1949904	Iron	1.20	Spike	P	0 - 20
1949904	Magnesium	1.37	Spike	P	0 - 20
1949904	Potassium	2.14	Spike	P	0 - 20
1949904	Sodium	0.284	Spike	P	0 - 20
1949904	Zinc	0.646	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P336291

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949904	Arsenic	0.413	Spike	P	0 - 20
1949904	Cadmium	1.93	Spike	P	0 - 20
1949904	Chromium	0.476	Spike	P	0 - 20
1949904	Copper	0.534	Spike	P	0 - 20
1949904	Lead	0.437	Spike	P	0 - 20
1949904	Nickel	0.114	Spike	P	0 - 20
1949904	Silver	1.09	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336481

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949769	Chloride	0.468	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P336484

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949769	Sulfate	0.213	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P336794

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949773	Ammonia-N	2.26	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P336569

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949918	Kjeldahl Nitrogen	4.05	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P336385

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949724	NO2NO3-N	1.91	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P336530

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949909	Total-P	4.27	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P336217

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949964	O-Phosphate-P	0.0881	Spike	P	0 - 20

Reference Method: EPA 8270D
Batch ID: P336154

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Aldrin	51.2	LCS	F	0 - 30
LFB	Alpha-BHC	2.57	LCS	P	0 - 30
LFB	Alpha-Chlordane	7.72	LCS	P	0 - 30
LFB	Beta-BHC	12.0	LCS	P	0 - 30
LFB	Carbophenothion	8.22	LCS	P	0 - 30
LFB	Chlorothalonil	20.1	LCS	P	0 - 30
LFB	Cypermethrin	15.7	LCS	P	0 - 30
LFB	DDD-p,p'	15.1	LCS	P	0 - 30
LFB	DDE-p,p'	5.92	LCS	P	0 - 30
LFB	DDT-p,p'	12.7	LCS	P	0 - 30
LFB	Delta-BHC	4.09	LCS	P	0 - 30
LFB	Dicofol	22.6	LCS	P	0 - 30
LFB	Dieldrin	8.34	LCS	P	0 - 30
LFB	Endosulfan I	5.88	LCS	P	0 - 30
LFB	Endosulfan II	18.9	LCS	P	0 - 30
LFB	Endosulfan Sulfate	26.7	LCS	P	0 - 30
LFB	Endrin	10.1	LCS	P	0 - 30
LFB	Endrin Aldehyde	23.9	LCS	P	0 - 30
LFB	Endrin Ketone	19.9	LCS	P	0 - 30
LFB	Gamma-BHC	9.60	LCS	P	0 - 30
LFB	Gamma-Chlordane	9.54	LCS	P	0 - 30
LFB	Heptachlor	7.86	LCS	P	0 - 30
LFB	Heptachlor Epoxide	9.01	LCS	P	0 - 30
LFB	Methoxychlor	1.66	LCS	P	0 - 30
LFB	Mirex	0.491	LCS	P	0 - 30
LFB	Permethrin	10.1	LCS	P	0 - 30
LFB	Trifluralin	2.69	LCS	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270D

Batch ID: P336157

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1948578	Acetochlor	0.995	Spike	P	0 - 30
1948578	Alachlor	1.43	Spike	P	0 - 30
1948578	Ametryn	12.8	Spike	P	0 - 30
1948578	Atrazine	2.80	Spike	P	0 - 30
1948578	Atrazine Desethyl	15.9	Spike	P	0 - 30
1948578	Azinphos Methyl	26.3	Spike	P	0 - 30
1948578	Bromacil	6.39	Spike	P	0 - 30
1948578	Butylate	5.16	Spike	P	0 - 30
1948578	Chlorpyrifos Ethyl	5.21	Spike	P	0 - 30
1948578	Chlorpyrifos Methyl	1.18	Spike	P	0 - 30
1948578	Cyanazine	18.9	Spike	P	0 - 30
1948578	Demeton	12.5	Spike	P	0 - 30
1948578	Diazinon	0.721	Spike	P	0 - 30
1948578	Disulfoton	7.99	Spike	P	0 - 30
1948578	EPTC	15.3	Spike	P	0 - 30
1948578	Ethion	67.9	Spike	F	0 - 30
1948578	Ethoprop	3.00	Spike	P	0 - 30
1948578	Fenamiphos	7.86	Spike	P	0 - 30
1948578	Fipronil	9.43	Spike	P	0 - 30
1948578	Fipronil Sulfide	4.07	Spike	P	0 - 30
1948578	Fipronil Sulfone	1.56	Spike	P	0 - 30
1948578	Fonofos	8.22	Spike	P	0 - 30
1948578	Hexazinone	6.53	Spike	P	0 - 30
1948578	Malathion	2.85	Spike	P	0 - 30
1948578	Metalaxyl	1.08	Spike	P	0 - 30
1948578	Metolachlor	2.03	Spike	P	0 - 30
1948578	Metribuzin	25.4	Spike	P	0 - 30
1948578	Mevinphos	4.66	Spike	P	0 - 30
1948578	Molinate	4.98	Spike	P	0 - 30
1948578	Norflurazon	8.24	Spike	P	0 - 30
1948578	Parathion Ethyl	2.23	Spike	P	0 - 30
1948578	Parathion Methyl	3.75	Spike	P	0 - 30
1948578	Pendimethalin	3.77	Spike	P	0 - 30
1948578	Phorate	2.35	Spike	P	0 - 30
1948578	Prometon	2.91	Spike	P	0 - 30
1948578	Prometryn	8.50	Spike	P	0 - 30
1948578	Simazine	3.80	Spike	P	0 - 30
1948578	Terbufos	10.7	Spike	P	0 - 30
1948578	Terbutylazine	0.366	Spike	P	0 - 30
LFB	Acetochlor	0.474	LCS	P	0 - 30
LFB	Alachlor	3.52	LCS	P	0 - 30
LFB	Ametryn	6.79	LCS	P	0 - 30
LFB	Atrazine	1.40	LCS	P	0 - 30
LFB	Atrazine Desethyl	0.848	LCS	P	0 - 30
LFB	Azinphos Methyl	3.83	LCS	P	0 - 30
LFB	Bromacil	0.981	LCS	P	0 - 30
LFB	Butylate	16.5	LCS	P	0 - 30
LFB	Chlorpyrifos Ethyl	2.32	LCS	P	0 - 30
LFB	Chlorpyrifos Methyl	1.36	LCS	P	0 - 30
LFB	Cyanazine	6.06	LCS	P	0 - 30
LFB	Demeton	53.6	LCS	F	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8270D
Batch ID: P336157

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	Diazinon	4.36	LCS	P	0 - 30
LFB	Disulfoton	8.55	LCS	P	0 - 30
LFB	EPTC	17.6	LCS	P	0 - 30
LFB	Ethion	11.9	LCS	P	0 - 30
LFB	Ethoprop	4.84	LCS	P	0 - 30
LFB	Fenamiphos	5.57	LCS	P	0 - 30
LFB	Fipronil	3.07	LCS	P	0 - 30
LFB	Fipronil Sulfide	2.53	LCS	P	0 - 30
LFB	Fipronil Sulfone	4.46	LCS	P	0 - 30
LFB	Fonofos	3.21	LCS	P	0 - 30
LFB	Hexazinone	3.81	LCS	P	0 - 30
LFB	Malathion	1.16	LCS	P	0 - 30
LFB	Metalaxyl	2.60	LCS	P	0 - 30
LFB	Metolachlor	0.233	LCS	P	0 - 30
LFB	Metribuzin	6.03	LCS	P	0 - 30
LFB	Mevinphos	20.2	LCS	P	0 - 30
LFB	Molinate	4.62	LCS	P	0 - 30
LFB	Norflurazon	1.58	LCS	P	0 - 30
LFB	Parathion Ethyl	2.11	LCS	P	0 - 30
LFB	Parathion Methyl	1.67	LCS	P	0 - 30
LFB	Pendimethalin	1.72	LCS	P	0 - 30
LFB	Phorate	4.51	LCS	P	0 - 30
LFB	Prometon	2.69	LCS	P	0 - 30
LFB	Prometryn	4.81	LCS	P	0 - 30
LFB	Simazine	5.55	LCS	P	0 - 30
LFB	Terbufos	1.35	LCS	P	0 - 30
LFB	Terbutylazine	0.794	LCS	P	0 - 30

Reference Method: EPA 8276
Batch ID: P336154

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1948659	Chlordane	6.83	Spike	P	0 - 30
1948659	Toxaphene	2.10	Spike	P	0 - 30
LFB	Chlordane	6.93	LCS	P	0 - 30
LFB	Toxaphene	2.36	LCS	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P335933

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1948475	Pyraclostrobin	12.2	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P335934

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1948475	Glyphosate	0.711	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P336107

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949632	Sucralose	2.51	Spike	P	0 - 30

Reference Method: SM 2120 B
Batch ID: P336158

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949892	Color (true)	1.38	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P336617

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949722	Total Alkalinity	1.19	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P336444

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949892	TDS	1.02	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P336620

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1949722	Fluoride	0.474	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P336635

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1950126	Organic Carbon	0.597	Spike	P	0 - 20

Reference Method: USGS O-2060-01
Batch ID: P336034

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1948788	2,4-D	2.26	Spike	P	0 - 30
1948788	Acetaminophen	2.70	Spike	P	0 - 30
1948788	Bentazon	1.00	Spike	P	0 - 30
1948788	Carbamazepine	2.94	Spike	P	0 - 30
1948788	Diuron	2.72	Spike	P	0 - 30
1948788	Fenuron	8.74	Spike	P	0 - 30
1948788	Fluridone	5.01	Spike	P	0 - 30
1948788	Imidacloprid	5.83	Spike	P	0 - 30
1948788	Linuron	0.427	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: USGS O-2060-01
Batch ID: P336034

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1948788	MCP	2.11	Spike	P	0 - 30
1948788	Primidone	0.0415	Spike	P	0 - 30
1948788	Triclopyr	6.48	Spike	P	0 - 30

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 1949901
Field Sample ID: G1TLHR0048-2017-05

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Acetaminophen-d4	98.6	P	30 - 160
EPA 8321B	Acetaminophen-d4	98.6	P	30 - 160
EPA 8321B	Carbamazepine-d10	91.4	P	30 - 160
EPA 8321B	Carbamazepine-d10	91.4	P	30 - 160
EPA 8321B	Diuron-d6	107	P	30 - 160
EPA 8321B	Diuron-d6	82.3	P	30 - 160
EPA 8321B	Diuron-d6	107	P	30 - 160
EPA 8321B	Diuron-d6	82.3	P	30 - 160
EPA 8321B	Glyphosate 13C 15N	115	P	40 - 160
EPA 8321B	Primidone-d5	96.1	P	30 - 160
EPA 8321B	Primidone-d5	96.1	P	30 - 160
EPA 8321B	Sucralose-d6	95.0	P	40 - 160
EPA 8321B	Sucralose-d6	95.0	P	40 - 160
USGS O-2060-01	2,4-D-d3	55.3	P	30 - 160
USGS O-2060-01	Acetaminophen-d4	98.6	P	30 - 160
USGS O-2060-01	Carbamazepine-d10	91.4	P	30 - 160
USGS O-2060-01	Diuron-d6	107	P	30 - 160
USGS O-2060-01	Diuron-d6	82.3	P	30 - 160
USGS O-2060-01	Primidone-d5	96.1	P	30 - 160
USGS O-2060-01	Sucralose-d6	95.0	P	40 - 160

Lab Sample ID: 1949902
Field Sample ID: G1TLHR0048-2017-05

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270D	Decachlorobiphenyl	81.8	P	39 - 118
EPA 8270D	Tetrachloro-m-xylene	55.2	P	32 - 103
EPA 8276	Decachlorobiphenyl	69.9	P	31 - 100

Lab Sample ID: 1949903
Field Sample ID: G1TLHR0048-2017-05

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8270D	Malathion-d7	100	P	53 - 135
EPA 8270D	Simazine-d10	103	P	71 - 126

Quality Assurance Report Calibration Verification

Reference Method: EPA 8321B
 Run ID: A80568
 Included Lab Sample IDs: 1949901

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	107	72.7	P/P	60 - 150

Reference Method: SM 2120 B
 Run ID: A80620
 Included Lab Sample IDs: 1949897

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	103	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Run ID: A80641
 Included Lab Sample IDs: 1949899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	101	100	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
 Run ID: A80658
 Included Lab Sample IDs: 1949897

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	100	P/P	95 - 105

Reference Method: EPA 8321B
 Run ID: A80660
 Included Lab Sample IDs: 1949901

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Glyphosate	95.0	101	P/P	60 - 150
Pyraclostrobin	79.4	81.4	P/P	60 - 150

Reference Method: USGS O-2060-01
 Run ID: A80693
 Included Lab Sample IDs: 1949901

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	98.5	91.6	P/P	50 - 150
Acetaminophen	117	115	P/P	60 - 150
Bentazon	100	98.3	P/P	50 - 150
Carbamazepine	115	113	P/P	60 - 150
Diuron	111	104	P/P	60 - 150
Fenuron	114	112	P/P	60 - 150
Fluridone	117	116	P/P	60 - 160
Imidacloprid	108	103	P/P	60 - 150
Linuron	110	115	P/P	60 - 150
MCPP	98.8	92.2	P/P	50 - 150
Primidone	119	114	P/P	60 - 150
Triclopyr	90.6	83.9	P/P	50 - 150

Quality Assurance Report Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A80702

Included Lab Sample IDs: 1949898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	102	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A80712

Included Lab Sample IDs: 1949904

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Arsenic	101	100	P/P	90 - 110
Cadmium	97.0	96.7	P/P	90 - 110
Chromium	94.6	94.3	P/P	90 - 110
Copper	102	101	P/P	90 - 110
Lead	97.2	97.7	P/P	90 - 110
Nickel	102	101	P/P	90 - 110
Silver	97.7	97.3	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A80735

Included Lab Sample IDs: 1949897

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Sulfate	101	101	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A80748

Included Lab Sample IDs: 1949904

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	95.6	94.8	P/P	90 - 110
Iron	94.1	90.8	P/P	90 - 110
Magnesium	95.9	93.6	P/P	90 - 110
Potassium	101	99.3	P/P	90 - 110
Sodium	98.2	95.9	P/P	90 - 110
Zinc	97.8	97.8	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A80770

Included Lab Sample IDs: 1949897

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	103	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A80770

Included Lab Sample IDs: 1949897

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	103	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-00
Run ID: A80775
Included Lab Sample IDs: 1949898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.0	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A80797
Included Lab Sample IDs: 1949898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.0	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A80809
Included Lab Sample IDs: 1949898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A80825
Included Lab Sample IDs: 1949898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	104	103	P/P	90 - 110

Reference Method: EPA 8270D
Run ID: A80861
Included Lab Sample IDs: 1949903

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetochlor	98.5		P	80 - 120
Alachlor	97.6		P	80 - 120
Ametryn	101		P	80 - 120
Atrazine	98.2		P	80 - 120
Atrazine Desethyl	90.0		P	80 - 120
Azinphos Methyl	91.0		P	80 - 120
Bromacil	105		P	80 - 120
Butylate	97.6		P	80 - 120
Chlorpyrifos Ethyl	97.4		P	80 - 120
Chlorpyrifos Methyl	98.2		P	80 - 120
Cyanazine	72.7*		F	80 - 120
Demeton	91.7		P	80 - 120
Diazinon	94.1		P	80 - 120
Disulfoton	95.6		P	80 - 120
EPTC	100		P	80 - 120
Ethion	90.9		P	80 - 120
Ethoprop	95.5		P	80 - 120
Fenamiphos	99.0		P	80 - 120
Fipronil	98.8		P	80 - 120
Fipronil Sulfide	99.3		P	80 - 120
Fipronil Sulfone	100		P	80 - 120
Fonofos	97.2		P	80 - 120
Hexazinone	94.0		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270D

Run ID: A80861

Included Lab Sample IDs: 1949903

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Malathion	95.6		P	80 - 120
Metalaxyl	101		P	80 - 120
Metolachlor	105		P	80 - 120
Metribuzin	92.2		P	80 - 120
Mevinphos	99.1		P	80 - 120
Molinate	92.6		P	80 - 120
Norflurazon	96.0		P	80 - 120
Parathion Ethyl	92.0		P	80 - 120
Parathion Methyl	93.8		P	80 - 120
Pendimethalin	97.3		P	80 - 120
Phorate	98.0		P	80 - 120
Prometon	101		P	80 - 120
Prometryn	101		P	80 - 120
Simazine	96.0		P	80 - 120
Terbufos	96.4		P	80 - 120
Terbuthylazine	96.7		P	80 - 120

Reference Method: EPA 8276

Run ID: A80885

Included Lab Sample IDs: 1949902

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chlordane	117		P	80 - 120
Toxaphene	117		P	80 - 120

Reference Method: EPA 8270D

Run ID: A81070

Included Lab Sample IDs: 1949902

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Aldrin	105		P	80 - 120
Alpha-BHC	101		P	80 - 120
Alpha-Chlordane	97.9		P	80 - 120
Beta-BHC	87.3		P	80 - 120
Carbophenothion	88.8		P	80 - 120
Chlorothalonil	94.0		P	80 - 120
Cypermethrin	88.1		P	80 - 120
DDD-p,p'	89.4		P	80 - 120
DDE-p,p'	103		P	80 - 120
DDT-p,p'	105		P	80 - 120
Delta-BHC	95.7		P	80 - 120
Dicofol	93.7		P	80 - 120
Dieldrin	99.7		P	80 - 120
Endosulfan I	101		P	80 - 120
Endosulfan II	104		P	80 - 120
Endosulfan Sulfate	87.7		P	80 - 120
Endrin	96.2		P	80 - 120
Endrin Aldehyde	88.8		P	80 - 120
Endrin Ketone	93.3		P	80 - 120
Gamma-BHC	99.6		P	80 - 120
Gamma-Chlordane	97.4		P	80 - 120

Quality Assurance Report Calibration Verification

Reference Method: EPA 8270D
Run ID: A81070
Included Lab Sample IDs: 1949902

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Heptachlor	102		P	80 - 120
Heptachlor Epoxide	98.0		P	80 - 120
Methoxychlor	105		P	80 - 120
Mirex	99.9		P	80 - 120
Permethrin	91.6		P	80 - 120
Trifluralin	99.4		P	80 - 120

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery			LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity							1.49	
EPA 200.7 Rev. 4.4	Calcium	101		97.8	96.5	95.5			0.567
	Iron	98.9		98.5	94.8	95.9			1.20
	Magnesium	105		98.3	100	102			1.37
	Potassium	106		104	101				2.14
	Sodium	102		101	97.3				0.284
	Zinc	98.9		96.5	98.2	97.6			0.646
EPA 200.8 Rev. 5.4	Arsenic	102		104	98.7	99.1			0.413
	Cadmium	98.4		100	97.2	99.1			1.93
	Chromium	97.0		94.4	93.9	94.3			0.476
	Copper	99.1		96.4	96.6	97.2			0.534
	Lead	97.6		98.5	97.2	97.6			0.437
	Nickel	101		97.7	97.8	97.9			0.114
	Silver	100		101	98.6	99.6			1.09
EPA 300.0 Rev. 2.1	Chloride	97.0		95.2	95.8			0.468	
	Sulfate	98.0		97.2	97.8			0.213	
EPA 350.1 Rev. 2.0	Ammonia-N	104		102	98.2				2.26
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.8		96.5	102				4.05
EPA 353.2 Rev. 2.0	NO2NO3-N	100		106	104				1.91
EPA 365.1 Rev. 2.0	Total-P	103		97.8	102				4.27
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.4		98.8	98.9				0.0881
EPA 8270D	Acetochlor	99.2	98.8	85.0	85.8		0.474		0.995
	Alachlor	100	96.6	90.2	91.6		3.52		1.43
	Aldrin	50.1	84.6	57.8			51.2		
	Alpha-BHC	110	108	107			2.57		
	Alpha-Chlordane	107	99.2	91.5			7.72		
	Ametryn	84.1	90.0	68.5	77.9		6.79		12.8
	Atrazine	103	104	95.4	92.4		1.40		2.80
	Atrazine Desethyl	82.5	83.2	98.4	83.9		0.848		15.9
	Azinphos Methyl	113	109	104	79.5		3.83		26.3
	Beta-BHC	138	123	88.8			12.0		
	Bromacil	100	99.2	76.0	86.7		0.981		6.39
	Butylate	59.7	50.6	26.3	27.7		16.5		5.16
	Carbophenothion	91.9	84.6	126			8.22		

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		Precision	
						LCS	MS
						SMP	
EPA 8270D	Chlorothalonil	111	91.0	4.12		20.1	
	Chlorpyrifos Ethyl	93.7	91.6	74.1	78.0	2.32	5.21
	Chlorpyrifos Methyl	100	99.0	80.4	79.5	1.36	1.18
	Cyanazine	161	171	152	126	6.06	18.9
	Cypermethrin	125	107	127		15.7	
	DDD-p,p'	124	107	78.4		15.1	
	DDE-p,p'	110	104	80.6		5.92	
	DDT-p,p'	122	107	74.9		12.7	
	Delta-BHC	142	147	110		4.09	
	Demeton	45.3	78.5	80.1	70.6	53.6	12.5
	Diazinon	95.9	91.8	80.7	80.1	4.36	0.721
	Dicofol	110	87.5	83.6		22.6	
	Dieldrin	112	103	92.0		8.34	
	Disulfoton	109	99.7	105	96.5	8.55	7.99
	Endosulfan I	115	109	109		5.88	
	Endosulfan II	145	120	112		18.9	
	Endosulfan Sulfate	125	95.7	96.5		26.7	
	Endrin	115	104	93.5		10.1	
	Endrin Aldehyde	126	98.9	65.2		23.9	
	Endrin Ketone	123	101	100		19.9	
	EPTC	35.9	42.9	22.6	26.3	17.6	15.3
	Ethion	92.2	81.8	41.0	83.1	11.9	67.9
	Ethoprop	102	96.8	93.3	90.6	4.84	3.00
	Fenamiphos	109	103	107	98.9	5.57	7.86
	Fipronil	102	98.4	76.9	85.3	3.07	9.43
	Fipronil Sulfide	93.2	90.9	71.5	75.1	2.53	4.07
	Fipronil Sulfone	89.5	85.6	67.0	68.4	4.46	1.56
	Fonofos	93.8	96.9	88.1	81.2	3.21	8.22
	Gamma-BHC	113	124	109		9.60	
	Gamma-Chlordane	104	94.6	90.4		9.54	
	Heptachlor	95.4	88.2	96.0		7.86	
	Heptachlor Epoxide	110	100	96.3		9.01	
	Hexazinone	90.3	93.8	84.5	90.2	3.81	6.53
	Malathion	99.9	98.8	80.2	75.4	1.16	2.85
	Metalaxyl	102	99.8	79.6	78.7	2.60	1.08
	Methoxychlor	129	127	113		1.66	
	Metolachlor	98.2	98.0	85.5	83.7	0.233	2.03
	Metribuzin	104	110	172	133	6.03	25.4
	Mevinphos	56.9	69.7	51.7	49.4	20.2	4.66
	Mirex	86.5	86.1	94.3		0.491	
	Molinate	50.4	48.2	32.8	31.2	4.62	4.98
	Norflurazon	93.3	91.9	76.2	82.8	1.58	8.24
	Parathion Ethyl	94.2	96.2	85.7	83.8	2.11	2.23
	Parathion Methyl	104	105	97.7	94.1	1.67	3.75
	Pendimethalin	105	107	99.3	95.6	1.72	3.77
	Permethrin	108	97.9	108		10.1	
	Phorate	87.8	84.0	71.9	73.6	4.51	2.35
	Prometon	94.7	97.3	98.3	95.5	2.69	2.91
	Prometryn	96.8	92.3	74.0	80.6	4.81	8.50
	Simazine	97.9	104	100	96.6	5.55	3.80
	Terbufos	97.1	95.8	92.2	82.9	1.35	10.7
	Terbuthylazine	98.6	99.4	93.2	92.9	0.794	0.366
	Trifluralin	97.1	94.5	97.0		2.69	
EPA 8276	Chlordane	87.0	93.3	72.5	77.6	6.93	6.83

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		Precision	
						LCS	MS
EPA 8276	Toxaphene	113	111	93.5	95.5	2.36	2.10
EPA 8321B	Glyphosate	87.6		91.0	91.7		0.711
	Pyraclostrobin	132		123	139		12.2
	Sucralose	67.4		76.2	82.2		2.51
	Color (true)					1.38	
SM 2120 B	Total Alkalinity	103				1.19	
SM 2320 B-97	TDS					1.02	
SM 2540 C-97	Fluoride	98.2		98.2	97.6		0.474
SM 4500 F-C-97	Organic Carbon	99.1		103			0.597
SM 5310 B-00	2,4-D	97.6		70.0	71.5		2.26
USGS O-2060-01	Acetaminophen	115		125	122		2.70
	Bentazon	75.8		45.1	44.6		1.00
	Carbamazepine	109		93.1	90.4		2.94
	Diuron	106		88.0	85.6		2.72
	Fenuron	109		74.1	66.5		8.74
	Fluridone	105		94.3	89.7		5.01
	Imidacloprid	98.3		174	164		5.83
	Linuron	100		99.5	99.0		0.427
	MCPP	92.1		59.4	58.1		2.11
	Primidone	93.9		82.9	82.9		0.0415
	Triclopyr	89.1		61.3	57.4		6.48

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1949897
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	1949904
EPA 200.8 Rev. 5.4 / E31780	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	1949904
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1949897
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1949897
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1949898
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1949898
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1949898
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1949898
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1949899
EPA 8270D / E31780	Organochlorine pesticides in water matrices by GC/MS-SIM	1949902
EPA 8270D / E31780	Organonitrogen and phosphorus pesticides (ultra trace level) in water matrices by GC/MS/MS	1949903
EPA 8276 / E31780	Organochlorine pesticides in water matrices by GC/MS with Negative Chemical Ionization	1949902
EPA 8321B / E31780	Analysis of sucralose in water matrices by HPLC/MS/MS	1949901
EPA 8321B / E31780	Glyphosate in water matrices by HPLC/MS/MS	1949901
EPA 8321B / E31780	Pyraclostrobin in water matrices by HPLC/MS/MS	1949901
SM 2120 B / E31780	True Color measured at 450 nm	1949897
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1949897

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1949897
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1949897
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1949897
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1949898
USGS O-2060-01 / E31780	Chlorinated (phenoxy acid) herbicides in water matrices by HPLC/MS/MS	1949901
USGS O-2060-01 / E31780	Urea herbicides and other chemicals in water matrices by HPLC/MS/MS	1949901

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/30/2017			12/01/2017 15:40	Tanya Welborn	1949897
EPA 200.7 Rev. 4.4	11/30/2017	12/04/2017	Elliott D. Healy	12/06/2017	Martin Acosta	1949904
EPA 200.8 Rev. 5.4	11/30/2017	12/04/2017	Elliott D. Healy	12/05/2017	Nadine Brooks	1949904
EPA 300.0 Rev. 2.1	11/30/2017			12/05/2017	Julia Storbeck	1949897
EPA 350.1 Rev. 2.0	11/30/2017			12/08/2017	Sofia Elnemr	1949898
EPA 351.2 Rev. 2.0	11/30/2017	12/07/2017	Adrian Shelby	12/08/2017	Joseph Suarez	1949898
EPA 353.2 Rev. 2.0	11/30/2017			12/05/2017	Lauren Bottorf	1949898
EPA 365.1 Rev. 2.0	11/30/2017	12/06/2017	Sima Feizi	12/11/2017	Beverly Y. Sanders	1949898
EPA 365.1 Rev. 2.0 dissolved	11/30/2017			12/01/2017 14:52	Megan Treadwell	1949899
EPA 8270D	11/30/2017	12/01/2017	John Kaba	12/05/2017	Vandana T. Nagar	1949903
	11/30/2017	12/01/2017	Rasheda Ghaffari	12/06/2017	Goutam Palui	1949902
EPA 8276	11/30/2017	12/01/2017	Rasheda Ghaffari	12/12/2017	John P. Burgos	1949902
EPA 8321B	11/30/2017	11/30/2017	Juyoung Kim	11/30/2017	Juyoung Kim	1949901
	11/30/2017	12/01/2017	Juyoung Kim	12/01/2017	Juyoung Kim	1949901
SM 2120 B	11/30/2017			12/01/2017 09:45	Tanya Welborn	1949897
SM 2320 B-97	11/30/2017			12/07/2017	Dale L. Simmons	1949897
SM 2540 C-97	11/30/2017			12/01/2017	DeAsia Armster	1949897
SM 2540 D-97	11/30/2017			12/01/2017	DeAsia Armster	1949897
SM 4500 F-C-97	11/30/2017			12/07/2017	Dale L. Simmons	1949897
SM 5310 B-00	11/30/2017			12/07/2017	Ping Hua	1949898
USGS O-2060-01	11/30/2017	12/04/2017	Hoor Shaik	12/05/2017	Julie Michelle Frank	1949901
	11/30/2017	12/04/2017	Hoor Shaik	12/06/2017	Julie Michelle Frank	1949901