

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: 1015

Project: SMP 2017-04-17-28

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 4-4-17

DO FDEP SOP FD 1000	Name	Date	Time CET	Temp C/F	D.O. Chan mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	L. Bachler	4/8/17	0715	23.9	8.43	8.43	100	53.2	1.09	P	L
ICV			0717	24.0	8.41	8.34	99	-	-	P	L
CCV			1600	24.9	8.27	8.13	98	-	-	P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.
 DO Acceptance criteria from Table ± 0.3 mg/L. Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
 Optical: DO gain range 0.85 to 1.15; DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Conc. FDEP SOP FD 1200	Name	Date	Time CET	Lot #	Expi. Date	Standard umhos/cm	Meter Reading umhos/cm	Pass / Fail	Lab / Field
Calibr.	L. Bachler	4/8/17	0719	2608442	7/18	1000	1000	P	L
ICV			0721	2610009	6/18	100	99	P	L
CCV			1603	2608435	7/18	501000	4944	P	L
CCV									

Report specific conductance as a whole number with no decimal figure. Conductivity Acceptance Criteria $\pm 5\%$

pH FDEP SOP FD 1100	Name	Date	Time CET	Lot #	Expi. Date	pH Buffer SU	Meter Reading SU	mV	Pass / Fail	Lab / Field
Calibr.	L. Bachler	4/8/17	0723	163419	6/18	7.0	7.0	56.6	P	L
Calibr.			0725	163418	5/18	4.0	4.0	119.9	P	L
Calibr.			0728	163419	6/18	10.0	10.0	233.5	P	L
ICV			0730	163419	6/18	10.0	10.0	-	P	L
CCV			1607	163419	6/18	7.0	7.0	-	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ; mV pH 4 Range $\pm 180 \pm 50$;
 mV pH 10 Range -180 ± 50 ; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CET	Zero the Sensor	ICV Value	Pass/ Fail	Lab/ Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater