

2017-08-14-28

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

Busta

Project:

SVP

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____Temperature (Quarterly) FT 1400Date of Last Temperature Verification 07/18/2017

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	L. Bachler	8-16-17	0715	25.3	8.21	8.22	100	-17.6	0.74	P	L
ICV			0717	25.2	8.23	8.24	100	—	—	P	L
CCV			1600	25.0	8.26	8.29	98	—	—	P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	L. Bachler	8-16-17	0719	2703D066	03/19	1000	1000	P	L
ICV			0721	2703994	02/18	100	102	P	L
CCV			1602	2611944	10/18	500	503	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	L. Bachler	8-16-17	0725	167097	10/18	7.0	7.0	-12.7	P	L
Calibr.			0727	167431	11/18	4.0	4.0	166.7	P	L
Calibr.			0729	164223	07/18	10.0	10.0	-189.4	P	L
ICV			0732	164223	07/18	10.0	10.0	—	P	L
CCV			1605	167431	11/18	4.0	4.0	—	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50;mV pH 4 Range +180 $\pm$ 50;mV pH 10 Range -180 $\pm$ 50; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mVDepth (Quarterly)

Date of Last Depth Verification _____

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater