

Chemical Analysis Report

SW-DIST-2017-09-06-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **SMP - DO Monitoring Retrieval**

Request ID: **RQ-2017-09-04-25**

Customer: **SW-DIST**

Project ID: **SMP**

Send Reports to:

FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
13051 N. Telecom Parkway
Temple Terrace, FL 33637-0926
Attn: Sarah Meyer

For additional information please contact

Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 22-SEP-2017 10:41



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Woods Creek Continuous DO

Collection Date/Time: 09/05/2017 11:00

Field ID: G1SW0087

Matrix: W-SURF-SLT

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1927898	EPA 180.1 Rev. 2.0	Turbidity	4.7		NTU	P331392	
	SM 2320 B-97	Total Alkalinity	101		mg CaCO3/L	P331538	
	SM 2540 D-97	TSS	7	I	mg/L	P331683	
	SM 4500 F-C-97	Fluoride	0.80		mg F/L	P331745	
1927899	EPA 350.1 Rev. 2.0	Ammonia-N	0.009		mg N/L	P331653	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.78		mg N/L	P331606	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P331668	
	EPA 365.1 Rev. 2.0	Total-P	0.16		mg P/L	P331566	
	SM 5310 B-00	Organic Carbon	6.8		mg C/L	P331765	
1927900	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.096		mg P/L	P331474	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P331392

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P331653

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P331606

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0

Batch ID: P331668

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0

Batch ID: P331566

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Batch ID: P331474

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2320 B-97
 Batch ID: P331538

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 D-97
 Batch ID: P331683

Component	Result	Code	Units
TSS	3	U	mg/L

Reference Method: SM 4500 F-C-97
 Batch ID: P331745

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
 Batch ID: P331765

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P331653

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	107		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P331606

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P331668

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P331566

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	98.4		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P331474

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P331474

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.7		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P331538

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	95 - 108

Reference Method: SM 4500 F-C-97
Batch ID: P331745

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.3		P	92 - 103

Reference Method: SM 5310 B-00
Batch ID: P331765

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P331653

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1927668	Ammonia-N	109	107	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P331606

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1927899	Kjeldahl Nitrogen	94.3	98.1	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P331668

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1927899	NO2NO3-N	105	105	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P331566

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1927899	Total-P	91.6	91.4	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P331474

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P331474

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1927900	O-Phosphate-P	97.6	100	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P331745

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1927215	Fluoride	93.5	93.5	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P331765

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1927217	Organic Carbon	95.3	94.8	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P331392

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1927786	Turbidity	2.87	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P331653

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1927668	Ammonia-N	2.11	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P331606

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1927899	Kjeldahl Nitrogen	2.94	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P331668

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1927899	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P331566

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1927899	Total-P	0.201	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P331474

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1927900	O-Phosphate-P	1.38	Spike	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P331538

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1927215	Total Alkalinity	0.608	Sample	P	0 - 4.8

Reference Method: SM 4500 F-C-97
Batch ID: P331745

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1927215	Fluoride	0.0	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P331765

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1927217	Organic Carbon	0.497	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A78829
Included Lab Sample IDs: 1927898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A78860
Included Lab Sample IDs: 1927900

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	102	102	P/P	90 - 110

Reference Method: SM 2320 B-97
Run ID: A78874
Included Lab Sample IDs: 1927898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	104	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A78921

Included Lab Sample IDs: 1927899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	106	106	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A78923

Included Lab Sample IDs: 1927899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	105	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A78929

Included Lab Sample IDs: 1927899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A78959

Included Lab Sample IDs: 1927898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	101	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A78968

Included Lab Sample IDs: 1927899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.3	95.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A78981

Included Lab Sample IDs: 1927899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	92.3	98.9	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements. Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				2.87	
EPA 350.1 Rev. 2.0	Ammonia-N	107	109	107		2.11
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	94.3	98.1		2.94
EPA 353.2 Rev. 2.0	NO2NO3-N	104	105	105		0.0

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 365.1 Rev. 2.0	Total-P	98.4	91.6	91.4			0.201
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.7	97.6	100			1.38
SM 2320 B-97	Total Alkalinity	103				0.608	
SM 4500 F-C-97	Fluoride	98.3	93.5	93.5			0.0
SM 5310 B-00	Organic Carbon	95.8	95.3	94.8			0.497

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1927898
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1927899
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1927899
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1927899
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1927899
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1927900
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1927898
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1927898
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1927898
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1927899

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/06/2017			09/06/2017 17:24	DeAsia Armster	1927898
EPA 350.1 Rev. 2.0	09/06/2017			09/14/2017	Sofia Elnemr	1927899
EPA 351.2 Rev. 2.0	09/06/2017	09/14/2017	Adrian Shelby	09/18/2017	Joseph Suarez	1927899
EPA 353.2 Rev. 2.0	09/06/2017			09/15/2017	Beverly Y. Sanders	1927899
EPA 365.1 Rev. 2.0	09/06/2017	09/13/2017	Sima Feizi	09/14/2017	Lipi Saha	1927899
EPA 365.1 Rev. 2.0 dissolved	09/06/2017			09/06/2017 14:29	Megan Treadwell	1927900
SM 2320 B-97	09/06/2017			09/08/2017	Dale L. Simmons	1927898
SM 2540 D-97	09/06/2017			09/07/2017	Yijie Li	1927898
SM 4500 F-C-97	09/06/2017			09/16/2017	Dale L. Simmons	1927898
SM 5310 B-00	09/06/2017			09/16/2017	Sofia Elnemr	1927899