



Queue: MICt

Batch Number: 3664

I. Receipt

No Exceptions were encountered.

II. Holding Times

Preparation: All holding times were met.

Analysis: T717727001 and 002 was received by the lab without sufficient processing time. The analysis was performed as soon as possible after receipt by the laboratory. The data is qualified to indicate the holding time violation.

III. Method

Analysis: EPA 1603

Preparation: None

IV. Preparation

Sample preparation proceeded normally.

V. Analysis

A. Calibration: All acceptance criteria were met.

B. Blanks: All acceptance criteria were met.

C. Duplicates: All acceptance criteria were met.

D. Spikes: All acceptance criteria were met.

E. Serial Diluion: All acceptance criteria were met.

F. Samples: Sample analyses proceeded normally.

G. Other:

I certify that this data package is in compliance with the terms and conditions agreed to by Advanced Environmental Laboratories, Inc. and by the client, both technically and for completeness, except for the conditions detailed above. The Quality Assurance Officer, or designee, as verified by the following signature, has authorized release of the data contained in this data package: