

RQ-2017-01-16-36

Boldly "X" this box if there is qualified data on this page.

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID: Lil Gus

Project: 26 CA 1001

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 12/22/16

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	A. Hurley	1/25/17	0800	21.5	8.8	8.8	100	60	1.2	P	L
ICV	"	"	"	21.5	8.8	8.8	100			P	L
CCV	Matt Seal	1/25/17	17:00	22.0	8.7	8.8	101	53		P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	A. Hurley	1/25/17	0800	260864	10/18	1000	1000	P	L
ICV	"	"	"	2604736	3/18	100	103	P	L
CCV	Matt Seal	1/25/17	17:15	260525	12/17	5000	5001	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	A. Hurley	1/25/17	0900	161856	4/18	7.0	7.0	-13	P	L
Calibr.	"	"	"	163027	5/18	4.0	4.0	163	P	L
Calibr.	"	"	"	160910	3/18	10.0	10.0	191	P	L
ICV	"	"	"	160910	3/18	10.0	10.0		P	L
CCV	Matt Seal	1/25/17	17:00	163027	10/18	4.0	4.0	164	P	L
CCV		25								

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification 12/22/16

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	A. Hurley	1/25/17	0800	0.000	6.00	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater