

RQ-2017-03-06-25

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID:

L1 Gas

Project:

SMP IRL

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on SiteTemperature (Quarterly) FT 1400

Date of Last Temperature Verification

1/23/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	3/9/17	828	23.9	8.4	8.4	100	60.4	1.24	P	L
ICV	//	//	830	23.9	8.4	8.4	100	60.4		P	L
CCV	//	//	1704	20.3	9.0	9.1	107	63.5		F	L
CCV	//	//	1710	20.6	9.0	9.3	104	64.5		P	L

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.**Rapid-Pulse Sensors:** DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.**Optical:** DO gain range 0.85 to 1.15; DO charge N/A.**Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	3/9/17	832	2610B64	10/18	1000	1000	P	L
ICV	//	//	834	2604736	3/18	100	101	P	L
CCV	//	//	1711	1610091	9/18	100k	97200	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	3/9/17	836	165674	8/18	7.0	7.0	-13.0	P	L
Calibr.			838	164668	7/18	4.0	4.0	165	P	L
Calibr.			840	160910	3/18	10.0	10.0	-188.2	P	L
ICV			842	160910	3/18	10.0	10.0	-187.7	P	L
CCV	//	//	1713	164668	7/18	4.0	4.0	163.2	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;mV pH 4 Range $+180 \pm 50$;mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

1/23/17

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	Drew Liddick	3/9/17	844	0.000	0.000	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater