

RQ-2017-04-03-23

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

AE3

Project:

SMP

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

11/23/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	4/11/17	800	20.5	9.0	9.0	100	49.2	1.14	P	L
ICV	↓	↓	0802	20.5	9.0	9.0	100	49.2	—	P	L
CCV	Drew Liddick	"	0805	22.3	8.7	8.6	99	59.4		P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard $\mu\text{mhos/cm}$	Meter Reading $\mu\text{mhos/cm}$	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	4/11/17	0804	2610B64	10/18	1000	1000	P	L
ICV	↓	↓	0805	2604736	3/18	100	103	P	L
CCV	Drew Liddick	"	1730	2511638	11/17	50 K	48940	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	4/11/17	0807	165674	8/18	7.0	7.0	-23.6	P	L
Calibr.	↓	↓	0808	164668	7/18	4.0	4.0	153.1	P	L
Calibr.	↓	↓	0809	164223	7/18	10.0	10.0	-199.6	P	L
ICV	↓	↓	0810	164223	7/18	10.0	10.0	-199.4	P	L
CCV	Drew Liddick	"	1732	164668	7/18	4.0	4.0	150.8	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;mV pH7 Range 0 ± 50 ;mV pH 4 Range $+180 \pm 50$;mV pH 10 Range -180 ± 50 ;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

11/23/17

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	Drew Liddick	4/11/17	0812	0.000	0.00	P	L

James P. Jones

ICV acceptance criteria $\pm 5\%$ or $\pm 0.05\text{m}$, whichever is greater