

RQ-2017-06-05-25

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name

Meter ID:

AG / ISI#770

Project:

Dade

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on SiteTemperature (Quarterly) FT 1400Date of Last Temperature Verification 3/13/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	6/5/17	740	23.8	8.5	8.5	100	40.0	1.23	P	L
ICV	//	//	742	23.8	8.5	8.5	100	40.0	-	P	L
CCV	//	//	1640	20.6	9.0	9.1	101	39.6	-	P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.**Rapid-Pulse Sensors:** DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.**Optical:** DO gain range 0.85 to 1.15; DO charge N/A.**Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	6/5/17	744	2610864	10/18	1000	1000	P	L
ICV	//	//	746	2606608	5/18	100	100	P	L
CCV	//	//	1642	1610691	9/18	100K	101295	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Drew Liddick	6/5/17	748	165674	8/18	7.0	7.0	-43.9	P	L
Calibr.			750	164668	7/18	4.0	4.0	133.8	P	L
Calibr.			752	164223	7/18	10.0	10.0	-222.8	P	L
ICV			754	164223	7/18	10.0	10.0	-221.1	P	L
CCV			1644	171241	3/19	4.0	4.0	136.5	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50;mV pH 10 Range -180 $\pm$ 50; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mVDepth (Quarterly)Date of Last Depth Verification 3/13/17

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	Drew Liddick	6/5/17	756	0.000	0.000	P	L

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater