

RQ-2017-07-10-12

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Meter ID:

L-1 G-5

Project:

Dade Green

Boldly "X" this box if there is qualified data on this page.

- Notes: (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 7/11/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. P. 155	8/15/17	0735	24.5	8.3	8.3	100	67.6	0.90	P	L
ICV	↓	↓	0756	24.4	8.4	8.4	100	67.6	—	P	L
CCV	D. L. 1414	"	1612	24.4	8.4	8.3	100	67.6	—	P	L
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
 Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard $\mu\text{mhos/cm}$	Meter Reading $\mu\text{mhos/cm}$	Pass / Fail	Lab / Field
Calibr.	J. P. 155	8/15/17	0737	240864	10/18	1000	1000	P	L
ICV	↓	↓	0758	260608	5/18	100	100	P	L
CCV	D. L. 1414	"	1614	2610861	10/18	1000	998	P	L
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. P. 155	8/15/17	0800	165674	8/18	7.0	7.0	-19.7	P	L
Calibr.	↓	↓	0801	171241	3/19	4.0	4.0	-153.4	P	L
Calibr.	↓	↓	0802	164223	7/18	10.0	10.0	-197.1	P	L
ICV	↓	↓	0803	164223	7/18	10.0	10.0	-197.1	P	L
CCV	D. L. 1414	"	1616	171241	3/19	4.0	4.0	-153.4	P	L
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU;

mV pH7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification 7/11/17

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air	J. P. 155	8/15/17	0805	0.000	0.00	P	L

ICV acceptance criteria $\pm 5\%$ or $\pm 0.05\text{m}$, whichever is greater