From:	June Flowers
To:	DEP-Overflowmicro; Matthews, Daisys
Cc:	Phil Loucks; Savage, Jacqueline
Subject:	Re: Your FCL report is ready. 329713SW1 DEP
Date:	Monday, May 1, 2017 12:04:36 PM

Hello,

I just confirmed that only 47 of the small wells fluoresced, so that # (2419.6) is the same as if 48 small and 49 large wells are positive…that is when we are to "J" flag the value.

Thank you,

June Flowers
Quality Systems Director
Flowers **C**hemical **L**aboratories, Inc.
407.339.5984 x212
www.flowerslabs.com

Celebrating 60 years 1957 - 2017
The information contained in this e-mail message is intended only for the personal and confidential use of the recipient(s) named above. This message may be client privileged communication and as such is privileged and confidential. If the reader of this message is not the intended recipient, you are hereby notified that you have received this message in error and that any review, dissemination, distribution, or copying of this message is strictly prohibited. If you received this message in error, please notify the sender immediately by e-mail, and delete the original message.

E-mail transmission cannot be guaranteed to be secure or error-free. The sender therefore does not accept liability for any damage caused by any virus transmitted by this e-mail or for any errors or omissions in the contents of this message, which arise as a result of e-mail transmission.

On May 1, 2017, at 9:02 AM, Phil wrote:

-------- Forwarded Message --------
Subject:RE: Your FCL report is ready.
Date: Fri, 28 Apr 2017 18:57:34 +0000
From:DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
To:phil@flowerslabs.com <phil@flowerslabs.com>
CC: Savage, Jacqueline <Jacqueline.Savage@dep.state.fl.us>, Matthews, Daisys <Daisys.Matthews@dep.state.fl.us>

```
Good Afternoon,

I'd like to request the addition of a "J" qualifier to sample
#G5SE0050.
Please, let me know if there are any questions or concerns.

Thank You,

Carina Tull
OPS Environmental Specialist I
FDEP Laboratory - Biology Program
```

Florida Department of Environmental Protection
Carina.Tull@dep.state.fl.us
Office: 850.245.8174

-----Original Message-----
From: labinfo@flowerslabs.com [mailto:labinfo@flowerslabs.com]
Sent: Friday, April 21, 2017 1:37 PM
To: phil@flowerslabs.com; Matthews, Daisys <Daisys.Matthews@dep.state.fl.us>; DEP-Overflowmicro <DEP-Overflowmicro@dep.state.fl.us>
Subject: Your FCL report is ready.

April 21, 2017

Thank you for doing business with Flowers Chemical Laboratories! Results for your samples sampled on April 19, 2017 with Project Number SMP G5 North is available at http://www.flowerslabs.com/Apps/WebObjects/FCLSite.woa/wa/arpt?ar=BYAx8Ftk66AHeS

Please note that Adobe Acrobat Reader is required to view this report. If you don't already have this program, it may be downloaded at http://www.adobe.com/products/acrobat/readstep2.html.
[Dep Customer Survey]<http://survey.dep.state.fl.us/?refemail=DEP-Overflowmicro@dep.state.fl.us>