

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boxed "X" this box if there is qualified data on the online

Meter Bio #2

RQ- 2015-05-01-28

Project Smp

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 03/06/2017

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Curtis Mussen	5/2/17	708	22.37	8.677	8.69	100.0	38.0	-	P	LAB
ICV	Curtis Mussen	5/2/17	710	22.35	8.677	8.71	100.4	38.0	-	P	LAB
CCV	Curtis Mussen	5/2/17	1442	22.65	8.627	9.80	113.5	39.0	-	Fail	LAB
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	Curtis Mussen	5/2/17	712	16111D	11/17	1000	1000	P	LAB
ICV	Curtis Mussen	5/2/17	713	16111D	11/17	1000	1000	P	LAB
CCV	Curtis Mussen	5/2/17	1444	256F25	10/17	100	102	P	LAB
CCV									

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Curtis Mussen	5/2/17	715	16111B	5/18	7.0	7.00	-18	P	LAB
Calibr.	Curtis Mussen	5/2/17	717	170310A	3/18	4.0	4.01	158	P	LAB
Calibr.	Curtis Mussen	5/2/17	719	16111C	5-18	10.01	9.97	-181	P	LAB
ICV	Curtis Mussen	5/2/17	720	16111C	5/18	10.01	9.92	-178	P	LAB
CCV	Curtis Mussen	5/2/17	1446	170310A	3-18	3.87	3.87	164	P	LAB
CCV						4.01				

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater