

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter Bio #2 RQ- 2017-082127 Project SMP

- Notes:** (1) Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up (e.g., 5.15 becomes 5.2).
 (2) Always wait for meter to stabilize before recording any readings.
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____

Temperature (Quarterly) FT 1400 Date of Last Temperature Verification 03/06/2017

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Curtis Musson	8/22/17	1708	23.90	8.434	8.43	100.0	S1	—	P	LAD
ICV	Curtis Musson	8/22/17	1710	23.93	8.434	8.44	100.1	S1	—	P	LAD
CCV	Curtis Musson	8/23/17	1348	25.43	8.203	8.23	100.3	S2	—	P	LAD
CCV											

Report DO mg/L with one decimal figure and DO % saturation as a whole number with no decimals.

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Curtis Musson	8/22/17	1712	170310B	09/2018	7.00	7.00	P	LAD
ICV	Curtis Musson	8/22/17	1721	170310D	3/2018	1000	1000	P	LAD
CCV	Curtis Musson	8/22/17	1723	170310D	3/2018	1000	1001	P	LAD
CCV	Curtis Musson	8/23/17	1352	170614D	6/2018	100	104	P	LAD

Report specific conductance as a whole number with no decimal figure.

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Curtis Musson	8/22/17	1712	170310B	09/2018	7.0	7.00	40	P	LAD
Calibr.	Curtis Musson	8/22/17	1715	170614A	06/2018	4.01	4.00	201	P	LAD
Calibr.	Curtis Musson	8/22/17	1717	170310C	9/2018	10.01	9.99	-115	P	LAD
ICV	Curtis Musson	8/22/17	1719	170614A	06/2018	4.01	3.81	210	P	LAD
CCV	Curtis Musson	8/23/17	1353	170310C	9/2018	10.01	10.01	-116	P	LAD
CCV										

Report pH with one decimal place; pH Acceptance criteria ± 0.2 SU; mV pH7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ; Slope from 7 to 10 and 4 to 7 must be between 165 and 180 mV

Depth (Quarterly)

Date of Last Depth Verification _____

Depth Sensor (Daily)	Name	Date	Time CT-ET	Zero the Sensor	ICV Value	Pass / Fail	Lab / Field
Pressure mode in air				0.000			

ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater