

Swamp SCE

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: Bio #2

RQ: 2018-12-18-34

Project: SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Curtis Mussen	12/20/17	738	20.78	8.950	8.95	100.0	64.5	-	P/F	D/F
ICV	Curtis Mussen	12/20/17	746	20.82	8.950	8.94	100.0	64.5	-	P/F	D/F
CCV	Curtis Mussen	12/20/17	1211	19.86	9.111	9.24	101.4	65	-	P/F	D/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Curtis Mussen	12/20/17	748	17100SD	10/2018	1000	1000	P/F	D/F
ICV	Curtis Mussen	12/20/17	750	17100SD	10/2018	1000	1000	P/F	D/F
CCV	Curtis Mussen	12/20/17	1213	17100SE	10/2018	100	102	P/F	D/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Curtis Mussen	12/20/17	752	170614B	12/2018	7.0	7.00	22	P/F	D/F
Calibr.	Curtis Mussen	12/20/17	753	17100SA	10/2018	4.01	4.01	190	P/F	D/F
Calibr.	Curtis Mussen	12/20/17	754	17100SC	4/2019	10.01	9.98	-139	P/F	D/F
ICV	Curtis Mussen	12/20/17	756	17100SA	10/2018	4.01	3.96	192	P/F	D/F
CCV	Curtis Mussen	12/20/17	1215	170614B	12/2018	7.0	7.06	22	P/F	D/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? ☒ Yes / No If yes, list date of last quarterly depth verification _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air	Curtis Mussen	12/20/17	750	0.00	0.00	P/F	D/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: