

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: DEP#154181 RQ: 2018-10-2208 Project: SUWANNEE

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 10-10-18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	P O'Connor	10-25-18	0717	23.1	8.56	8.55	100.0	—	—	<u>P</u> /F	<u>L</u> /F
ICV	V	V	0720	23.3	8.53	8.54	100.1	—	—	<u>P</u> /F	<u>L</u> /F
CCV	P O'Connor	10-25-18	1350	22.1	8.12	8.68	99.1	—	—	<u>P</u> /F	<u>L</u> /F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Explr. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	P O'Connor	10-25-18	0725	180711A	07-2019	1000	1000	<u>P</u> /F	<u>L</u> /F
ICV	V	V	0727	180321A	04-2019	100	101.8	<u>P</u> /F	<u>L</u> /F
CCV	P O'Connor	10-25-18	1355	1410718	12-2018	147	153	<u>P</u> /F	<u>L</u> /F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Explr. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	P O'Connor	10-25-18	0729	2803E45	03-2020	7.0	7.00	-29.3	<u>P</u> /F	<u>L</u> /F
Calibr.			0732	2806975	06-2020	4.0	4.00	153.5	<u>P</u> /F	<u>L</u> /F
Calibr.			0735	2712855	06-2019	10.0	9.97	-196.6	<u>P</u> /F	<u>P</u> /F
ICV	V	V	0740	2803E45	03-2020	7.00	7.07	-28.2	<u>P</u> /F	<u>P</u> /F
CCV	P O'Connor	10-25-18	1558	2806975	06-2020	4.00	4.03	150.0	<u>P</u> /F	<u>L</u> /F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification 10-10-18

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

BP QC/Scanned 11/5/18 JP 10/29/18