

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: **YSI Pro DSS**

RQ: **2018-10-29-26**

Project: **LSJR East 1**

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification _____

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	AKalmbacher	11/7/2018	0825	23.0	759.5	8.57	8.58	100.0			P / F	L / F
ICV			0828	23.0	759.6	8.57	8.57	99.9			(P) F	(L) F
CCV	AKalmbacher	11/7/2018	1330	24.6	757.4	8.32	8.29	99.7			(P) F	(L) F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	AKalmbacher	11/7/2018	0835	1014138	9/2019	1413	1413	P / F	L / F
ICV				18032A	4/2019	100	100-1	(P) F	(L) F
CCV	AKalmbacher	11/7/2018	1333	1014138	9/2019	1413	1418	(P) F	(L) F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	AKalmbacher	11/7/2018	0840	2803E45	3/2020	7.	23.7	7.06	-20.6	P / F	L / F
Calibr.				2806975	6/2020	4.	23.9	3.96	153.9	P / F	L / F
Calibr.				2712855	6/2019	10.	24.0	10.01	198.4	P / F	L / F
ICV				2803E45	3/2020	7.0	23.9	7.08	-28.8	(P) F	(L) F
CCV	AKalmbacher	11/7/2018	1337	2803E45		10.0	24.1	10.09	-203.2	(P) F	(L) F
CCV				2712855	6/2019					P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: _____

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

BP Scanned/QC 11/15/18 JP 11/8/18