

Chemical Analysis Report

NE-DIST-2018-01-17-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Suwannee North**
Request ID: **RQ-2018-01-15-24**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Jacksonville, FL 32256-7577
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator Date Certified: 05-FEB-2018 15:53



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Field Blank

Collection Date/Time: 01/16/2018 11:15

Field ID: BLANK01162017

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1959852	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P338370	
	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P338474	
		Sulfate	0.20	U	mg SO4/L	P338476	
	SM 2120 B	Color (true)	2.5	U	PCU	P338399	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P338674	
	SM 2540 C-97	TDS	15	U	mg/L	P338695	
	SM 2540 D-97	TSS	2	U	mg/L	P338895	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P339050	
1959856	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P338724	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P338490	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P338570	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P338426	
	SM 5310 B-00	Organic Carbon	0.50	U	mg C/L	P338527	
1959860	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P338372	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 4500 F-C-97: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Jerry Branch @ SE CR 25A

Collection Date/Time: 01/16/2018 11:10

Field ID: 21020136

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1959853	EPA 180.1 Rev. 2.0	Turbidity	2.8		NTU	P338370	
	EPA 300.0 Rev. 2.1	Chloride	11		mg Cl/L	P338474	
		Sulfate	2.6		mg SO4/L	P338476	
	SM 2120 B	Color (true)	310		PCU	P338399	
	SM 2320 B-97	Total Alkalinity	1.3	I	mg CaCO3/L	P338674	
	SM 2540 C-97	TDS	109		mg/L	P338697	
	SM 2540 D-97	TSS	2	I	mg/L	P338897	
	SM 4500 F-C-97	Fluoride	0.30		mg F/L	P339050	
1959857	EPA 350.1 Rev. 2.0	Ammonia-N	0.012		mg N/L	P338724	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.89		mg N/L	P338490	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.007	I	mg N/L	P338570	
	EPA 365.1 Rev. 2.0	Total-P	0.37		mg P/L	P338426	
	SM 5310 B-00	Organic Carbon	33		mg C/L	P338527	
1959861	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.33		mg P/L	P338372	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 4500 F-C-97: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Cannon Creek @ Cannon Creek Rd.

Collection Date/Time: 01/16/2018 12:05

Field ID: 21030138

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1959864	EPA 180.1 Rev. 2.0	Turbidity	2.7		NTU	P338370	

Field ID: 21030138

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1959864	EPA 300.0 Rev. 2.1	Chloride	14		mg Cl/L	P338475	
		Sulfate	2.9		mg SO4/L	P338477	
	SM 2120 B	Color (true)	66		PCU	P338399	
	SM 2320 B-97	Total Alkalinity	53		mg CaCO3/L	P338674	
	SM 2540 C-97	TDS	106		mg/L	P338697	
	SM 2540 D-97	TSS	2	I	mg/L	P338897	
	SM 4500 F-C-97	Fluoride	0.060	I	mg F/L	P339050	
1959865	EPA 350.1 Rev. 2.0	Ammonia-N	0.006		mg N/L	P338725	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.38		mg N/L	P33892	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P338570	
	EPA 365.1 Rev. 2.0	Total-P	0.025		mg P/L	P338425	
	SM 5310 B-00	Organic Carbon	10		mg C/L	P338527	
1959866	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.011		mg P/L	P338375	
1959894	EPA 8321B	Sucralose**	0.25		ug/L	P338431	
	USGS O-2060-01	2,4-D	0.0020	U	ug/L	P338349	
		Diuron	0.0020	U	ug/L	P338349	
		Fenuron	0.0080	U	ug/L	P338349	
		Imidacloprid	0.012		ug/L	P338349	MS
		Bentazon	0.0018	I	ug/L	P338349	
		Linuron	0.0040	U	ug/L	P338349	
		Acetaminophen**	0.0080	U	ug/L	P338349	
		MCP	0.0020	U	ug/L	P338349	
		Carbamazepine**	4.0E-04	U	ug/L	P338349	
		Triclopyr**	0.0040	U	ug/L	P338349	
		Primidone**	0.0040	U	ug/L	P338349	
		Fluridone**	0.0036		ug/L	P338349	RPD

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 4500 F-C-97: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Rocky CR. 136

Collection Date/Time: 01/16/2018 10:30

Field ID: 21010016

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1959854	EPA 180.1 Rev. 2.0	Turbidity	2.4		NTU	P338370	
	EPA 300.0 Rev. 2.1	Chloride	12		mg Cl/L	P338474	
		Sulfate	1.7		mg SO4/L	P338476	
	SM 2120 B	Color (true)	270		PCU	P338399	
	SM 2320 B-97	Total Alkalinity	7.5		mg CaCO3/L	P338674	
	SM 2540 C-97	TDS	89		mg/L	P338697	
	SM 2540 D-97	TSS	3	I	mg/L	P338897	
1959858	SM 4500 F-C-97	Fluoride	0.16		mg F/L	P339050	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.01		mg N/L	P338724	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.75		mg N/L	P338490	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.021		mg N/L	P338570	

Field ID: 21010016

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1959858	EPA 365.1 Rev. 2.0	Total-P	0.32		mg P/L	P338426	
	SM 5310 B-00	Organic Carbon	22		mg C/L	P338527	
1959862	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.28		mg P/L	P338375	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 4500 F-C-97: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Peacock Lake Center

Collection Date/Time: 01/16/2018 09:40

Field ID: 21020097

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1959855	EPA 180.1 Rev. 2.0	Turbidity	5.0		NTU	P338370	
	EPA 300.0 Rev. 2.1	Chloride	7.3		mg Cl/L	P338474	
		Sulfate	3.7		mg SO4/L	P338476	
	SM 2120 B	Color (true)	73	A	PCU	P338399	
	SM 2320 B-97	Total Alkalinity	10		mg CaCO3/L	P338674	
	SM 2540 C-97	TDS	55		mg/L	P338697	
	SM 2540 D-97	TSS	7	I	mg/L	P338897	
	SM 4500 F-C-97	Fluoride	0.11		mg F/L	P339050	
1959859	EPA 350.1 Rev. 2.0	Ammonia-N	0.23		mg N/L	P338724	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P338490	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.051		mg N/L	P338570	
	EPA 365.1 Rev. 2.0	Total-P	0.33		mg P/L	P338426	
	SM 5310 B-00	Organic Carbon	8.7		mg C/L	P338527	
1959863	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.23		mg P/L	P338376	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 4500 F-C-97: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P338370

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P338474

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P338475

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338476

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338477

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P338724

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P338725

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P338490

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P33892

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P338570

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P338425

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P338426

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P338372

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P338375

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P338376

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8321B
Batch ID: P338431

Component	Result	Code	Units
Sucralose	0.010	U	ug/L

Reference Method: SM 2120 B
Batch ID: P338399

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P338674

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P338695

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-97
Batch ID: P338697

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P338895

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report Method Blank Results

Reference Method: SM 2540 D-97
Batch ID: P338897

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P339050

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P338527

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: USGS O-2060-01
Batch ID: P338349

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
Acetaminophen	0.0080	U	ug/L
Bentazon	8.0E-04	U	ug/L
Carbamazepine	4.0E-04	U	ug/L
Diuron	0.0020	U	ug/L
Fenuron	0.0080	U	ug/L
Fluridone	4.0E-04	U	ug/L
Imidacloprid	0.0020	U	ug/L
Linuron	0.0040	U	ug/L
MCP	0.0020	U	ug/L
Primidone	0.0040	U	ug/L
Triclopyr	0.0040	U	ug/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338474

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338475

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338476

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	94.4		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338477

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	95.2		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P338724

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P338725

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	95.6		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P338490

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P33892

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	98.9		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P338570

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO ₂ NO ₃ -N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P338425

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P338426

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P338372

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.7		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P338375

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P338376

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.5		P	90 - 110

Reference Method: EPA 8321B
 Batch ID: P338431

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sucralose	102		P	40 - 150

Reference Method: SM 2320 B-97
 Batch ID: P338674

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	98.8		P	95 - 108

Reference Method: SM 4500 F-C-97
 Batch ID: P339050

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	95.4		P	92 - 103

Reference Method: SM 5310 B-00
 Batch ID: P338527

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	93.2		P	85 - 115

Reference Method: USGS O-2060-01
 Batch ID: P338349

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	75.3		P	40 - 150
Acetaminophen	98.5		P	30 - 150
Bentazon	78.6		P	30 - 150
Carbamazepine	114		P	40 - 150
Diuron	102		P	40 - 150
Fenuron	114		P	40 - 150
Fluridone	120		P	40 - 150
Imidacloprid	113		P	40 - 160
Linuron	110		P	40 - 150
MCPP	79.5		P	40 - 150
Primidone	109		P	40 - 150
Triclopyr	79.5		P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338474

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959559	Chloride	96.3		P	90 - 110
1959560	Chloride	96.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338475

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959939	Chloride	90.6		P	90 - 110
1960003	Chloride	96.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338476

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959559	Sulfate	94.0		P	90 - 110
1959560	Sulfate	94.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338477

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959939	Sulfate	91.0		P	90 - 110
1960003	Sulfate	93.2		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P338724

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959728	Ammonia-N	98.9	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P338725

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959865	Ammonia-N	95.1	93.9	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P338490

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959857	Kjeldahl Nitrogen	106	96.9	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P33892

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959865	Kjeldahl Nitrogen	104	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P338570

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P338570

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959801	NO2NO3-N	103	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P338425

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959865	Total-P	101	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P338426

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959931	Total-P	99.0	98.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P338372

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959528	O-Phosphate-P	97.0	97.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P338375

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959945	O-Phosphate-P	100	100	P/P	90 - 110

Reference Method: EPA 8321B
Batch ID: P338431

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959537	Sucralose	87.4	97.6	P/P	40 - 150

Reference Method: SM 5310 B-00
Batch ID: P338527

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959801	Organic Carbon	92.9	92.0	P/P	85 - 115

Reference Method: USGS O-2060-01
Batch ID: P338349

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959579	2,4-D	113	103	P/P	40 - 150
1959579	Acetaminophen	87.9	109	P/P	30 - 150
1959579	Bentazon	148	105	P/P	30 - 150
1959579	Carbamazepine	100	105	P/P	40 - 150
1959579	Diuron	74.3	85.4	P/P	40 - 150
1959579	Fenuron	107	93.4	P/P	40 - 150
1959579	Fluridone	83.5	47.8	P/P	40 - 150
1959579	Imidacloprid	161	223	F/F	40 - 160
1959579	Linuron	87.7	87.0	P/P	40 - 150
1959579	MCP	97.1	75.7	P/P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: USGS O-2060-01
Batch ID: P338349

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1959579	Primidone	121	96.4	P/P	40 - 150
1959579	Triclopyr	90.9	75.1	P/P	40 - 150

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P338370

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959800	Turbidity	1.59	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338474

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959559	Chloride	0.667	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338475

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1960003	Chloride	0.134	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P338476

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959559	Sulfate	0.663	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P338724

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959728	Ammonia-N	1.56	Spike	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P338725

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959865	Ammonia-N	1.17	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P338490

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959857	Kjeldahl Nitrogen	5.39	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P33892

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959865	Kjeldahl Nitrogen	2.67	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P338570

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959801	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P338425

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959865	Total-P	0.561	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P338426

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959931	Total-P	0.188	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P338372

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959528	O-Phosphate-P	0.617	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P338375

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959945	O-Phosphate-P	0.300	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P338376

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959863	O-Phosphate-P	0.233	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P338431

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959537	Sucralose	11.0	Spike	P	0 - 30

Reference Method: SM 2120 B
Batch ID: P338399

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959855	Color (true)	0.0648	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P338674

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1960002	Total Alkalinity	1.01	Sample	P	0 - 4.8

Reference Method: SM 2540 C-97
Batch ID: P338695

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959526	TDS	2.63	Sample	P	0 - 10

Reference Method: SM 2540 C-97
Batch ID: P338697

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959665	TDS	2.82	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P339050

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1960002	Fluoride	1.96	Spike	P	0 - 3.0

Reference Method: SM 5310 B-00
Batch ID: P338527

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959801	Organic Carbon	0.583	Spike	P	0 - 20

Reference Method: USGS O-2060-01
Batch ID: P338349

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959579	2,4-D	8.07	Spike	P	0 - 30
1959579	Acetaminophen	21.2	Spike	P	0 - 30

Quality Assurance Report Precision

Reference Method: USGS O-2060-01
 Batch ID: P338349

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1959579	Bentazon	12.6	Spike	P	0 - 30
1959579	Carbamazepine	4.22	Spike	P	0 - 30
1959579	Diuron	12.2	Spike	P	0 - 30
1959579	Fenuron	13.3	Spike	P	0 - 30
1959579	Fluridone	36.3	Spike	F	0 - 30
1959579	Imidacloprid	28.3	Spike	P	0 - 30
1959579	Linuron	0.843	Spike	P	0 - 30
1959579	MCP	23.6	Spike	P	0 - 30
1959579	Primidone	22.8	Spike	P	0 - 30
1959579	Triclopyr	19.0	Spike	P	0 - 30

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
 Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 1959894
 Field Sample ID: 21030138

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	Acetaminophen-d4	96.5	P	30 - 160
EPA 8321B	Carbamazepine-d10	103	P	30 - 160
EPA 8321B	Diuron-d6	87.3	P	30 - 160
EPA 8321B	Primidone-d5	102	P	30 - 160
EPA 8321B	Sucralose-d6	104	P	40 - 160
USGS O-2060-01	2,4-D-d3	86.9	P	30 - 160
USGS O-2060-01	Acetaminophen-d4	96.5	P	30 - 160
USGS O-2060-01	Carbamazepine-d10	103	P	30 - 160
USGS O-2060-01	Diuron-d6	87.3	P	30 - 160
USGS O-2060-01	Primidone-d5	102	P	30 - 160
USGS O-2060-01	Sucralose-d6	104	P	40 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 180.1 Rev. 2.0
 Run ID: A81498
 Included Lab Sample IDs: 1959852, 1959853, 1959854, 1959855, 1959864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	99.0	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Run ID: A81499
 Included Lab Sample IDs: 1959860, 1959861, 1959862, 1959863, 1959866

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.2	P/P	90 - 110
O-Phosphate-P	98.8	99.5	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A81499

Included Lab Sample IDs: 1959860, 1959861, 1959862, 1959863, 1959866

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.2	99.6	P/P	90 - 110
O-Phosphate-P	99.6	98.8	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A81514

Included Lab Sample IDs: 1959852, 1959853, 1959854, 1959855, 1959864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	98.9	P/P	85 - 115
Color (true)	98.6	100	P/P	85 - 115
Color (true)	99.5	98.6	P/P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A81524

Included Lab Sample IDs: 1959852, 1959853, 1959854, 1959855, 1959864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	103	107	P/P	90 - 110
Chloride	105	103	P/P	90 - 110
Sulfate	100	106	P/P	90 - 110
Sulfate	105	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A81551

Included Lab Sample IDs: 1959857, 1959858, 1959859

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	98.2	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A81556

Included Lab Sample IDs: 1959856, 1959865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	99.8	P/P	90 - 110
Total-P	99.8	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A81564

Included Lab Sample IDs: 1959856, 1959857, 1959858, 1959859, 1959865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.2	95.3	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A81579

Included Lab Sample IDs: 1959856, 1959857, 1959858, 1959859, 1959865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A81585

Included Lab Sample IDs: 1959856, 1959857, 1959858, 1959859, 1959865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	95.1	96.1	P/P	90 - 110
Kjeldahl Nitrogen	96.1	98.4	P/P	90 - 110
Kjeldahl Nitrogen	98.9	100	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A81591

Included Lab Sample IDs: 1959894

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	104	101	P/P	60 - 150

Reference Method: SM 2320 B-97

Run ID: A81614

Included Lab Sample IDs: 1959852, 1959853, 1959854, 1959855, 1959864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	97.1	99.0	P/P	90 - 110

Reference Method: USGS O-2060-01

Run ID: A81623

Included Lab Sample IDs: 1959894

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	103	83.9	P/P	50 - 150
Acetaminophen	107	113	P/P	60 - 150
Bentazon	111	97.5	P/P	50 - 150
Carbamazepine	118	120	P/P	60 - 150
Diuron	113	114	P/P	60 - 150
Fenuron	109	109	P/P	60 - 150
Fluridone	114	118	P/P	60 - 160
Imidacloprid	112	112	P/P	60 - 150
Linuron	107	113	P/P	60 - 150
MCPP	103	101	P/P	50 - 150
Primidone	106	102	P/P	60 - 150
Triclopyr	109	88.0	P/P	50 - 150

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A81629

Included Lab Sample IDs: 1959856, 1959857, 1959858, 1959859, 1959865

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	91.4	92.2	P/P	90 - 110
Ammonia-N	97.0	95.6	P/P	90 - 110
Ammonia-N	98.9	97.0	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A81747

Included Lab Sample IDs: 1959852, 1959853, 1959854, 1959855, 1959864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision		
					LCS	SMP	MS
EPA 180.1 Rev. 2.0	Turbidity					1.59	
EPA 300.0 Rev. 2.1	Chloride	98.5	96.3	96.8		0.667	
	Chloride	98.4	96.7	90.6		0.134	
	Sulfate	94.4	94.0	94.6		0.663	
	Sulfate	95.2	93.2	91.0			
	Sulfate	95.2	93.2	91.0			
EPA 350.1 Rev. 2.0	Ammonia-N	99.9	98.9	101			1.56
	Ammonia-N	95.6	95.1	93.9			1.17
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	102	106	96.9			5.39
	Kjeldahl Nitrogen	98.9	104	100			2.67
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	104	103	103			0.0
EPA 365.1 Rev. 2.0	Total-P	103	101	101			0.561
	Total-P	102	99.0	98.8			0.188
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.7	97.0	97.6			0.617
	O-Phosphate-P	98.8	100	100			0.300
	O-Phosphate-P	99.5					0.233
EPA 8321B	Sucralose	102	87.4	97.6			11.0
SM 2120 B	Color (true)					0.0648	
SM 2320 B-97	Total Alkalinity	98.8				1.01	
SM 2540 C-97	TDS					2.63	
	TDS					2.82	
SM 4500 F-C-97	Fluoride	95.4					1.96
SM 5310 B-00	Organic Carbon	93.2	92.9	92.0			0.583
USGS O-2060-01	2,4-D	75.3	113	103			8.07
	Acetaminophen	98.5	87.9	109			21.2
	Bentazon	78.6	148	105			12.6
	Carbamazepine	114	100	105			4.22
	Diuron	102	74.3	85.4			12.2
	Fenuron	114	107	93.4			13.3
	Fluridone	120	83.5	47.8			36.3
	Imidacloprid	113	161	223			28.3
	Linuron	110	87.7	87.0			0.843
	MCPP	79.5	97.1	75.7			23.6
	Primidone	109	121	96.4			22.8
	Triclopyr	79.5	90.9	75.1			19.0

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1959852, 1959853, 1959854, 1959855, 1959864
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1959852, 1959853, 1959854, 1959855, 1959864
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1959852, 1959853, 1959854, 1959855, 1959864
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1959856, 1959857, 1959858, 1959859, 1959865

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1959856, 1959857, 1959858, 1959859, 1959865
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1959856, 1959857, 1959858, 1959859, 1959865
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1959856, 1959857, 1959858, 1959859, 1959865
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1959860, 1959861, 1959862, 1959863, 1959866
EPA 8321B / E31780	Analysis of sucralose in water matrices by HPLC/MS/MS	1959894
SM 2120 B / E31780	True Color measured at 450 nm	1959852, 1959853, 1959854, 1959855, 1959864
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	1959852, 1959853, 1959854, 1959855, 1959864
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1959852, 1959853, 1959854, 1959855, 1959864
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1959852, 1959853, 1959854, 1959855, 1959864
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1959852, 1959853, 1959854, 1959855, 1959864
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1959856, 1959857, 1959858, 1959859, 1959865
USGS O-2060-01 / E31780	Chlorinated (phenoxy acid) herbicides in water matrices by HPLC/MS/MS	1959894
USGS O-2060-01 / E31780	Urea herbicides and other chemicals in water matrices by HPLC/MS/MS	1959894

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	01/17/2018			01/17/2018 16:19	DeAsia Armster	1959852, 1959853, 1959854, 1959855, 1959864
EPA 300.0 Rev. 2.1	01/17/2018			01/18/2018	Julia Storbeck	1959852, 1959853, 1959854, 1959855, 1959864
EPA 350.1 Rev. 2.0	01/17/2018			01/24/2018	Ping Hua	1959856, 1959857, 1959858, 1959859, 1959865
EPA 351.2 Rev. 2.0	01/17/2018	01/19/2018	Adrian Shelby	01/22/2018	Joseph Suarez	1959856, 1959857, 1959858, 1959859, 1959865
EPA 353.2 Rev. 2.0	01/17/2018			01/22/2018	Lauren Bottorf	1959856, 1959857, 1959858, 1959859, 1959865
EPA 365.1 Rev. 2.0	01/17/2018	01/18/2018	Sima Feizi	01/19/2018	Lipi Saha	1959856, 1959857, 1959858, 1959859, 1959865
EPA 365.1 Rev. 2.0 dissolved	01/17/2018			01/17/2018 16:13	Megan Treadwell	1959860
	01/17/2018			01/17/2018 16:27	Megan Treadwell	1959863
	01/17/2018			01/17/2018 16:36	Megan Treadwell	1959866
	01/17/2018			01/17/2018 16:47	Megan Treadwell	1959861
	01/17/2018			01/17/2018 16:48	Megan Treadwell	1959862
EPA 8321B	01/17/2018	01/19/2018	Juyoung Kim	01/20/2018	Juyoung Kim	1959894
SM 2120 B	01/17/2018			01/17/2018 17:09	Tanya Welborn	1959852
	01/17/2018			01/17/2018 17:12	Tanya Welborn	1959855
	01/17/2018			01/17/2018 17:13	Tanya Welborn	1959864
	01/17/2018			01/17/2018 17:26	Tanya Welborn	1959853
	01/17/2018			01/17/2018 17:29	Tanya Welborn	1959854
SM 2320 B-97	01/17/2018			01/23/2018	Dale L. Simmons	1959852, 1959853, 1959854, 1959855, 1959864
SM 2540 C-97	01/17/2018			01/19/2018	Yijie Li	1959852, 1959853, 1959854, 1959855, 1959864
SM 2540 D-97	01/17/2018			01/19/2018	Yijie Li	1959852, 1959853, 1959854, 1959855, 1959864
SM 4500 F-C-97	01/17/2018			01/30/2018	Lauren Bottorf	1959852, 1959853, 1959854, 1959855, 1959864
SM 5310 B-00	01/17/2018			01/20/2018	Megan Treadwell	1959856, 1959857, 1959858, 1959859, 1959865
USGS O-2060-01	01/17/2018	01/19/2018	Hoor Shaik	01/20/2018	Julie Michelle Frank	1959894