

Chemical Analysis Report

WQAP-2018-07-27-06

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **2018 Continuous Monitoring**
Request ID: **RQ-2018-07-23-26**
Customer: **WQAP**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
MS #3525
Tallahassee, FL 32399-2400
Attn: Curtis Musson

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 21-AUG-2018 15:20

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Craig Cr @ Hendricks Ave

Collection Date/Time: 07/26/2018 10:40

Field ID: 20030793

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2011556	EPA 180.1 Rev. 2.0	Turbidity	6.4		NTU	P349326	
	EPA 300.0 Rev. 2.1	Chloride	18		mg Cl/L	P349592	
		Sulfate	34		mg SO4/L	P349595	
	SM 2120 B	Color (true)	89	A	PCU	P349300	
	SM 2320 B-97	Total Alkalinity	148	A	mg CaCO3/L	P349484	
	SM 2540 C-97	TDS	247		mg/L	P349675	
	SM 2540 D-97	TSS	3	I	mg/L	P349368	
	SM 4500 F-C-97	Fluoride	0.16		mg F/L	P349488	
2011557	EPA 350.1 Rev. 2.0	Ammonia-N	0.13		mg N/L	P349571	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.73		mg N/L	P349446	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.26		mg N/L	P349405	
	EPA 365.1 Rev. 2.0	Total-P	0.076		mg P/L	P349514	
	SM 5310 B-00	Organic Carbon	12		mg C/L	P349690	
2011558	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.045		mg P/L	P349303	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Craig Cr @ Hendricks Ave

Collection Date/Time: 07/26/2018 10:40

Field ID: 20030793

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2011566	EPA 8321B	2,4-D	0.0030	I	ug/L	P349361	MS
		Sucralose**	0.31		ug/L	P349535	
		Bentazon	0.014		ug/L	P349361	
		MCPD	0.0020	U	ug/L	P349361	
		Triclopyr**	0.0040	U	ug/L	P349361	
		Imidacloprid	0.029		ug/L	P349361	
		Diuron	0.0070	I	ug/L	P349361	
		Pyraclostrobin**	0.0020	U	ug/L	P349361	
		Primidone**	0.0040	U	ug/L	P349361	
		Linuron	0.0040	U	ug/L	P349361	
		Acetaminophen**	0.0080	U	ug/L	P349361	
		Fenuron	0.016	U	ug/L	P349361	
		Imazapyr**	0.13		ug/L	P349361	
		Carbamazepine**	8.0E-04	U	ug/L	P349361	
		Fluridone**	0.0093		ug/L	P349361	
		Hydrocodone**	8.0E-04	U	ug/L	P349361	
		Naproxen**	0.0080	U	ug/L	P349361	
		Ibuprofen**	0.020	U	ug/L	P349361	

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349326

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349592

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349595

Component	Result	Code	Units
Sulfate	0.20	U	mg SO ₄ /L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P349571

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P349446

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P349405

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P349514

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349303

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8321B
Batch ID: P349361

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
Acetaminophen	0.0080	U	ug/L
Bentazon	8.0E-04	U	ug/L
Carbamazepine	8.0E-04	U	ug/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 8321B
Batch ID: P349361

Component	Result	Code	Units
Diuron	0.0020	U	ug/L
Fenuron	0.016	U	ug/L
Fluridone	4.0E-04	U	ug/L
Hydrocodone	8.0E-04	U	ug/L
Ibuprofen	0.020	U	ug/L
Imazapyr	0.0040	U	ug/L
Imidacloprid	0.0020	U	ug/L
Linuron	0.0040	U	ug/L
MCPP	0.0020	U	ug/L
Naproxen	0.0080	U	ug/L
Primidone	0.0040	U	ug/L
Pyraclostrobin	0.0020	U	ug/L
Triclopyr	0.0040	U	ug/L

Reference Method: EPA 8321B
Batch ID: P349535

Component	Result	Code	Units
Sucralose	0.010	U	ug/L

Reference Method: SM 2120 B
Batch ID: P349300

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P349484

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P349675

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P349368

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P349488

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report Method Blank Results

Reference Method: SM 5310 B-00
Batch ID: P349690

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P349592

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.1		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P349595

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.2		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P349571

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P349446

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	99.1		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P349405

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P349514

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	95.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P349303

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.1		P	90 - 110

Reference Method: EPA 8321B
 Batch ID: P349361

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	126		P	40 - 150
Acetaminophen	85.6		P	30 - 150
Bentazon	88.1		P	30 - 150
Carbamazepine	115		P	40 - 150
Diuron	90.5		P	40 - 150
Fenuron	105		P	40 - 150
Fluridone	105		P	40 - 150
Hydrocodone	98.0		P	40 - 150
Ibuprofen	112		P	40 - 150
Imazapyr	108		P	40 - 150

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P349361

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Imidacloprid	98.7		P	40 - 160
Linuron	80.2		P	40 - 150
MCPP	126		P	40 - 150
Naproxen	120		P	40 - 150
Primidone	115		P	40 - 150
Pyraclostrobin	86.9		P	40 - 150
Triclopyr	133		P	40 - 150

Reference Method: EPA 8321B
Batch ID: P349535

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sucralose	114		P	40 - 150

Reference Method: SM 2120 B
Batch ID: P349300

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	98.1		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P349484

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P349488

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.6		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P349690

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	93.9		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P349592

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011483	Chloride	102		P	90 - 110
2011594	Chloride	108		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P349595

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011483	Sulfate	99.8		P	90 - 110
2011594	Sulfate	104		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P349571

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011573	Ammonia-N	100	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P349446

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011557	Kjeldahl Nitrogen	100	97.4	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P349405

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011557	NO2NO3-N	104		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P349514

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011557	Total-P	101	92.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P349303

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011490	O-Phosphate-P	95.9	95.2	P/P	90 - 110

Reference Method: EPA 8321B
 Batch ID: P349361

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011588	2,4-D	124	153	P/F	40 - 150
2011588	Acetaminophen	81.1	88.4	P/P	30 - 150
2011588	Bentazon	46.6	45.3	P/P	30 - 150
2011588	Carbamazepine	71.8	68.5	P/P	40 - 150
2011588	Diuron	60.6	47.7	P/P	40 - 150
2011588	Fenuron	87.3	85.3	P/P	40 - 150
2011588	Fluridone	93.2	74.2	P/P	40 - 150
2011588	Hydrocodone	81.9	81.0	P/P	40 - 150

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 8321B
Batch ID: P349361

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011588	Ibuprofen	73.6	74.9	P/P	40 - 150
2011588	Imazapyr	101	88.9	P/P	40 - 150
2011588	Imidacloprid	134	149	P/P	40 - 160
2011588	Linuron	48.6	59.1	P/P	40 - 150
2011588	MCPP	121	119	P/P	40 - 150
2011588	Naproxen	103	96.9	P/P	40 - 150
2011588	Primidone	87.6	82.0	P/P	40 - 150
2011588	Pyraclostrobin	75.3	75.1	P/P	40 - 150
2011588	Triclopyr	121	121	P/P	40 - 150

Reference Method: EPA 8321B
Batch ID: P349535

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011995	Sucralose	111	109	P/P	40 - 150

Reference Method: SM 4500 F-C-97
Batch ID: P349488

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011556	Fluoride	96.9	97.7	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P349690

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2011652	Organic Carbon	96.0	95.3	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P349326

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011482	Turbidity	4.15	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349592

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011483	Chloride	4.04	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P349595

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011483	Sulfate	3.88	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P349571

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011573	Ammonia-N	0.163	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P349446

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011557	Kjeldahl Nitrogen	1.91	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P349405

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011557	NO2NO3-N	0.433	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P349514

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011557	Total-P	7.70	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P349303

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011490	O-Phosphate-P	0.733	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P349361

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011588	2,4-D	14.7	Spike	P	0 - 30
2011588	Acetaminophen	8.52	Spike	P	0 - 30
2011588	Bentazon	2.31	Spike	P	0 - 30
2011588	Carbamazepine	4.61	Spike	P	0 - 30
2011588	Diuron	21.9	Spike	P	0 - 30
2011588	Fenuron	2.34	Spike	P	0 - 30
2011588	Fluridone	9.41	Spike	P	0 - 30
2011588	Hydrocodone	1.06	Spike	P	0 - 30
2011588	Ibuprofen	1.74	Spike	P	0 - 30
2011588	Imazapyr	4.03	Spike	P	0 - 30
2011588	Imidacloprid	6.51	Spike	P	0 - 30
2011588	Linuron	19.4	Spike	P	0 - 30
2011588	MCPP	1.76	Spike	P	0 - 30
2011588	Naproxen	6.02	Spike	P	0 - 30
2011588	Primidone	6.60	Spike	P	0 - 30
2011588	Pyraclostrobin	0.342	Spike	P	0 - 30
2011588	Triclopyr	0.157	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P349535

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011995	Sucralose	1.55	Spike	P	0 - 30

Reference Method: SM 2120 B
Batch ID: P349300

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011556	Color (true)	0.254	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P349484

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011556	Total Alkalinity	0.661	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P349675

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2010788	TDS	7.23	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P349488

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: SM 4500 F-C-97
Batch ID: P349488

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011556	Fluoride	0.520	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P349690

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2011652	Organic Carbon	0.565	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 2011566
Field Sample ID: 20030793

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	114	P	30 - 160
EPA 8321B	Acetaminophen-d4	96.2	P	30 - 160
EPA 8321B	Carbamazepine-d10	66.3	P	30 - 160
EPA 8321B	Diuron-d6	53.5	P	30 - 160
EPA 8321B	Ibuprofen-d3	91.4	P	30 - 160
EPA 8321B	Primidone-d5	87.7	P	30 - 160
EPA 8321B	Sucralose-d6	88.0	P	40 - 160

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A85528
Included Lab Sample IDs: 2011556

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.4	98.7	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A85530
Included Lab Sample IDs: 2011558

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.0	98.2	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A85537
Included Lab Sample IDs: 2011556

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	102	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A85559
Included Lab Sample IDs: 2011557

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A85560
Included Lab Sample IDs: 2011556

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.9	102	P/P	90 - 110
Sulfate	100	103	P/P	90 - 110

Reference Method: SM 2320 B-97
Run ID: A85607
Included Lab Sample IDs: 2011556

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	103	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Run ID: A85607
Included Lab Sample IDs: 2011556

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.5	99.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A85630
Included Lab Sample IDs: 2011557

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	94.7	95.2	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A85647

Included Lab Sample IDs: 2011557

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	96.6	99.4	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A85648

Included Lab Sample IDs: 2011557

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	101	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A85704

Included Lab Sample IDs: 2011557

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.1	98.5	P/P	90 - 110

Reference Method: EPA 8321B

Run ID: A85787

Included Lab Sample IDs: 2011566

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	99.0	106	P/P	60 - 160

Reference Method: EPA 8321B

Run ID: A85801

Included Lab Sample IDs: 2011566

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetaminophen	114	113	P/P	60 - 160
Carbamazepine	109	101	P/P	60 - 160
Diuron	97.0	92.0	P/P	60 - 150
Fenuron	104	101	P/P	60 - 150
Fluridone	103	100	P/P	60 - 160
Hydrocodone	93.6	96.3	P/P	60 - 150
Imazapyr	94.8	87.1	P/P	50 - 150
Imidacloprid	96.2	95.8	P/P	60 - 150
Linuron	97.4	94.2	P/P	60 - 150
Primidone	97.8	110	P/P	60 - 150
Pyraclostrobin	90.0	85.0	P/P	60 - 150

Reference Method: EPA 8321B

Run ID: A85823

Included Lab Sample IDs: 2011566

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	111	146	P/P	50 - 150
Bentazon	99.3	101	P/P	50 - 160
Ibuprofen	108	119	P/P	50 - 160
MCP	107	115	P/P	50 - 150
Naproxen	90.3	101	P/P	50 - 160
Triclopyr	121	133	P/P	50 - 160

Quality Assurance Report Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					4.15	
EPA 300.0 Rev. 2.1	Chloride	97.1	102	108		4.04	
	Sulfate	97.2	99.8	104		3.88	
EPA 350.1 Rev. 2.0	Ammonia-N	101	100	101			0.163
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	99.1	100	97.4			1.91
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	104				0.433
EPA 365.1 Rev. 2.0	Total-P	95.8	101	92.0			7.70
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.1	95.9	95.2			0.733
EPA 8321B	2,4-D	126	124	153			14.7
	Acetaminophen	85.6	81.1	88.4			8.52
	Bentazon	88.1	46.6	45.3			2.31
	Carbamazepine	115	71.8	68.5			4.61
	Diuron	90.5	60.6	47.7			21.9
	Fenuron	105	87.3	85.3			2.34
	Fluridone	105	93.2	74.2			9.41
	Hydrocodone	98.0	81.9	81.0			1.06
	Ibuprofen	112	73.6	74.9			1.74
	Imazapyr	108	101	88.9			4.03
	Imidacloprid	98.7	134	149			6.51
	Linuron	80.2	48.6	59.1			19.4
	MCPP	126	121	119			1.76
	Naproxen	120	103	96.9			6.02
	Primidone	115	87.6	82.0			6.60
	Pyraclostrobin	86.9	75.3	75.1			0.342
	Sucralose	114	111	109			1.55
	Triclopyr	133	121	121			0.157
SM 2120 B	Color (true)	98.1				0.254	
SM 2320 B-97	Total Alkalinity	104				0.661	
SM 2540 C-97	TDS					7.23	
SM 4500 F-C-97	Fluoride	97.6	96.9	97.7			0.520
SM 5310 B-00	Organic Carbon	93.9	96.0	95.3			0.565

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2011556
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2011556
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2011556
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2011557
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2011557
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2011557
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2011557
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2011558
EPA 8321B / E31780	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS	2011566
EPA 8321B / E31780	Water soluble organic compounds by HPLC/MS/MS	2011566
SM 2120 B / E31780	True Color measured at 450 nm	2011556
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2011556

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2011556
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2011556
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2011556
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2011557

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	07/27/2018			07/27/2018 16:39	Tanya Welborn	2011556
EPA 300.0 Rev. 2.1	07/27/2018			08/02/2018 20:06	Lipi Saha	2011556
EPA 350.1 Rev. 2.0	07/27/2018			08/02/2018 11:15	Ping Hua	2011557
EPA 351.2 Rev. 2.0	07/27/2018	07/31/2018 11:15	Adrian Shelby	08/02/2018 13:12	Joseph Suarez	2011557
EPA 353.2 Rev. 2.0	07/27/2018			07/30/2018 13:18	Lauren Bottorf	2011557
EPA 365.1 Rev. 2.0	07/27/2018	08/01/2018 12:35	Sima Feizi	08/02/2018 09:28	Beverly Y. Sanders	2011557
EPA 365.1 Rev. 2.0 dissolved	07/27/2018			07/27/2018 13:47	Julia Storbeck	2011558
EPA 8321B	07/27/2018	08/02/2018 09:00	Hoor Shaik	08/10/2018 03:58	Juyoung Kim	2011566
	07/27/2018	08/02/2018 09:00	Hoor Shaik	08/10/2018 20:29	Juyoung Kim	2011566
	07/27/2018	08/03/2018 09:00	Mohammad Ghaffari	08/09/2018 03:28	Mohammad Ghaffari	2011566
SM 2120 B	07/27/2018			07/27/2018 11:49	Tanya Welborn	2011556
SM 2320 B-97	07/27/2018			08/01/2018 00:55	Dale L. Simmons	2011556
SM 2540 C-97	07/27/2018			07/28/2018 20:08	Yijie Li	2011556
SM 2540 D-97	07/27/2018			07/28/2018 23:02	Yijie Li	2011556
SM 4500 F-C-97	07/27/2018			08/01/2018 00:47	Dale L. Simmons	2011556
SM 5310 B-00	07/27/2018			08/03/2018 23:24	Julia Storbeck	2011557