

Chemical Analysis Report

NE-DIST-2018-10-18-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **LSJR West2**
Request ID: **RQ-2018-10-15-09**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256-7577
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 21-NOV-2018 11:33

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light-colored rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: YELLOW WATER CREEK AT STATE FOREST

Collection Date/Time: 10/17/2018 10:10

Field ID: G2NE1163

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2033855	EPA 180.1 Rev. 2.0	Turbidity	55		NTU	P353460	
	EPA 300.0 Rev. 2.1	Chloride	8.8		mg Cl/L	P353721	
		Sulfate	5.0		mg SO4/L	P353723	
		Color (true)	97		PCU	P353303	
	SM 2120 B	Total Alkalinity	27		mg CaCO3/L	P353711	
	SM 2540 C-97	TDS	71		mg/L	P353838	
	SM 2540 D-97	TSS	10		mg/L	P353587	
	SM 4500 F-C-97	Fluoride	0.045	I	mg F/L	P353714	
2033860	EPA 350.1 Rev. 2.0	Ammonia-N	0.016		mg N/L	P353862	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.54		mg N/L	P353582	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.077		mg N/L	P353353	
	EPA 365.1 Rev. 2.0	Total-P	0.061		mg P/L	P353342	
	SM 5310 B-00	Organic Carbon	13		mg C/L	P353622	
2033865	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.011		mg P/L	P353313	
2033888	EPA 200.7 Rev. 4.4	Calcium	11.7		mg/L	P353462	
		Iron	1.27E+03		ug/L	P353462	
		Magnesium	1.79		mg/L	P353462	
		Potassium	0.69	I	mg/L	P353462	
		Sodium	6.2		mg/L	P353462	
	EPA 200.8 Rev. 5.4	Zinc	5.0	U	ug/L	P353462	
		Arsenic	0.72		ug/L	P353462	
		Cadmium	0.028	I	ug/L	P353462	
		Chromium	1.9		ug/L	P353462	
		Copper	0.85		ug/L	P353462	
		Lead	1.54		ug/L	P353462	
		Nickel	0.59	I	ug/L	P353462	
		Silver	0.010	U	ug/L	P353462	

Ref. Method and Comment:
 EPA 200.8 Rev. 5.4: The lead result was confirmed on 11/15/2018.

Sample Location: YELLOW WATER CR AT SR 228

Collection Date/Time: 10/17/2018 10:45

Field ID: 20030549

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2033856	EPA 180.1 Rev. 2.0	Turbidity	5.8		NTU	P353459	
	EPA 300.0 Rev. 2.1	Chloride	12		mg Cl/L	P353721	
		Sulfate	1.2		mg SO4/L	P353723	
	SM 2120 B	Color (true)	260		PCU	P353297	
	SM 2320 B-97	Total Alkalinity	15		mg CaCO3/L	P353711	
	SM 2540 C-97	TDS	80		mg/L	P353838	
	SM 2540 D-97	TSS	4	I	mg/L	P353587	
	SM 4500 F-C-97	Fluoride	0.048	I	mg F/L	P353714	
2033861	EPA 350.1 Rev. 2.0	Ammonia-N	0.024		mg N/L	P353862	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.80		mg N/L	P353582	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.062		mg N/L	P353353	
	EPA 365.1 Rev. 2.0	Total-P	0.053		mg P/L	P353342	
	SM 5310 B-00	Organic Carbon	22		mg C/L	P353622	
2033866	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.020		mg P/L	P353313	
2033889	EPA 200.7 Rev. 4.4	Calcium	8.40		mg/L	P353462	
		Iron	1.77E+03		ug/L	P353462	
		Magnesium	1.80		mg/L	P353462	
		Potassium	0.48	I	mg/L	P353462	
		Sodium	6.9		mg/L	P353462	
		Zinc	5.9	I	ug/L	P353462	
	EPA 200.8 Rev. 5.4	Arsenic	0.64		ug/L	P353462	
		Cadmium	0.020	U	ug/L	P353462	
		Chromium	0.92	I	ug/L	P353462	
		Copper	0.81		ug/L	P353462	
		Lead	1.03		ug/L	P353462	
		Nickel	0.75	I	ug/L	P353462	
		Silver	0.010	U	ug/L	P353462	

Ref. Method and Comment:
 EPA 200.8 Rev. 5.4: The lead result was confirmed on 11/15/2018.

Sample Location: YELLOW WATER CR AT BICENTENNIAL

Collection Date/Time: 10/17/2018 11:10

Field ID: G2NE1153

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2033857	EPA 180.1 Rev. 2.0	Turbidity	7.8	A	NTU	P353459	
	EPA 300.0 Rev. 2.1	Chloride	10		mg Cl/L	P353721	
		Sulfate	0.69		mg SO4/L	P353723	
		Color (true)	370		PCU	P353297	
	SM 2320 B-97	Total Alkalinity	13		mg CaCO3/L	P353711	
	SM 2540 C-97	TDS	96		mg/L	P353838	
	SM 2540 D-97	TSS	6	I	mg/L	P353587	
	SM 4500 F-C-97	Fluoride	0.049	I	mg F/L	P353714	
2033862	EPA 350.1 Rev. 2.0	Ammonia-N	0.040		mg N/L	P353862	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.1		mg N/L	P353582	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.051		mg N/L	P353353	
	EPA 365.1 Rev. 2.0	Total-P	0.065		mg P/L	P353342	
	SM 5310 B-00	Organic Carbon	27		mg C/L	P353622	
2033867	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.020		mg P/L	P353313	
2033890	EPA 200.7 Rev. 4.4	Calcium	7.71		mg/L	P353462	
		Iron	2.13E+03		ug/L	P353462	
		Magnesium	1.58		mg/L	P353462	
		Potassium	0.48	I	mg/L	P353462	
		Sodium	5.9		mg/L	P353462	
		Zinc	7.4	I	ug/L	P353462	
	EPA 200.8 Rev. 5.4	Arsenic	0.64		ug/L	P353462	
		Cadmium	0.020	U	ug/L	P353462	
		Chromium	0.97	I	ug/L	P353462	
		Copper	0.76	I	ug/L	P353462	
		Lead	0.89		ug/L	P353462	
		Nickel	0.82		ug/L	P353462	
		Silver	0.010	U	ug/L	P353462	

Ref. Method and Comment:

EPA 200.8 Rev. 5.4: The lead result was confirmed in the undigested sample on 11/08/2018.

Sample Location: MCGIRTS CREEK AT OLD PLANK ROAD

Collection Date/Time: 10/17/2018 09:30

Field ID: 20030346

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2033858	EPA 180.1 Rev. 2.0	Turbidity	9.1		NTU	P353459	
	EPA 300.0 Rev. 2.1	Chloride	16		mg Cl/L	P353721	
		Sulfate	8.3		mg SO4/L	P353723	
		Color (true)	92		PCU	P353297	
	SM 2120 B	Total Alkalinity	40		mg CaCO3/L	P353711	
	SM 2540 C-97	TDS	106	A	mg/L	P353838	
	SM 2540 D-97	TSS	19	A	mg/L	P353587	
	SM 4500 F-C-97	Fluoride	0.044	I	mg F/L	P353714	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.031		mg N/L	P353862	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.36		mg N/L	P353582	
2033863	EPA 353.2 Rev. 2.0	NO2NO3-N	0.071		mg N/L	P353353	
	EPA 365.1 Rev. 2.0	Total-P	0.034		mg P/L	P353342	
	SM 5310 B-00	Organic Carbon	8.7		mg C/L	P353622	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.015		mg P/L	P353313	
	dissolved						
2033891	EPA 200.7 Rev. 4.4	Calcium	20.0		mg/L	P353462	
		Iron	4.09E+03		ug/L	P353462	
		Magnesium	1.58		mg/L	P353462	
		Potassium	1.1	I	mg/L	P353462	
		Sodium	9.3		mg/L	P353462	
		Zinc	10	I	ug/L	P353462	
	EPA 200.8 Rev. 5.4	Arsenic	0.86		ug/L	P353462	
		Cadmium	0.021	I	ug/L	P353462	
		Chromium	1.1	I	ug/L	P353462	
		Copper	1.08		ug/L	P353462	
		Lead	14.5		ug/L	P353462	
		Nickel	0.70	I	ug/L	P353462	
		Silver	0.010	U	ug/L	P353462	

Ref. Method and Comment:

EPA 200.8 Rev. 5.4: The lead result was confirmed on 11/15/2018.

Sample Location: McGIRTS CREEK AT HELSEMA RD

Collection Date/Time: 10/17/2018 09:10

Field ID: 20030877

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2033859	EPA 180.1 Rev. 2.0	Turbidity	4.4		NTU	P353459	
	EPA 300.0 Rev. 2.1	Chloride	9.4	A	mg Cl/L	P353721	
		Sulfate	8.0	A	mg SO4/L	P353723	
		Color (true)	46	A	PCU	P353298	
	SM 2120 B	Total Alkalinity	41		mg CaCO3/L	P353711	
	SM 2540 C-97	TDS	84		mg/L	P353838	
	SM 2540 D-97	TSS	3	I	mg/L	P353587	
	SM 4500 F-C-97	Fluoride	0.041	I	mg F/L	P353714	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.040		mg N/L	P353862	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.24		mg N/L	P353582	
2033864	EPA 353.2 Rev. 2.0	NO2NO3-N	0.038		mg N/L	P353353	
	EPA 365.1 Rev. 2.0	Total-P	0.029		mg P/L	P353342	
	SM 5310 B-00	Organic Carbon	5.1		mg C/L	P353622	
	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.012		mg P/L	P353313	
	EPA 200.7 Rev. 4.4	Calcium	18.1		mg/L	P353462	
2033892		Iron	610		ug/L	P353462	
		Magnesium	1.13		mg/L	P353462	
		Potassium	0.55	I	mg/L	P353462	
		Sodium	5.8		mg/L	P353462	
		Zinc	5.0	U	ug/L	P353462	
		Arsenic	0.25		ug/L	P353462	
		Cadmium	0.020	U	ug/L	P353462	
		Chromium	0.40	U	ug/L	P353462	
		Copper	0.20	U	ug/L	P353462	
		Lead	1.01		ug/L	P353462	
	EPA 200.8 Rev. 5.4	Nickel	0.20	U	ug/L	P353462	
		Silver	0.010	U	ug/L	P353462	

Sample Location: blank-101718

Collection Date/Time: 10/17/2018 10:20

Field ID: Field Blank

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2033870	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P353459	
	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P353721	
		Sulfate	0.20	U	mg SO4/L	P353723	
		Color (true)	2.5	U	PCU	P353298	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P353711	
	SM 2540 C-97	TDS	15	U	mg/L	P353834	
	SM 2540 D-97	TSS	2	U	mg/L	P353583	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P353714	
2033871	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P353862	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P353606	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P353353	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P353342	
	SM 5310 B-00	Organic Carbon	0.50	U	mg C/L	P353622	
2033872	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P353313	
2033893	EPA 200.7 Rev. 4.4	Calcium	0.075	U	mg/L	P353462	
		Iron	30	U	ug/L	P353462	
		Magnesium	0.040	U	mg/L	P353462	
		Potassium	0.30	U	mg/L	P353462	
		Sodium	0.50	U	mg/L	P353462	
		Zinc	5.0	U	ug/L	P353462	
		Arsenic	0.050	U	ug/L	P353462	
	EPA 200.8 Rev. 5.4	Cadmium	0.020	U	ug/L	P353462	
		Chromium	0.40	U	ug/L	P353462	
		Copper	0.20	U	ug/L	P353462	
		Lead	0.050	U	ug/L	P353462	
		Nickel	0.20	U	ug/L	P353462	
		Silver	0.010	U	ug/L	P353462	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P353459

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P353460

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P353462

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P353462

Component	Result	Code	Units
Arsenic	0.050	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.20	U	ug/L
Lead	0.050	U	ug/L
Nickel	0.20	U	ug/L
Silver	0.010	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P353721

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P353723

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353862

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353582

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353606

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P353353

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P353342

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P353313

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P353297

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2120 B
Batch ID: P353298

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2120 B
Batch ID: P353303

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P353711

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 C-97
Batch ID: P353834

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-97
Batch ID: P353838

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P353583

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P353587

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P353714

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P353622

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P353462

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	103		P	85 - 115
Iron	103		P	85 - 115
Magnesium	104		P	85 - 115
Potassium	98.6		P	85 - 115
Sodium	106		P	85 - 115
Zinc	92.4		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P353462

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Arsenic	102		P	85 - 115
Cadmium	101		P	85 - 115
Chromium	101		P	85 - 115
Copper	102		P	85 - 115
Lead	94.9		P	85 - 115
Nickel	102		P	85 - 115
Silver	98.7		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P353721

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P353723

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	99.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353862

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	100		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353582

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	105		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353606

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P353353

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P353342

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P353313

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	102		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P353297

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2120 B
Batch ID: P353298

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2120 B
Batch ID: P353303

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P353711

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P353714

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.9		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P353622

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P353462

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033892	Calcium	99.8		P	80 - 120
2033892	Iron	104		P	80 - 120
2033892	Magnesium	104		P	80 - 120
2033892	Potassium	101		P	80 - 120
2033892	Sodium	95.1		P	80 - 120
2033892	Zinc	91.2		P	80 - 120
2033902	Calcium	99.7	102	P/P	80 - 120
2033902	Iron	102	101	P/P	80 - 120
2033902	Magnesium	101	99.8	P/P	80 - 120
2033902	Potassium	97.3	95.1	P/P	80 - 120
2033902	Sodium	97.6	94.4	P/P	80 - 120
2033902	Zinc	91.9	92.7	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P353462

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033892	Arsenic	102		P	80 - 120
2033892	Cadmium	103		P	80 - 120
2033892	Chromium	103		P	80 - 120
2033892	Copper	101		P	80 - 120
2033892	Lead	95.8		P	80 - 120
2033892	Nickel	101		P	80 - 120
2033892	Silver	99.6		P	80 - 120
2033902	Arsenic	102	101	P/P	80 - 120
2033902	Cadmium	101	98.9	P/P	80 - 120
2033902	Chromium	101	100	P/P	80 - 120
2033902	Copper	103	103	P/P	80 - 120
2033902	Lead	94.2	94.2	P/P	80 - 120
2033902	Nickel	101	101	P/P	80 - 120
2033902	Silver	99.0	98.3	P/P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P353721

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033859	Chloride	102		P	90 - 110
2033971	Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P353723

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033859	Sulfate	101		P	90 - 110
2033971	Sulfate	102		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353862

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033864	Ammonia-N	98.7	100	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353582

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033810	Kjeldahl Nitrogen	98.6	98.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353606

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033879	Kjeldahl Nitrogen	104	106	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P353353

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033863	NO2NO3-N	99.6	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P353342

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033864	Total-P	100	98.3	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P353313

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033785	O-Phosphate-P	97.4	97.7	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P353714

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033920	Fluoride	97.6	97.6	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P353622

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2033531	Organic Carbon	95.2	94.9	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P353459

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033857	Turbidity	6.62	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P353460

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033972	Turbidity	10.7	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P353462

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033902	Calcium	1.64	Spike	P	0 - 20
2033902	Iron	0.566	Spike	P	0 - 20
2033902	Magnesium	0.856	Spike	P	0 - 20
2033902	Potassium	2.14	Spike	P	0 - 20
2033902	Sodium	1.85	Spike	P	0 - 20
2033902	Zinc	0.876	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P353462

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033902	Arsenic	0.588	Spike	P	0 - 20
2033902	Cadmium	1.75	Spike	P	0 - 20
2033902	Chromium	0.815	Spike	P	0 - 20
2033902	Copper	0.0989	Spike	P	0 - 20
2033902	Lead	0.0603	Spike	P	0 - 20
2033902	Nickel	0.133	Spike	P	0 - 20
2033902	Silver	0.647	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P353721

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033859	Chloride	0.142	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P353723

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033859	Sulfate	0.486	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P353862

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033864	Ammonia-N	0.771	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353582

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033810	Kjeldahl Nitrogen	0.0855	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P353606

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033879	Kjeldahl Nitrogen	1.98	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P353353

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033863	NO2NO3-N	0.738	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P353342

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033864	Total-P	2.00	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P353313

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033785	O-Phosphate-P	0.308	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P353297

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033805	Color (true)	1.95	Sample	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P353298

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033859	Color (true)	2.17	Sample	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
Batch ID: P353303

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033804	Color (true)	6.90	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P353711

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033920	Total Alkalinity	0.631	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P353834

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033096	TDS	3.65	Sample	P	0 - 10

Reference Method: SM 2540 C-97
Batch ID: P353838

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033858	TDS	8.45	Sample	P	0 - 10

Reference Method: SM 2540 D-97
Batch ID: P353587

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033858	TSS	6.19	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P353714

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033920	Fluoride	0.0	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P353622

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2033531	Organic Carbon	0.154	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A87127

Included Lab Sample IDs: 2033855, 2033856, 2033857, 2033858, 2033859, 2033870

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	101	P/P	90 - 115
Color (true)	101	102	P/P	90 - 115
Color (true)	103	102	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A87131

Included Lab Sample IDs: 2033865, 2033866, 2033867, 2033868, 2033869, 2033872

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.1	98.6	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A87160

Included Lab Sample IDs: 2033860, 2033861, 2033862, 2033863, 2033864, 2033871

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A87192

Included Lab Sample IDs: 2033860, 2033861, 2033862

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A87195

Included Lab Sample IDs: 2033863, 2033864

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	100	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A87200

Included Lab Sample IDs: 2033855, 2033856, 2033857, 2033858, 2033859, 2033870

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	101	101	P/P	95 - 105
Turbidity	102	101	P/P	95 - 105

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A87212

Included Lab Sample IDs: 2033855, 2033856, 2033857, 2033858, 2033859, 2033870

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Chloride	99.9	101	P/P	90 - 110
Sulfate	100	101	P/P	90 - 110
Sulfate	99.2	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A87220

Included Lab Sample IDs: 2033871

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	101	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A87260

Included Lab Sample IDs: 2033860, 2033861, 2033862, 2033863, 2033864, 2033871

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.4	96.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A87281

Included Lab Sample IDs: 2033860, 2033861, 2033862, 2033863, 2033864, 2033871

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	95.4	97.5	P/P	90 - 110
Kjeldahl Nitrogen	97.0	95.4	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A87299

Included Lab Sample IDs: 2033855, 2033856, 2033857, 2033858, 2033859, 2033870

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A87299

Included Lab Sample IDs: 2033855, 2033856, 2033857, 2033858, 2033859, 2033870

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	100	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A87337

Included Lab Sample IDs: 2033888, 2033889, 2033890, 2033891, 2033892, 2033893

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Potassium	92.2	92.2	P/P	90 - 110
Potassium	94.3	92.2	P/P	90 - 110
Sodium	94.0	94.9	P/P	90 - 110
Sodium	98.1	94.0	P/P	90 - 110
Zinc	92.8	92.8	P/P	90 - 110
Zinc	95.0	92.8	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A87365

Included Lab Sample IDs: 2033888, 2033889, 2033890, 2033891, 2033892, 2033893

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	102	99.2	P/P	90 - 110
Calcium	99.2	98.1	P/P	90 - 110
Iron	102	99.9	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A87365

Included Lab Sample IDs: 2033888, 2033889, 2033890, 2033891, 2033892, 2033893

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Iron	99.9	99.6	P/P	90 - 110
Magnesium	97.7	97.0	P/P	90 - 110
Magnesium	99.5	97.7	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A87366

Included Lab Sample IDs: 2033860, 2033861, 2033862, 2033863, 2033864, 2033871

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	103	P/P	90 - 110
Ammonia-N	103	101	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A87611

Included Lab Sample IDs: 2033888, 2033889, 2033890, 2033891, 2033892, 2033893

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Arsenic	97.6	97.6	P/P	90 - 110
Arsenic	99.2	97.6	P/P	90 - 110
Cadmium	96.0	98.3	P/P	90 - 110
Cadmium	98.2	96.0	P/P	90 - 110
Chromium	98.6	99.7	P/P	90 - 110
Chromium	99.7	100	P/P	90 - 110
Copper	100	98.9	P/P	90 - 110
Copper	98.9	99.1	P/P	90 - 110
Lead	93.7	93.6	P/P	90 - 110
Lead	94.1	93.7	P/P	90 - 110
Nickel	100	99.9	P/P	90 - 110
Nickel	99.9	99.9	P/P	90 - 110
Silver	92.9	93.8	P/P	90 - 110
Silver	93.8	92.9	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision	
						LCS	SMP MS
EPA 180.1 Rev. 2.0	Turbidity						6.62
	Turbidity						10.7
EPA 200.7 Rev. 4.4	Calcium	103	99.8	99.7	102		1.64
	Iron	103	104	102	101		0.566
	Magnesium	104	104	101	99.8		0.856
	Potassium	98.6	101	97.3	95.1		2.14
	Sodium	106	95.1	97.6	94.4		1.85
	Zinc	92.4	91.2	91.9	92.7		0.876
EPA 200.8 Rev. 5.4	Arsenic	102	102	102	101		0.588
	Cadmium	101	103	101	98.9		1.75
	Chromium	101	103	101	100		0.815
	Copper	102	101	103	103		0.0989
	Lead	94.9	95.8	94.2	94.2		0.0603
	Nickel	102	101	101	101		0.133
	Silver	98.7	99.6	99.0	98.3		0.647
EPA 300.0 Rev. 2.1	Chloride	100	102	102			0.142
	Sulfate	99.7	101	102			0.486
EPA 350.1 Rev. 2.0	Ammonia-N	100	98.7	100			0.771
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	105	98.6	98.5			0.0855
	Kjeldahl Nitrogen	103	104	106			1.98
EPA 353.2 Rev. 2.0	NO2NO3-N	100	99.6	101			0.738
EPA 365.1 Rev. 2.0	Total-P	102	100	98.3			2.00
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	102	97.4	97.7			0.308
SM 2120 B	Color (true)	101					1.95
	Color (true)	101					2.17
	Color (true)	101					6.90
SM 2320 B-97	Total Alkalinity	103					0.631
SM 2540 C-97	TDS						3.65
	TDS						8.45
SM 2540 D-97	TSS						6.19
SM 4500 F-C-97	Fluoride	98.9	97.6	97.6			0.0
SM 5310 B-00	Organic Carbon	96.8	95.2	94.9			0.154

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2033888, 2033889, 2033890, 2033891, 2033892, 2033893
EPA 200.8 Rev. 5.4 / E31780	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2033888, 2033889, 2033890, 2033891, 2033892, 2033893
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2033860, 2033861, 2033862, 2033863, 2033864, 2033871
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2033860, 2033861, 2033862, 2033863, 2033864, 2033871
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2033860, 2033861, 2033862, 2033863, 2033864, 2033871
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2033860, 2033861, 2033862, 2033863, 2033864, 2033871

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 365.1 Rev. 2.0 dissolved / E31780 SM 2120 B / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L True Color measured at 450 nm	2033865, 2033866, 2033867, 2033868, 2033869, 2033872 2033855, 2033856, 2033857, 2033858, 2033859, 2033870
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2033860, 2033861, 2033862, 2033863, 2033864, 2033871

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	10/18/2018			10/18/2018 15:59	Megan Treadwell	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
EPA 200.7 Rev. 4.4	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/26/2018 22:03	Betina Topolski	2033893
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/26/2018 22:39	Betina Topolski	2033888
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/26/2018 23:02	Betina Topolski	2033889
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/26/2018 23:07	Betina Topolski	2033890
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/26/2018 23:12	Betina Topolski	2033891
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/26/2018 23:17	Betina Topolski	2033892
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/29/2018 20:16	Betina Topolski	2033893
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/29/2018 21:22	Betina Topolski	2033888
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/29/2018 21:29	Betina Topolski	2033889
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/29/2018 21:36	Betina Topolski	2033890
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/29/2018 21:43	Betina Topolski	2033891
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	10/29/2018 21:50	Betina Topolski	2033892
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	11/07/2018 19:08	Robert K Palmer	2033893
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	11/07/2018 21:12	Robert K Palmer	2033888
EPA 200.8 Rev. 5.4	10/18/2018	10/22/2018 14:45	Elliott D. Healy	11/07/2018 21:21	Robert K Palmer	2033889
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	11/07/2018 21:31	Robert K Palmer	2033890
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	11/07/2018 21:40	Robert K Palmer	2033891
	10/18/2018	10/22/2018 14:45	Elliott D. Healy	11/07/2018 21:50	Robert K Palmer	2033892
	10/18/2018					
EPA 300.0 Rev. 2.1	10/18/2018			10/24/2018 22:19	Lipi Saha	2033859
	10/18/2018			10/24/2018 23:44	Lipi Saha	2033855
	10/18/2018			10/24/2018 23:56	Lipi Saha	2033856
	10/18/2018			10/25/2018 00:32	Lipi Saha	2033857
	10/18/2018			10/25/2018 00:44	Lipi Saha	2033858
	10/18/2018			10/25/2018 00:56	Lipi Saha	2033870
EPA 350.1 Rev. 2.0	10/18/2018			10/29/2018 12:55	Ping Hua	2033864
	10/18/2018			10/29/2018 12:59	Ping Hua	2033861
	10/18/2018			10/29/2018 13:07	Ping Hua	2033862
	10/18/2018			10/29/2018 13:08	Ping Hua	2033863
	10/18/2018			10/29/2018 13:10	Ping Hua	2033860
	10/18/2018			10/29/2018 13:16	Ping Hua	2033871
EPA 351.2 Rev. 2.0	10/18/2018	10/24/2018 11:00	Adrian Shelby	10/25/2018 11:01	Joseph Suarez	2033860
	10/18/2018	10/24/2018 11:00	Adrian Shelby	10/25/2018 11:02	Joseph Suarez	2033861
	10/18/2018	10/24/2018 11:00	Adrian Shelby	10/25/2018 11:04	Joseph Suarez	2033862
	10/18/2018	10/24/2018 11:00	Adrian Shelby	10/25/2018 11:05	Joseph Suarez	2033863
	10/18/2018	10/24/2018 11:00	Adrian Shelby	10/25/2018 11:07	Joseph Suarez	2033864
	10/18/2018	10/24/2018 11:00	Adrian Shelby	10/25/2018 11:15	Joseph Suarez	2033871
EPA 353.2 Rev. 2.0	10/18/2018			10/19/2018 13:15	Lauren Bottorf	2033863
	10/18/2018			10/19/2018 13:22	Lauren Bottorf	2033871

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 353.2 Rev. 2.0	10/18/2018			10/19/2018 13:24	Lauren Bottorf	2033861
	10/18/2018			10/19/2018 13:26	Lauren Bottorf	2033862
	10/18/2018			10/19/2018 13:27	Lauren Bottorf	2033860
	10/18/2018			10/19/2018 13:28	Lauren Bottorf	2033864
EPA 365.1 Rev. 2.0	10/18/2018	10/19/2018 12:10	Sima Feizi	10/22/2018 10:16	Beverly Y. Sanders	2033860
	10/18/2018	10/19/2018 12:10	Sima Feizi	10/22/2018 10:17	Beverly Y. Sanders	2033861
	10/18/2018	10/19/2018 12:10	Sima Feizi	10/22/2018 10:18	Beverly Y. Sanders	2033862
	10/18/2018	10/19/2018 12:10	Sima Feizi	10/22/2018 12:06	Beverly Y. Sanders	2033863
	10/18/2018	10/19/2018 12:10	Sima Feizi	10/22/2018 12:07	Beverly Y. Sanders	2033864
	10/18/2018	10/19/2018 12:10	Sima Feizi	10/23/2018 11:42	Beverly Y. Sanders	2033871
EPA 365.1 Rev. 2.0 dissolved	10/18/2018			10/18/2018 17:36	Anthony C. Onicha	2033865
	10/18/2018			10/18/2018 17:37	Anthony C. Onicha	2033866
	10/18/2018			10/18/2018 17:38	Anthony C. Onicha	2033867
	10/18/2018			10/18/2018 17:39	Anthony C. Onicha	2033868
	10/18/2018			10/18/2018 17:40	Anthony C. Onicha	2033869
	10/18/2018			10/18/2018 17:41	Anthony C. Onicha	2033872
SM 2120 B	10/18/2018			10/18/2018 15:38	Chelsea Hernandez	2033858
	10/18/2018			10/18/2018 15:42	Chelsea Hernandez	2033859
	10/18/2018			10/18/2018 15:43	Chelsea Hernandez	2033870
	10/18/2018			10/18/2018 16:06	Chelsea Hernandez	2033855
	10/18/2018			10/18/2018 16:07	Chelsea Hernandez	2033856
	10/18/2018			10/18/2018 16:08	Chelsea Hernandez	2033857
SM 2320 B-97	10/18/2018			10/24/2018 12:57	Dale L. Simmons	2033855
	10/18/2018			10/24/2018 13:08	Dale L. Simmons	2033856
	10/18/2018			10/24/2018 13:18	Dale L. Simmons	2033857
	10/18/2018			10/24/2018 13:29	Dale L. Simmons	2033858
	10/18/2018			10/24/2018 13:39	Dale L. Simmons	2033859
	10/18/2018			10/24/2018 13:53	Dale L. Simmons	2033870
SM 2540 C-97	10/18/2018			10/19/2018 15:48	Yijie Li	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
SM 2540 D-97	10/18/2018			10/19/2018 15:48	Yijie Li	2033855, 2033856, 2033857, 2033858, 2033859, 2033870
SM 4500 F-C-97	10/18/2018			10/24/2018 12:57	Dale L. Simmons	2033855
	10/18/2018			10/24/2018 13:08	Dale L. Simmons	2033856
	10/18/2018			10/24/2018 13:18	Dale L. Simmons	2033857
	10/18/2018			10/24/2018 13:29	Dale L. Simmons	2033858
	10/18/2018			10/24/2018 13:39	Dale L. Simmons	2033859
	10/18/2018			10/24/2018 13:53	Dale L. Simmons	2033870
SM 5310 B-00	10/18/2018			10/24/2018 07:17	Megan Treadwell	2033860
	10/18/2018			10/24/2018 07:30	Megan Treadwell	2033861

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 5310 B-00	10/18/2018			10/24/2018 07:44	Megan Treadwell	2033862
	10/18/2018			10/24/2018 08:00	Megan Treadwell	2033863
	10/18/2018			10/24/2018 08:17	Megan Treadwell	2033864
	10/18/2018			10/24/2018 08:30	Megan Treadwell	2033871