

Chemical Analysis Report

NE-DIST-2018-12-20-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **unnamed 2156**
Request ID: **RQ-2018-12-17-08**
Customer: **NE-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
8800 Baymeadows Way West
Suite 100
Jacksonville, FL 32256
Attn: Jeremy Parrish

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 08-JAN-2019 16:25

A handwritten signature in black ink, appearing to read "Liang Lin", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: UNNAMED BRANCH ABOVE ALLIGATOR CREEK

Collection Date/Time: 12/19/2018 11:25

Field ID: G4NE0288

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2050260	EPA 180.1 Rev. 2.0	Turbidity	3.1		NTU	P356502	
	EPA 300.0 Rev. 2.1	Chloride	31		mg Cl/L	P356721	
		Sulfate	54		mg SO4/L	P356723	
		Color (true)	310		PCU	P356474	
	SM 2320 B-97	Total Alkalinity	75		mg CaCO3/L	P356661	
	SM 2540 C-97	TDS	283		mg/L	P356692	
	SM 2540 D-97	TSS	2	U	mg/L	P356584	
	SM 4500 F-C-97	Fluoride	0.10		mg F/L	P356662	
2050261	EPA 350.1 Rev. 2.0	Ammonia-N	0.017		mg N/L	P356623	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.4		mg N/L	P356642	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.028		mg N/L	P356514	
	EPA 365.1 Rev. 2.0	Total-P	0.27		mg P/L	P356680	
	SM 5310 B-00	Organic Carbon	34		mg C/L	P356718	
2050262	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.21		mg P/L	P356473	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P356502

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P356721

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P356723

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P356623

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P356642

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P356514

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P356680

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P356473

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P356474

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P356661

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P356692

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P356584

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P356662

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P356718

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P356721

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.3		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P356723

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P356623

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	100		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P356642

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	106		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P356514

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P356680

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P356473

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.2		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P356474

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	102		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P356661

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	98 - 107

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P356662

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.3		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P356718

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P356721

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2050014	Chloride	96.6		P	90 - 110
2050221	Chloride	94.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P356723

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2050014	Sulfate	96.0		P	90 - 110
2050221	Sulfate	91.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P356623

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2050061	Ammonia-N	103	98.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P356642

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2050281	Kjeldahl Nitrogen	115	109	F/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P356514

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2050222	NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P356680

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2050298	Total-P	98.2	99.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P356473

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2050062	O-Phosphate-P	93.2	95.0	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P356662

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2050297	Fluoride	98.7	98.7	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P356718

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2050119	Organic Carbon	92.0	92.9	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P356502

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050118	Turbidity	3.08	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P356721

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050221	Chloride	1.45	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P356723

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050221	Sulfate	1.36	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P356623

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050061	Ammonia-N	3.32	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P356642

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050281	Kjeldahl Nitrogen	3.96	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P356514

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050222	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P356680

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050298	Total-P	1.49	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P356473

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050062	O-Phosphate-P	1.71	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
Batch ID: P356474

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050278	Color (true)	0.312	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P356661

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050297	Total Alkalinity	0.165	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P356692

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050149	TDS	2.95	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P356662

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050297	Fluoride	0.0	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P356718

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2050119	Organic Carbon	0.548	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A88469

Included Lab Sample IDs: 2050262

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	97.3	97.2	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A88470

Included Lab Sample IDs: 2050260

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	102	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A88486

Included Lab Sample IDs: 2050260

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A88490

Included Lab Sample IDs: 2050261

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.0	99.5	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A88500

Included Lab Sample IDs: 2050261

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	102	103	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A88542

Included Lab Sample IDs: 2050260

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	98.4	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A88542

Included Lab Sample IDs: 2050260

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.6	100	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A88546

Included Lab Sample IDs: 2050260

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	103	P/P	90 - 110
Sulfate	99.8	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-00
Run ID: A88576
Included Lab Sample IDs: 2050261

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.2	94.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A88585
Included Lab Sample IDs: 2050261

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A88587
Included Lab Sample IDs: 2050261

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.6	90.5	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		Precision	
					LCS	MS
EPA 180.1 Rev. 2.0	Turbidity				3.08	
EPA 300.0 Rev. 2.1	Chloride	99.3	94.6	96.6	1.45	
	Sulfate	97.5	91.8	96.0	1.36	
EPA 350.1 Rev. 2.0	Ammonia-N	100	103	98.5		3.32
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	106	115	109		3.96
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	100	100	100		0.0
EPA 365.1 Rev. 2.0	Total-P	101	98.2	99.8		1.49
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.2	93.2	95.0		1.71
SM 2120 B	Color (true)	102			0.312	
SM 2320 B-97	Total Alkalinity	101			0.165	
SM 2540 C-97	TDS				2.95	
SM 4500 F-C-97	Fluoride	96.3	98.7	98.7		0.0
SM 5310 B-00	Organic Carbon	95.4	92.0	92.9		0.548

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2050260
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2050260
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2050260
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2050261
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2050261
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2050261
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2050261
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2050262
SM 2120 B / E31780	True Color measured at 450 nm	2050260
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2050260
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2050260
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2050260
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2050260
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2050261

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	12/20/2018			12/20/2018 15:16	Megan Treadwell	2050260
EPA 300.0 Rev. 2.1	12/20/2018			12/29/2018 01:33	Lipi Saha	2050260
EPA 350.1 Rev. 2.0	12/20/2018			12/21/2018 10:42	Ping Hua	2050261
EPA 351.2 Rev. 2.0	12/20/2018	12/27/2018 14:15	Adrian Shelby	01/02/2019 10:44	Joseph Suarez	2050261
EPA 353.2 Rev. 2.0	12/20/2018			12/21/2018 10:20	Lauren Bottorf	2050261
EPA 365.1 Rev. 2.0	12/20/2018	12/28/2018 11:30	Joseph Suarez	01/02/2019 09:19	Beverly Y. Sanders	2050261
EPA 365.1 Rev. 2.0 dissolved	12/20/2018			12/20/2018 14:43	Jhamieka S. Greenwood	2050262
SM 2120 B	12/20/2018			12/20/2018 16:11	Chelsea Hernandez	2050260
SM 2320 B-97	12/20/2018			12/27/2018 23:03	Dale L. Simmons	2050260
SM 2540 C-97	12/20/2018			12/21/2018 11:15	Yijie Li	2050260
SM 2540 D-97	12/20/2018			12/21/2018 11:15	Yijie Li	2050260
SM 4500 F-C-97	12/20/2018			12/27/2018 23:03	Dale L. Simmons	2050260
SM 5310 B-00	12/20/2018			12/21/2018 22:01	Megan Treadwell	2050261