

Chemical Analysis Report

CEN-DIST-2018-05-11-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Tilden X4 + D&B / Starke**
Request ID: **RQ-2018-05-07-63**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Thekkekalathil Chandrasekhar, Chemist Administrator

Date Certified: 29-MAY-2018 15:44



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Lake Starke @ Center

Collection Date/Time: 05/10/2018 10:34

Field ID: G2CE0094

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1992431	EPA 180.1 Rev. 2.0	Turbidity	7.5		NTU	P345148	
	EPA 300.0 Rev. 2.1	Chloride	28		mg Cl/L	P345433	
		Sulfate	21		mg SO4/L	P345435	
	SM 2120 B	Color (true)	20		PCU	P345138	
	SM 2320 B-97	Total Alkalinity	59		mg CaCO3/L	P345483	
	SM 2540 C-97	TDS	154		mg/L	P345721	
	SM 2540 D-97	TSS	6	I	mg/L	P345713	
	SM 4500 F-C-97	Fluoride	0.14		mg F/L	P345486	
1992432	EPA 350.1 Rev. 2.0	Ammonia-N	0.005	I	mg N/L	P345385	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.5		mg N/L	P345535	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P345409	
	EPA 365.1 Rev. 2.0	Total-P	0.020		mg P/L	P345443	
	SM 5310 B-00	Organic Carbon	12		mg C/L	P345545	
1992433	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P345152	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P345148

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P345433

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P345435

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P345385

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P345535

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P345409

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P345443

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P345152

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P345138

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P345483

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P345721

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P345713

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P345486

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P345545

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P345433

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P345435

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	95.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P345385

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P345535

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P345409

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P345443

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P345152

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.3		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P345483

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P345486

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	95.6		P	91 - 105

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
 Batch ID: P345545

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.4		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P345433

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1992246	Chloride	96.9		P	90 - 110
1992309	Chloride	98.5		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P345435

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1992246	Sulfate	93.8		P	90 - 110
1992309	Sulfate	96.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P345385

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1992310	Ammonia-N	97.5	99.3	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P345535

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1992301	Kjeldahl Nitrogen	100	90.7	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P345409

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1992301	NO2NO3-N	99.9	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P345443

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1992598	Total-P	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P345152

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1992307	O-Phosphate-P	101	100	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P345486

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1992225	Fluoride	94.9	93.8	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P345545

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1992342	Organic Carbon	93.3	94.1	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P345148

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992295	Turbidity	3.43	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P345433

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992246	Chloride	0.405	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P345435

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992246	Sulfate	2.03	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P345385

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992310	Ammonia-N	1.48	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P345535

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992301	Kjeldahl Nitrogen	7.45	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P345409

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P345409

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992301	NO2NO3-N	0.380	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P345443

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992598	Total-P	1.27	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P345152

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992307	O-Phosphate-P	0.298	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P345138

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992297	Color (true)	4.75	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P345483

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992225	Total Alkalinity	0.607	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P345721

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992341	TDS	9.42	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P345486

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992225	Fluoride	0.943	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P345545

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1992342	Organic Carbon	0.675	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A83828
Included Lab Sample IDs: 1992431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	99.4	99.3	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A83829
Included Lab Sample IDs: 1992431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	102	101	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A83830
Included Lab Sample IDs: 1992433

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.4	98.4	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A83885
Included Lab Sample IDs: 1992431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	101	P/P	90 - 110
Sulfate	99.4	102	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A83925
Included Lab Sample IDs: 1992432

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	100	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A83934
Included Lab Sample IDs: 1992432

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	101	P/P	90 - 110

Reference Method: SM 2320 B-97
Run ID: A83969
Included Lab Sample IDs: 1992431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.5	106	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 4500 F-C-97
Run ID: A83969
Included Lab Sample IDs: 1992431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.1	99.5	P/P	90 - 110

Reference Method: SM 5310 B-00
Run ID: A83989
Included Lab Sample IDs: 1992432

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.0	95.4	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A84044
Included Lab Sample IDs: 1992432

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	98.8	94.3	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A84048
Included Lab Sample IDs: 1992432

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				3.43	
EPA 300.0 Rev. 2.1	Chloride	98.5	96.9	98.5	0.405	
	Sulfate	95.7	93.8	96.8	2.03	
EPA 350.1 Rev. 2.0	Ammonia-N	101	97.5	99.3		1.48
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	100	90.7		7.45
EPA 353.2 Rev. 2.0	NO2NO3-N	100	99.9	100		0.380
EPA 365.1 Rev. 2.0	Total-P	102	101	102		1.27
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.3	101	100		0.298
SM 2120 B	Color (true)				4.75	
SM 2320 B-97	Total Alkalinity	102			0.607	
SM 2540 C-97	TDS				9.42	
SM 4500 F-C-97	Fluoride	95.6	94.9	93.8		0.943
SM 5310 B-00	Organic Carbon	95.4	93.3	94.1		0.675

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
---------------------	-------------	--------------------

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1992431
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1992431
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1992431
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1992432
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1992432
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1992432
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1992432
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1992433
SM 2120 B / E31780	True Color measured at 450 nm	1992431
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1992431
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1992431
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	1992431
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1992431
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1992432

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	05/11/2018			05/11/2018 14:35	Tanya Welborn	1992431
EPA 300.0 Rev. 2.1	05/11/2018			05/15/2018 18:53	Lipi Saha	1992431
EPA 350.1 Rev. 2.0	05/11/2018			05/15/2018 11:29	Ping Hua	1992432
EPA 351.2 Rev. 2.0	05/11/2018	05/17/2018 15:45	Adrian Shelby	05/18/2018 20:48	Joseph Suarez	1992432
EPA 353.2 Rev. 2.0	05/11/2018			05/16/2018 12:29	Lauren Bottorf	1992432
EPA 365.1 Rev. 2.0	05/11/2018	05/16/2018 08:57	Sima Feizi	05/21/2018 13:00	Beverly Y. Sanders	1992432
EPA 365.1 Rev. 2.0 dissolved	05/11/2018			05/11/2018 15:24	Julia Storbeck	1992433
SM 2120 B	05/11/2018			05/11/2018 14:57	Tanya Welborn	1992431
SM 2320 B-97	05/11/2018			05/16/2018 06:31	Dale L. Simmons	1992431
SM 2540 C-97	05/11/2018			05/16/2018 15:26	Yijie Li	1992431
SM 2540 D-97	05/11/2018			05/16/2018 15:26	Yijie Li	1992431
SM 4500 F-C-97	05/11/2018			05/16/2018 06:31	Dale L. Simmons	1992431
SM 5310 B-00	05/11/2018			05/17/2018 04:53	Megan Treadwell	1992432