

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: Betty/67

RQ: 2017-09-25-36

Project: SPRINGS

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 7A 12/17 12/1/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	T. Riordan	1/16/18	720	21.11	8.898	8.90	100.0	51.2	1.06343	<u>P</u> /F	L/ <u>F</u>
ICV				21.14	8.898	8.89	100.0	51.2		<u>P</u> /F	L/ <u>F</u>
CCV	Leif Boman		16:02	23.75	8.450	8.44	99.7	51.2	1.0647	<u>P</u> /F	<u>P</u> /F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	T. Riordan	1/16/18	720	170831	9/18	1000	1000	<u>P</u> /F	L/ <u>F</u>
ICV				171023	11/18	100	99	<u>P</u> /F	L/ <u>F</u>
CCV	Leif Boman		16:02	2601329	01/18	10,000	10,077	<u>P</u> /F	<u>P</u> /F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	T. Riordan	1/16/18	720	165674	8/18	7.0	6.98	-24.4	<u>P</u> /F	L/ <u>F</u>
Calibr.				172189	4/19	4.0	4.01	153.3	<u>P</u> /F	L/ <u>F</u>
Calibr.				174997	7/19	10.0	9.98	-201.1	<u>P</u> /F	L/ <u>F</u>
ICV				171629	4/19	4.0	4.02	153.0	<u>P</u> /F	L/ <u>F</u>
CCV	Leif Boman		16:02	174997	7/19	10.0	9.88	195.6	<u>P</u> /F	<u>P</u> /F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 176.7, slope from 4 to 7 177.7 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: 300P NON RQ: 2018-01-15-36 Project: SMP

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site _____

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 12/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Mike Drennan	1/16/18	805	21.47	8.829	8.83	100.00	50.2	1.03811	P/F	L/F
ICV			1	21.53	8.829	8.82	99.9	50.2		P/F	L/F
CCV			1520	21.34	8.863	8.92	100.7	49.2	1.03811	P/F	L/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.
 Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.
 Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot # <u>LAP</u>	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.			805	2702761	1/19	50000	50000	P/F	L/F
ICV			1	170831	9/18	10000	9996	P/F	L/F
CCV				2601329	1/18			P/F	L/F
CCV			1520	2601329	1/18	10000	9985	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.			805	165674	8/18	7.0	7.00	-7.0	P/F	L/F
Calibr.				171629	4/19	4.0	4.00	169.1	P/F	L/F
Calibr.				174997	7/19	10.0	10.00	-181.2	P/F	L/F
ICV				174997	7/19	10.0	9.98	-180.6	P/F	L/F
CCV			1520	174997	7/19	10.0	9.98	-180.7	P/F	L/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;
 If mV are recorded: slope from 7 to 10 174.2, slope from 4 to 7 176.1 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: