

Chemical Analysis Report

CEN-DIST-2018-05-02-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Withlacoochee River**
Request ID: **RQ-2018-04-30-32**
Customer: **CEN-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 17-MAY-2018 12:58



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

**Sample Location: WITHLACOOCHIE RIVER @ UPSTREAM OF
SR 33**

Collection Date/Time: 05/01/2018 11:09

Field ID: G4CE0138

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1989234	EPA 180.1 Rev. 2.0	Turbidity	1.5		NTU	P344508	
	EPA 300.0 Rev. 2.1	Chloride	18		mg Cl/L	P344993	
		Sulfate	0.20	I	mg SO4/L	P344997	
	SM 2120 B	Color (true)	440		PCU	P344493	
	SM 2320 B-97	Total Alkalinity	3.5		mg CaCO3/L	P345071	
	SM 2540 C-97	TDS	130		mg/L	P345190	
	SM 2540 D-97	TSS	2	I	mg/L	P345183	
	SM 4500 F-C-97	Fluoride	0.068	I	mg F/L	P345075	
1989235	EPA 350.1 Rev. 2.0	Ammonia-N	0.11		mg N/L	P344926	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.6		mg N/L	P344640	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P344593	
	EPA 365.1 Rev. 2.0	Total-P	0.091		mg P/L	P344796	
	SM 5310 B-00	Organic Carbon	44		mg C/L	P344911	
1989236	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.050		mg P/L	P344568	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Chloride: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

EPA 300.0 Rev. 2.1: Sulfate: : Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

SM 5310 B-00: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P344508

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P344993

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P344997

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P344926

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P344640

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P344593

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P344796

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P344568

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P344493

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P345071

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P345190

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P345183

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P345075

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Quality Assurance Report Method Blank Results

Reference Method: SM 5310 B-00
Batch ID: P344911

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P344993

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P344997

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	93.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P344926

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P344640

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	107		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P344593

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P344796

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	105		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P344568

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P345071

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
-----------	---------	---------	-----------	----------------

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 2320 B-97
Batch ID: P345071

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	105		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P345075

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	95.1		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P344911

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P344993

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1989206	Chloride	95.2		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P344997

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1989296	Sulfate	95.8		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P344640

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1989273	Kjeldahl Nitrogen	97.6	99.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P344593

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1989210	NO2NO3-N	102	104	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P344796

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1989676	Total-P	102	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P344568

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P344568

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1989160	O-Phosphate-P	97.3	98.2	P/P	90 - 110

Reference Method: SM 4500 F-C-97
 Batch ID: P345075

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1989011	Fluoride	94.5	95.0	P/P	91 - 105

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P344508

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989151	Turbidity	4.11	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P344993

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989206	Chloride	0.290	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P344997

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989206	Sulfate	0.477	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P344926

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989235	Ammonia-N	0.0706	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P344640

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989273	Kjeldahl Nitrogen	1.55	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P344593

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P344593

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989210	NO2NO3-N	0.971	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P344796

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989676	Total-P	2.10	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P344568

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989160	O-Phosphate-P	0.921	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P344493

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989206	Color (true)	0.722	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P345071

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989011	Total Alkalinity	0.856	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P345190

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989208	TDS	5.12	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P345075

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989011	Fluoride	0.472	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P344911

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1989235	Organic Carbon	0.126	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A83602
Included Lab Sample IDs: 1989234

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	101	102	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A83605
Included Lab Sample IDs: 1989234

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A83633
Included Lab Sample IDs: 1989236

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	95.7	95.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A83636
Included Lab Sample IDs: 1989235

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A83698
Included Lab Sample IDs: 1989235

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	104	106	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A83732
Included Lab Sample IDs: 1989234

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	102	P/P	90 - 110
Sulfate	98.7	101	P/P	90 - 110

Reference Method: SM 5310 B-00
Run ID: A83745
Included Lab Sample IDs: 1989235

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.6	98.3	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A83756

Included Lab Sample IDs: 1989235

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	99.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A83782

Included Lab Sample IDs: 1989235

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	103	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A83812

Included Lab Sample IDs: 1989234

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.1	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A83812

Included Lab Sample IDs: 1989234

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.6	98.1	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				4.11	
EPA 300.0 Rev. 2.1	Chloride	99.7	95.2		0.290	
	Sulfate	93.7	95.8		0.477	
EPA 350.1 Rev. 2.0	Ammonia-N	101				0.0706
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	107	97.6 99.8			1.55
EPA 353.2 Rev. 2.0	NO2NO3-N	102	102 104			0.971
EPA 365.1 Rev. 2.0	Total-P	105	102 99.0			2.10
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101	97.3 98.2			0.921
SM 2120 B	Color (true)				0.722	
SM 2320 B-97	Total Alkalinity	105			0.856	
SM 2540 C-97	TDS				5.12	
SM 4500 F-C-97	Fluoride	95.1	94.5 95.0			0.472
SM 5310 B-00	Organic Carbon	98.8				0.126

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
---------------------	-------------	--------------------

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1989234
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1989234
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1989234
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1989235
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1989235
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1989235
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1989235
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1989236
SM 2120 B / E31780	True Color measured at 450 nm	1989234
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1989234
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1989234
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1989234
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1989234
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1989235

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	05/02/2018			05/02/2018 17:27	William L. Brown IV	1989234
EPA 300.0 Rev. 2.1	05/02/2018			05/08/2018 16:27	Lipi Saha	1989234
EPA 350.1 Rev. 2.0	05/02/2018			05/08/2018 15:04	Ping Hua	1989235
EPA 351.2 Rev. 2.0	05/02/2018	05/04/2018 11:20	Adrian Shelby	05/07/2018 10:52	Joseph Suarez	1989235
EPA 353.2 Rev. 2.0	05/02/2018			05/03/2018 14:37	Lauren Bottorf	1989235
EPA 365.1 Rev. 2.0	05/02/2018	05/07/2018 13:25	Sima Feizi	05/09/2018 08:18	Beverly Y. Sanders	1989235
EPA 365.1 Rev. 2.0 dissolved	05/02/2018			05/02/2018 16:59	Megan Treadwell	1989236
SM 2120 B	05/02/2018			05/02/2018 16:21	Tanya Welborn	1989234
SM 2320 B-97	05/02/2018			05/10/2018 02:46	Dale L. Simmons	1989234
SM 2540 C-97	05/02/2018			05/05/2018 23:12	Yijie Li	1989234
SM 2540 D-97	05/02/2018			05/05/2018 23:12	Yijie Li	1989234
SM 4500 F-C-97	05/02/2018			05/10/2018 02:46	Dale L. Simmons	1989234
SM 5310 B-00	05/02/2018			05/07/2018 16:36	Megan Treadwell	1989235