

Chemical Analysis Report

CEN-DIST-2018-06-19-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Reedy Creek Bw Russell + D&B**

Request ID: **RQ-2018-06-18-34**

Customer: **CEN-DIST**

Project ID: **SMP**

Send Reports to:

FL Dept. of Environmental Protection
FL Dept. of Environmental Protection
3319 Maguire Blvd., Suite 232
Orlando, FL 32803-3767
Attn: Mike Linger

For additional information please contact

Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 10-JUL-2018 11:38



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Reedy Creek @ 200m upstream of confluence w/ Dead River **Collection Date/Time: 06/18/2018 10:29**

Field ID: G4CE0145

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2001825	EPA 180.1 Rev. 2.0	Turbidity	2.6		NTU	P347145	
	EPA 300.0 Rev. 2.1	Chloride	17		mg Cl/L	P347541	
		Sulfate	17		mg SO4/L	P347545	
		Color (true)	680		PCU	P347138	
	SM 2320 B-97	Total Alkalinity	28		mg CaCO3/L	P347375	RPD
	SM 2540 C-97	TDS	212		mg/L	P347485	
	SM 2540 D-97	TSS	4	I	mg/L	P347506	
	SM 4500 F-C-97	Fluoride	0.078	I	mg F/L	P347379	
2001826	EPA 350.1 Rev. 2.0	Ammonia-N	0.034		mg N/L	P347385	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.8		mg N/L	P347821	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.008	I	mg N/L	P347511	
	EPA 365.1 Rev. 2.0	Total-P	0.088		mg P/L	P347322	
	SM 5310 B-00	Organic Carbon	57		mg C/L	P347412	
2001827	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.033		mg P/L	P347131	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Chloride: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Reedy Creek @ 200m upstream of confluence w/ Dead River **Collection Date/Time: 06/18/2018 10:32**

Field ID: G4CE0145_DUP

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2001828	EPA 180.1 Rev. 2.0	Turbidity	2.2		NTU	P347146	
	EPA 300.0 Rev. 2.1	Chloride	17		mg Cl/L	P347541	
		Sulfate	17		mg SO4/L	P347545	
		Color (true)	660		PCU	P347138	
	SM 2320 B-97	Total Alkalinity	27		mg CaCO3/L	P347375	RPD
	SM 2540 C-97	TDS	226		mg/L	P347485	
	SM 2540 D-97	TSS	3	I	mg/L	P347506	
	SM 4500 F-C-97	Fluoride	0.078	I	mg F/L	P347379	
2001830	EPA 350.1 Rev. 2.0	Ammonia-N	0.033		mg N/L	P347385	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.5		mg N/L	P347821	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.007	I	mg N/L	P347511	
	EPA 365.1 Rev. 2.0	Total-P	0.088		mg P/L	P347322	
	SM 5310 B-00	Organic Carbon	56		mg C/L	P347412	
2001832	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.034		mg P/L	P347131	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Chloride: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Sample Location: Field Blank

Collection Date/Time: 06/18/2018 11:30

Field ID: BLANK_06182018

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2001829	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P347146	
	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P347538	
		Sulfate	0.20	U	mg SO4/L	P347542	
	SM 2120 B	Color (true)	2.5	U	PCU	P347138	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P347375	RPD
	SM 2540 C-97	TDS	15	U	mg/L	P347482	
	SM 2540 D-97	TSS	2	U	mg/L	P347503	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P347379	
2001831	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P347385	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P347821	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P347368	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P347322	
	SM 5310 B-00	Organic Carbon	0.50	U	mg C/L	P347527	
2001833	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P347131	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P347145

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P347146

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347538

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347541

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347542

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347545

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347385

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P347821

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347368

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347511

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347322

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P347131

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P347138

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P347375

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P347482

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 C-97
Batch ID: P347485

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P347503

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P347506

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 4500 F-C-97
Batch ID: P347379

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P347412

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Reference Method: SM 5310 B-00
Batch ID: P347527

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347538

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.8		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347541

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	105		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347542

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.6		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347545

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	106		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347385

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	96.3		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P347821

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	94.9		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347368

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347511

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347322

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	96.4		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P347131

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P347375

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	100		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P347379

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.9		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P347412

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	96.7		P	85 - 115

Reference Method: SM 5310 B-00
Batch ID: P347527

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	95.7		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347538

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2001937	Chloride	97.0		P	90 - 110
2001981	Chloride	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347541

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1999148	Chloride	106		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347542

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2001937	Sulfate	97.0		P	90 - 110
2001981	Sulfate	103		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347545

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1999148	Sulfate	105		P	90 - 110
1999689	Sulfate	107		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347385

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2001697	Ammonia-N	98.3	95.1	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P347821

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2001872	Kjeldahl Nitrogen	99.6	112	P/F	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347368

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2001873	NO2NO3-N	99.4	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347511

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2002230	NO2NO3-N	102	103	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347322

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2001869	Total-P	97.6	97.8	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P347379

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2002011	Fluoride	98.3	97.1	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P347412

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2001636	Organic Carbon	95.8	98.4	P/P	85 - 115

Reference Method: SM 5310 B-00
Batch ID: P347527

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2002105	Organic Carbon	96.6	95.7	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P347145

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001635	Turbidity	2.43	Sample	P	0 - 20

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P347146

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001866	Turbidity	0.623	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347538

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001937	Chloride	3.30	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347541

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1999689	Chloride	5.23	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347542

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001937	Sulfate	3.54	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347545

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1999689	Sulfate	6.12	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347385

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001697	Ammonia-N	3.14	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P347821

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001872	Kjeldahl Nitrogen	9.94	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347368

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001873	NO2NO3-N	0.349	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347511

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2002230	NO2NO3-N	0.487	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347322

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001869	Total-P	0.273	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P347131

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001657	O-Phosphate-P	0.312	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P347138

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001884	Color (true)	2.91	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P347375

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2002011	Total Alkalinity	6.11	Sample	F	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P347482

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001113	TDS	1.11	Sample	P	0 - 10

Reference Method: SM 2540 C-97
Batch ID: P347485

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001635	TDS	2.60	Sample	P	0 - 10

Quality Assurance Report Precision

Reference Method: SM 4500 F-C-97
Batch ID: P347379

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2002011	Fluoride	0.940	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P347412

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2001636	Organic Carbon	1.45	Spike	P	0 - 20

Reference Method: SM 5310 B-00
Batch ID: P347527

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2002105	Organic Carbon	0.795	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A84681

Included Lab Sample IDs: 2001827, 2001832, 2001833

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	99.6	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A84682

Included Lab Sample IDs: 2001825, 2001828, 2001829

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	101	P/P	90 - 115
Color (true)	102	102	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A84684

Included Lab Sample IDs: 2001825, 2001828, 2001829

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	99.0	P/P	95 - 105
Turbidity	99.5	98.5	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A84781

Included Lab Sample IDs: 2001831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A84785

Included Lab Sample IDs: 2001825, 2001828, 2001829

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.0	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A84785

Included Lab Sample IDs: 2001825, 2001828, 2001829

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	102	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A84788

Included Lab Sample IDs: 2001826, 2001830, 2001831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A84794

Included Lab Sample IDs: 2001826, 2001830

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
-----------	---------	---------	------------	----------------

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-00

Run ID: A84794

Included Lab Sample IDs: 2001826, 2001830

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.4	97.0	P/P	90 - 110
Organic Carbon	97.0	95.6	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A84842

Included Lab Sample IDs: 2001826, 2001830

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A84844

Included Lab Sample IDs: 2001831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	95.6	95.6	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A84847

Included Lab Sample IDs: 2001825, 2001828, 2001829

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Chloride	99.2	99.2	P/P	90 - 110
Sulfate	103	103	P/P	90 - 110
Sulfate	99.3	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A84885

Included Lab Sample IDs: 2001826, 2001830

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	99.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A84929

Included Lab Sample IDs: 2001831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.8	99.8	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A85016

Included Lab Sample IDs: 2001826, 2001830, 2001831

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	96.7	101	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					2.43	
	Turbidity					0.623	
EPA 300.0 Rev. 2.1	Chloride	97.8	97.0	103		3.30	
	Chloride	105	106			5.23	
	Sulfate	98.6	97.0	103		3.54	
	Sulfate	106	107	105		6.12	
EPA 350.1 Rev. 2.0	Ammonia-N	96.3	98.3	95.1			3.14
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	94.9	99.6	112			9.94
EPA 353.2 Rev. 2.0	NO2NO3-N	100	99.4	100			0.349
	NO2NO3-N	102	102	103			0.487
EPA 365.1 Rev. 2.0	Total-P	96.4	97.6	97.8			0.273
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	101					0.312
SM 2120 B	Color (true)					2.91	
SM 2320 B-97	Total Alkalinity	100				6.11	
SM 2540 C-97	TDS					1.11	
	TDS					2.60	
SM 4500 F-C-97	Fluoride	97.9	98.3	97.1			0.940
SM 5310 B-00	Organic Carbon	96.7	95.8	98.4			1.45
	Organic Carbon	95.7	96.6	95.7			0.795

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2001825, 2001828, 2001829
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2001825, 2001828, 2001829
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2001825, 2001828, 2001829
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2001826, 2001830, 2001831
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2001826, 2001830, 2001831
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2001826, 2001830, 2001831
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2001826, 2001830, 2001831
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2001827, 2001832, 2001833
SM 2120 B / E31780	True Color measured at 450 nm	2001825, 2001828, 2001829
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	2001825, 2001828, 2001829
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2001825, 2001828, 2001829
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2001825, 2001828, 2001829
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2001825, 2001828, 2001829
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2001826, 2001830, 2001831

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/19/2018			06/19/2018 16:44	William L. Brown IV	2001825, 2001828, 2001829
EPA 300.0 Rev. 2.1	06/19/2018			06/25/2018 16:27	Lipi Saha	2001829
	06/19/2018			06/26/2018 15:02	Lipi Saha	2001825
	06/19/2018			06/26/2018 15:14	Lipi Saha	2001828
EPA 350.1 Rev. 2.0	06/19/2018			06/22/2018 12:58	Julia Storbeck	2001826
	06/19/2018			06/22/2018 13:00	Julia Storbeck	2001830
	06/19/2018			06/22/2018 13:03	Julia Storbeck	2001831
EPA 351.2 Rev. 2.0	06/19/2018	07/02/2018 09:30	Adrian Shelby	07/03/2018 20:33	Virginia P. Brown	2001826
	06/19/2018	07/02/2018 09:30	Adrian Shelby	07/03/2018 20:35	Virginia P. Brown	2001830
	06/19/2018	07/02/2018 09:30	Adrian Shelby	07/03/2018 20:36	Virginia P. Brown	2001831
EPA 353.2 Rev. 2.0	06/19/2018			06/22/2018 14:03	Lauren Bottorf	2001831
	06/19/2018			06/26/2018 12:46	Lauren Bottorf	2001826
	06/19/2018			06/26/2018 12:48	Lauren Bottorf	2001830
EPA 365.1 Rev. 2.0	06/19/2018	06/22/2018 12:15	Sima Feizi	06/27/2018 13:26	Beverly Y. Sanders	2001826
	06/19/2018	06/22/2018 12:15	Sima Feizi	06/27/2018 13:27	Beverly Y. Sanders	2001830
	06/19/2018	06/22/2018 12:15	Sima Feizi	06/29/2018 10:54	Beverly Y. Sanders	2001831
EPA 365.1 Rev. 2.0 dissolved	06/19/2018			06/19/2018 15:07	Megan Treadwell	2001827
	06/19/2018			06/19/2018 15:08	Megan Treadwell	2001832
	06/19/2018			06/19/2018 15:09	Megan Treadwell	2001833
SM 2120 B	06/19/2018			06/19/2018 16:03	Tanya Welborn	2001829
	06/19/2018			06/19/2018 16:25	Tanya Welborn	2001825
	06/19/2018			06/19/2018 16:26	Tanya Welborn	2001828
SM 2320 B-97	06/19/2018			06/22/2018 05:20	Lauren Bottorf	2001825
	06/19/2018			06/22/2018 05:30	Lauren Bottorf	2001828
	06/19/2018			06/22/2018 05:44	Lauren Bottorf	2001829
SM 2540 C-97	06/19/2018			06/20/2018 15:01	Yijie Li	2001825, 2001828, 2001829
SM 2540 D-97	06/19/2018			06/20/2018 15:01	Yijie Li	2001825, 2001828, 2001829
SM 4500 F-C-97	06/19/2018			06/22/2018 05:20	Lauren Bottorf	2001825
	06/19/2018			06/22/2018 05:30	Lauren Bottorf	2001828
	06/19/2018			06/22/2018 05:44	Lauren Bottorf	2001829
SM 5310 B-00	06/19/2018			06/22/2018 22:13	Blanca Fach	2001826
	06/19/2018			06/22/2018 23:00	Blanca Fach	2001830
	06/19/2018			06/26/2018 01:30	Megan Treadwell	2001831