

Wood-McGrath, Avril

From: June Flowers <june@flowerslabs.com>
Sent: Tuesday, August 21, 2018 12:14 PM
To: Wood-McGrath, Avril
Cc: DEP-Overflowmicro; Power, Diane; Michael Conner (michael@flowerslabs.com)
Subject: Re: 374433 J question

Yes Avril,

I am including Michael Connor in this email for training - he was not told to add the explanation:

"J" flag indicates: All Quanti-Tray wells were positive. Result may be higher than reported.

The revised report will soon be processed.

Thank you

June Flowers
Quality Systems Director
Flowers Chemical Laboratories, Inc.
481 Newburyport Ave., Altamonte Springs FL 32701
PO Box 150597, Altamonte Springs FL 32715-0597
407.339.5984 x212
www.flowerslabs.com

Celebrating 61 years 1957 - 2018

On Aug 21, 2018, at 11:09 AM, Wood-McGrath, Avril <Avril.Wood-McGrath@dep.state.fl.us> wrote:

Hi Ms. June,
Are the J qualifiers on 374433 because all wells were positive?

Avril Wood-McGrath
OPS Environmental Specialist II
FDEP Laboratory - Biology Program
Florida Department of Environmental Protection
Avril.Wood-McGrath@floridadep.gov
Office: 850.245.8193

