

Chemical Analysis Report

SO-DIST-2018-06-14-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **Tellywog**
Request ID: **RQ-2018-06-04-55**
Customer: **SO-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FLDEP South District Office
2295 Victoria Ave. Suite 364
Fort Myers, FL 33901-3381
Attn: Nicole Reistad

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 09-JUL-2018 17:58



NON-CONFORMANCE REPORT INCLUDED

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Telegraph Swamp South

Collection Date/Time: 06/13/2018 09:10

Field ID: G3SD0100

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2000777	EPA 180.1 Rev. 2.0	Turbidity	2.6		NTU	P346930	
	EPA 300.0 Rev. 2.1	Chloride	73		mg Cl/L	P347396	
		Sulfate	13		mg SO4/L	P347399	
	SM 2120 B	Color (true)	330		PCU	P346913	
	SM 2320 B-97	Total Alkalinity	73		mg CaCO3/L	P347268	RPD
	SM 2540 C-97	TDS	313		mg/L	P347171	
	SM 2540 D-97	TSS	4	I	mg/L	P347174	
	SM 4500 F-C-97	Fluoride	0.13		mg F/L	P347272	
2000778	EPA 350.1 Rev. 2.0	Ammonia-N	0.044		mg N/L	P347346	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.6		mg N/L	P347968	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.036		mg N/L	P347510	
	EPA 365.1 Rev. 2.0	Total-P	0.078		mg P/L	P347192	
	SM 5310 B-00	Organic Carbon	32		mg C/L	P347186	
2000779	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.042		mg P/L	P346927	
2000798	EPA 200.7 Rev. 4.4	Calcium	38.7		mg/L	P347060	
		Iron	1.02E+03		ug/L	P347060	
		Magnesium	7.73		mg/L	P347060	
		Potassium	2.0		mg/L	P347060	
		Sodium	36.5		mg/L	P347060	
	EPA 200.8 Rev. 5.4	Zinc	5.0	U	ug/L	P347060	
		Arsenic	1.16		ug/L	P347060	
		Cadmium	0.020	U	ug/L	P347060	
		Chromium	0.71	I	ug/L	P347060	
		Copper	0.68	I	ug/L	P347060	
		Lead	0.19	I	ug/L	P347060	
		Nickel	0.43	I	ug/L	P347060	
		Silver	0.010	U	ug/L	P347060	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Bee Branch @LoblollyBay

Collection Date/Time: 06/13/2018 10:50

Field ID: G3SD0085

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2000771	EPA 180.1 Rev. 2.0	Turbidity	2.7		NTU	P346930	
	EPA 300.0 Rev. 2.1	Chloride	27	A	mg Cl/L	P347396	
		Sulfate	8.6	A	mg SO4/L	P347399	
	SM 2120 B	Color (true)	470		PCU	P346913	
	SM 2320 B-97	Total Alkalinity	14		mg CaCO3/L	P347268	RPD
	SM 2540 C-97	TDS	187		mg/L	P347171	
	SM 2540 D-97	TSS	4	I	mg/L	P347174	
	SM 4500 F-C-97	Fluoride	0.080	I	mg F/L	P347272	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.029		mg N/L	P347342	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.9		mg N/L	P347968	MS
2000773	EPA 353.2 Rev. 2.0	NO2NO3-N	0.065		mg N/L	P347510	
	EPA 365.1 Rev. 2.0	Total-P	0.15		mg P/L	P347192	
	SM 5310 B-00	Organic Carbon	39		mg C/L	P347186	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.099		mg P/L	P346927	
	dissolved						

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Sample Location: Pollywog Creek @ SR 78

Collection Date/Time: 06/13/2018 11:35

Field ID: G3SD0098

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2000772	EPA 180.1 Rev. 2.0	Turbidity	3.1		NTU	P346930	
	EPA 300.0 Rev. 2.1	Chloride	13		mg Cl/L	P347396	
		Sulfate	2.4		mg SO4/L	P347399	
	SM 2120 B	Color (true)	630		PCU	P346913	
	SM 2320 B-97	Total Alkalinity	15		mg CaCO3/L	P347268	RPD
	SM 2540 C-97	TDS	151		mg/L	P347171	
	SM 2540 D-97	TSS	4	I	mg/L	P347174	
	SM 4500 F-C-97	Fluoride	0.071	I	mg F/L	P347272	
2000774	EPA 350.1 Rev. 2.0	Ammonia-N	0.045	Y	mg N/L	P347346	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	2.0	Y	mg N/L	P347968	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.019	Y	mg N/L	P347510	
	EPA 365.1 Rev. 2.0	Total-P	0.40	Y	mg P/L	P347192	
	SM 5310 B-00	Organic Carbon	40	Y	mg C/L	P347186	
2000776	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.31		mg P/L	P346927	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 350.1 Rev. 2.0: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

EPA 351.2 Rev. 2.0: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

EPA 353.2 Rev. 2.0: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

EPA 365.1 Rev. 2.0: Total-P: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

SM 5310 B-00: The sample was received unpreserved: pH > 2. Sample preserved in laboratory.

Sample Location: Pollywog Creek @ SR 78

Collection Date/Time: 06/13/2018 11:50

Field ID: BLANK_06132018

Matrix: W-FIELD-BK

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2000780	EPA 180.1 Rev. 2.0	Turbidity	0.10	U	NTU	P346930	
	EPA 300.0 Rev. 2.1	Chloride	0.20	U	mg Cl/L	P347396	
		Sulfate	0.20	U	mg SO4/L	P347399	
	SM 2120 B	Color (true)	2.5	U	PCU	P346913	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P347268	RPD
	SM 2540 C-97	TDS	15	U	mg/L	P347171	
	SM 2540 D-97	TSS	2	U	mg/L	P347175	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P347272	
2000781	EPA 350.1 Rev. 2.0	Ammonia-N	0.002	U	mg N/L	P347346	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.080	U	mg N/L	P347968	MS
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.004	U	mg N/L	P347510	
	EPA 365.1 Rev. 2.0	Total-P	0.002	U	mg P/L	P347192	
	SM 5310 B-00	Organic Carbon	0.50	U	mg C/L	P347186	
2000782	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.004	U	mg P/L	P346928	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Non-Conformance Report

NCR ID: 7084

Event(s)

SO-DIST-2018-06-14-02

Job(s)

TLH-2018-06-14-25

Sample(s)

2000774

Test(s)

NCR Type: SHIPPING/RECEIVING

NCR Category: Preservation Not Intact

Observation: Sample not preserved to pH < 2.

Resolution: Sample preserved in the laboratory with sulfuric acid on 6-14-18 at 1:02pm

Sulfuric Acid Lot:SA8085170 EXP:03-26-19

Authorized by/Date: Joshua Ayres 6/20/2018

The Non-Conformance Report details exceptions or problems encountered with the events/jobs/samples/test.
Please address questions to:

Chemistry	Colin Wright	(850) 245-8085
Biology	Cheryl Swanson	(850) 245-8177

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P346930

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P347060

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P347060

Component	Result	Code	Units
Arsenic	0.050	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.20	U	ug/L
Lead	0.050	U	ug/L
Nickel	0.20	U	ug/L
Silver	0.010	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347396

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347399

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347342

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347346

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P347968

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347510

Component	Result	Code	Units
NO ₂ NO ₃ -N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347192

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P346927

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P346928

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P346913

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P347268

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P347171

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P347174

Component	Result	Code	Units
TSS	2	U	mg/L

Quality Assurance Report

Method Blank Results

Reference Method: SM 2540 D-97
Batch ID: P347175

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P347272

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P347186

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P347060

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	104		P	85 - 115
Iron	106		P	85 - 115
Magnesium	111		P	85 - 115
Potassium	105		P	85 - 115
Sodium	101		P	85 - 115
Zinc	103		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P347060

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Arsenic	98.0		P	85 - 115
Cadmium	99.5		P	85 - 115
Chromium	96.0		P	85 - 115
Copper	96.6		P	85 - 115
Lead	96.2		P	85 - 115
Nickel	98.3		P	85 - 115
Silver	101		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347396

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	99.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347399

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347342

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	103		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347346

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.9		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P347968

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347510

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	100		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347192

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.4		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P346927

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.3		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P346928

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	101		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P347268

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P347272

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.2		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P347186

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	97.2		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
 Batch ID: P347060

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000558	Calcium	104		P	80 - 120
2000558	Iron	109		P	80 - 120
2000558	Magnesium	106		P	80 - 120
2000558	Potassium	103		P	80 - 120
2000558	Sodium	102		P	80 - 120
2000558	Zinc	106		P	80 - 120
2000941	Calcium	104		P	80 - 120
2000941	Iron	106	105	P/P	80 - 120
2000941	Magnesium	107	106	P/P	80 - 120
2000941	Potassium	102	100	P/P	80 - 120
2000941	Sodium	103		P	80 - 120
2000941	Zinc	107	105	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
 Batch ID: P347060

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000558	Arsenic	96.8		P	80 - 120
2000558	Cadmium	97.7		P	80 - 120
2000558	Chromium	93.7		P	80 - 120
2000558	Copper	93.9		P	80 - 120
2000558	Lead	94.0		P	80 - 120
2000558	Nickel	95.2		P	80 - 120
2000558	Silver	96.3		P	80 - 120
2000941	Arsenic	101	99.0	P/P	80 - 120
2000941	Cadmium	103	99.6	P/P	80 - 120
2000941	Chromium	93.6	95.3	P/P	80 - 120
2000941	Copper	98.8	98.2	P/P	80 - 120
2000941	Lead	99.5	99.3	P/P	80 - 120
2000941	Nickel	99.5	99.0	P/P	80 - 120
2000941	Silver	101	99.6	P/P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P347396

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000771	Chloride	97.8		P	90 - 110
2001110	Chloride	96.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P347399

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000771	Sulfate	101		P	90 - 110
2001110	Sulfate	97.1		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P347342

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000529	Ammonia-N	101	99.1	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347346

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000977	Ammonia-N	96.2	99.5	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P347968

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000667	Kjeldahl Nitrogen	110	105	F/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347510

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000773	NO2NO3-N	96.2	97.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347192

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000938	Total-P	95.6	98.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P346927

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000498	O-Phosphate-P	98.1	99.3	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P346928

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000978	O-Phosphate-P	99.9	100	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P347272

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000974	Fluoride	97.0	95.9	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P347186

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2000492	Organic Carbon	112	102	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P346930

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000711	Turbidity	3.97	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P347060

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000941	Calcium	1.21	Spike	P	0 - 20
2000941	Iron	1.21	Spike	P	0 - 20
2000941	Magnesium	0.340	Spike	P	0 - 20
2000941	Potassium	0.644	Spike	P	0 - 20
2000941	Sodium	0.835	Spike	P	0 - 20
2000941	Zinc	2.06	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P347060

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000941	Arsenic	1.99	Spike	P	0 - 20
2000941	Cadmium	3.63	Spike	P	0 - 20
2000941	Chromium	1.70	Spike	P	0 - 20
2000941	Copper	0.564	Spike	P	0 - 20
2000941	Lead	0.198	Spike	P	0 - 20
2000941	Nickel	0.434	Spike	P	0 - 20
2000941	Silver	1.28	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347396

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000771	Chloride	3.48	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347399

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000771	Sulfate	4.43	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347342

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000529	Ammonia-N	1.24	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P347346

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000977	Ammonia-N	3.02	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P347968

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000667	Kjeldahl Nitrogen	4.11	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347510

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000773	NO2NO3-N	1.16	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347192

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000938	Total-P	2.64	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P346927

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000498	O-Phosphate-P	0.731	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P346928

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000978	O-Phosphate-P	0.599	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P346913

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000726	Color (true)	0.208	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P347268

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000974	Total Alkalinity	2.09	Sample	F	0 - 1.6

Quality Assurance Report Precision

Reference Method: SM 2540 C-97
Batch ID: P347171

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000975	TDS	9.84	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P347272

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000974	Fluoride	0.953	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P347186

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2000492	Organic Carbon	6.60	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B

Run ID: A84589

Included Lab Sample IDs: 2000771, 2000772, 2000777, 2000780

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	100	100	P/P	90 - 115
Color (true)	99.8	100	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A84591

Included Lab Sample IDs: 2000775, 2000776, 2000779, 2000782

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.1	98.4	P/P	90 - 110
O-Phosphate-P	99.9	98.1	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A84592

Included Lab Sample IDs: 2000771, 2000772, 2000777, 2000780

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	98.5	P/P	95 - 105

Reference Method: EPA 200.7 Rev. 4.4

Run ID: A84686

Included Lab Sample IDs: 2000798

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	102	103	P/P	90 - 110
Iron	102	104	P/P	90 - 110
Magnesium	103	105	P/P	90 - 110
Potassium	101	103	P/P	90 - 110
Sodium	101	103	P/P	90 - 110
Zinc	101	103	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A84687

Included Lab Sample IDs: 2000798

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Arsenic	99.1	98.6	P/P	90 - 110
Cadmium	98.8	99.0	P/P	90 - 110
Copper	99.9	101	P/P	90 - 110
Lead	100	99.8	P/P	90 - 110
Nickel	98.8	99.7	P/P	90 - 110
Silver	99.7	98.7	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A84709

Included Lab Sample IDs: 2000773, 2000774, 2000778, 2000781

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	97.2	95.8	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A84742

Included Lab Sample IDs: 2000771, 2000772, 2000777, 2000780

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	99.5	103	P/P	90 - 110
Sulfate	99.0	104	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A84752

Included Lab Sample IDs: 2000771, 2000772, 2000777, 2000780

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	101	102	P/P	90 - 110
Total Alkalinity	102	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A84752

Included Lab Sample IDs: 2000771, 2000772, 2000777, 2000780

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	101	P/P	90 - 110
Fluoride	101	97.1	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A84761

Included Lab Sample IDs: 2000773, 2000774, 2000778

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	99.2	100	P/P	90 - 110

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A84763

Included Lab Sample IDs: 2000798

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chromium	93.6	95.6	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A84774

Included Lab Sample IDs: 2000781

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	99.5	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A84776

Included Lab Sample IDs: 2000773

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	102	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A84777

Included Lab Sample IDs: 2000774, 2000778, 2000781

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	102	103	P/P	90 - 110
Ammonia-N	103	103	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A84842

Included Lab Sample IDs: 2000773, 2000774, 2000778, 2000781

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A85040

Included Lab Sample IDs: 2000773, 2000774, 2000778, 2000781

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	108	97.8	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
EPA 180.1 Rev. 2.0	Turbidity					3.97	
EPA 200.7 Rev. 4.4	Calcium	104	104	104			1.21
	Iron	106	109	106	105		1.21
	Magnesium	111	106	107	106		0.340
	Potassium	105	103	102	100		0.644
	Sodium	101	102	103			0.835
	Zinc	103	106	107	105		2.06
EPA 200.8 Rev. 5.4	Arsenic	98.0	96.8	101	99.0		1.99
	Cadmium	99.5	97.7	103	99.6		3.63
	Chromium	96.0	93.7	93.6	95.3		1.70
	Copper	96.6	93.9	98.8	98.2		0.564
	Lead	96.2	94.0	99.5	99.3		0.198
	Nickel	98.3	95.2	99.5	99.0		0.434
	Silver	101	96.3	101	99.6		1.28
EPA 300.0 Rev. 2.1	Chloride	99.9	97.8	96.9		3.48	
	Sulfate	101	101	97.1		4.43	
EPA 350.1 Rev. 2.0	Ammonia-N	103	101	99.1			1.24
	Ammonia-N	99.9	96.2	99.5			3.02
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	110	105			4.11
EPA 353.2 Rev. 2.0	NO2NO3-N	100	96.2	97.8			1.16
EPA 365.1 Rev. 2.0	Total-P	97.4	95.6	98.2			2.64
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.3	98.1	99.3			0.731
	O-Phosphate-P	101	99.9	100			0.599
SM 2120 B	Color (true)					0.208	
SM 2320 B-97	Total Alkalinity	101				2.09	
SM 2540 C-97	TDS					9.84	
SM 4500 F-C-97	Fluoride	96.2	97.0	95.9			0.953
SM 5310 B-00	Organic Carbon	97.2	112	102			6.60

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2000771, 2000772, 2000777, 2000780
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2000798
EPA 200.8 Rev. 5.4 / E31780	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2000798
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2000771, 2000772, 2000777, 2000780
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2000771, 2000772, 2000777, 2000780
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2000773, 2000774, 2000778, 2000781
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2000773, 2000774, 2000778, 2000781
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2000773, 2000774, 2000778, 2000781
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2000773, 2000774, 2000778, 2000781
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2000775, 2000776, 2000779, 2000782
SM 2120 B / E31780	True Color measured at 450 nm	2000771, 2000772, 2000777, 2000780
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO3/L	2000771, 2000772, 2000777, 2000780
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2000771, 2000772, 2000777, 2000780

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2000771, 2000772, 2000777, 2000780
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2000771, 2000772, 2000777, 2000780
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2000773, 2000774, 2000778, 2000781

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/14/2018			06/14/2018 17:30	William L. Brown IV	2000771, 2000772, 2000777, 2000780
EPA 200.7 Rev. 4.4	06/14/2018	06/18/2018 14:45	Elliott D. Healy	06/19/2018 15:07	Betina Topolski	2000798
EPA 200.8 Rev. 5.4	06/14/2018	06/18/2018 14:45	Elliott D. Healy	06/19/2018 20:54	Robert K Palmer	2000798
	06/14/2018	06/18/2018 14:45	Elliott D. Healy	06/21/2018 18:29	Robert K Palmer	2000798
EPA 300.0 Rev. 2.1	06/14/2018			06/22/2018 14:59	Lipi Saha	2000771
	06/14/2018			06/22/2018 15:47	Lipi Saha	2000772
	06/14/2018			06/22/2018 15:59	Lipi Saha	2000777
	06/14/2018			06/22/2018 16:11	Lipi Saha	2000780
EPA 350.1 Rev. 2.0	06/14/2018			06/21/2018 11:58	Ping Hua	2000773
	06/14/2018			06/22/2018 09:04	Ping Hua	2000778
	06/14/2018			06/22/2018 09:05	Ping Hua	2000781
	06/14/2018			06/22/2018 09:19	Ping Hua	2000774
EPA 351.2 Rev. 2.0	06/14/2018	07/05/2018 09:15	Adrian Shelby	07/05/2018 16:39	Tanya Welborn	2000773
	06/14/2018	07/05/2018 09:15	Adrian Shelby	07/05/2018 16:40	Tanya Welborn	2000774
	06/14/2018	07/05/2018 09:15	Adrian Shelby	07/05/2018 16:42	Tanya Welborn	2000778
	06/14/2018	07/05/2018 09:15	Adrian Shelby	07/05/2018 16:43	Tanya Welborn	2000781
EPA 353.2 Rev. 2.0	06/14/2018			06/26/2018 12:17	Lauren Bottorf	2000773
	06/14/2018			06/26/2018 12:23	Lauren Bottorf	2000781
	06/14/2018			06/26/2018 12:26	Lauren Bottorf	2000774
	06/14/2018			06/26/2018 12:28	Lauren Bottorf	2000778
EPA 365.1 Rev. 2.0	06/14/2018	06/20/2018 11:10	Sima Feizi	06/21/2018 11:28	Lipi Saha	2000773
	06/14/2018	06/20/2018 11:10	Sima Feizi	06/21/2018 11:29	Lipi Saha	2000774
	06/14/2018	06/20/2018 11:10	Sima Feizi	06/21/2018 11:30	Lipi Saha	2000778
	06/14/2018	06/20/2018 11:10	Sima Feizi	06/22/2018 10:46	Beverly Y. Sanders	2000781
EPA 365.1 Rev. 2.0 dissolved	06/14/2018			06/14/2018 15:19	Megan Treadwell	2000775
	06/14/2018			06/14/2018 15:22	Megan Treadwell	2000782
	06/14/2018			06/14/2018 15:50	Megan Treadwell	2000776
	06/14/2018			06/14/2018 15:51	Megan Treadwell	2000779
SM 2120 B	06/14/2018			06/14/2018 15:30	Tanya Welborn	2000780
	06/14/2018			06/14/2018 15:38	Tanya Welborn	2000771
	06/14/2018			06/14/2018 15:39	Tanya Welborn	2000772, 2000777
SM 2320 B-97	06/14/2018			06/21/2018 10:58	Lauren Bottorf	2000771
	06/14/2018			06/21/2018 11:08	Lauren Bottorf	2000772
	06/14/2018			06/21/2018 11:19	Lauren Bottorf	2000777
	06/14/2018			06/21/2018 12:12	Lauren Bottorf	2000780
SM 2540 C-97	06/14/2018			06/15/2018 15:30	William L. Brown IV	2000771, 2000772, 2000777, 2000780
SM 2540 D-97	06/14/2018			06/15/2018 15:30	William L. Brown IV	2000771, 2000772, 2000777, 2000780
SM 4500 F-C-97	06/14/2018			06/21/2018 10:58	Lauren Bottorf	2000771
	06/14/2018			06/21/2018 11:08	Lauren Bottorf	2000772
	06/14/2018			06/21/2018 11:19	Lauren Bottorf	2000777

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
SM 4500 F-C-97	06/14/2018			06/21/2018 12:12	Lauren Bottorf	2000780
SM 5310 B-00	06/14/2018			06/19/2018 06:43	Megan Treadwell	2000773
	06/14/2018			06/19/2018 06:59	Megan Treadwell	2000774
	06/14/2018			06/19/2018 07:14	Megan Treadwell	2000778
	06/14/2018			06/19/2018 07:30	Megan Treadwell	2000781