

Chemical Analysis Report

SO-DIST-2018-06-28-02

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780


Event Description: **Telegraph**
Request ID: **RQ-2018-06-25-14**
Customer: **SO-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
FLDEP South District Office
2295 Victoria Ave. Suite 364
Fort Myers, FL 33901-3381
Attn: Nicole Reistad

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 26-JUL-2018 11:29

A handwritten signature in black ink, appearing to read "Colin Wright", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Metals and Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Telegraph Swamp South

Collection Date/Time: 06/27/2018 13:00

Field ID: G3SD0100

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2003764	EPA 180.1 Rev. 2.0	Turbidity	4.6		NTU	P347703	
	EPA 300.0 Rev. 2.1	Chloride	96		mg Cl/L	P347951	
		Sulfate	16		mg SO4/L	P347957	
		Color (true)	240		PCU	P347690	
	SM 2320 B-97	Total Alkalinity	102		mg CaCO3/L	P348098	
	SM 2540 C-97	TDS	389	A	mg/L	P347930	
	SM 2540 D-97	TSS	3	IA	mg/L	P347874	
	SM 4500 F-C-97	Fluoride	0.18		mg F/L	P348102	
2003765	EPA 350.1 Rev. 2.0	Ammonia-N	0.13		mg N/L	P348265	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	1.6		mg N/L	P348186	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.080		mg N/L	P347768	
	EPA 365.1 Rev. 2.0	Total-P	0.17		mg P/L	P347884	
	SM 5310 B-00	Organic Carbon	30		mg C/L	P348327	
2003766	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.10		mg P/L	P347708	
2003782	EPA 200.7 Rev. 4.4	Calcium	51.0		mg/L	P347853	
		Iron	1.56E+03		ug/L	P347853	
		Magnesium	10.1		mg/L	P347853	
		Potassium	2.9		mg/L	P347853	
		Sodium	51.2		mg/L	P347853	
	EPA 200.8 Rev. 5.4	Zinc	5.0	U	ug/L	P347853	
		Arsenic	1.10		ug/L	P347853	
		Cadmium	0.020	U	ug/L	P347853	
		Chromium	0.86	I	ug/L	P347853	
		Copper	0.32	I	ug/L	P347853	
		Lead	0.12	I	ug/L	P347853	
		Nickel	0.44	I	ug/L	P347853	
		Silver	0.010	U	ug/L	P347853	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Sulfate: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 351.2 Rev. 2.0: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P347703

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P347853

Component	Result	Code	Units
Calcium	0.075	U	mg/L
Iron	30	U	ug/L
Magnesium	0.040	U	mg/L
Potassium	0.30	U	mg/L
Sodium	0.50	U	mg/L
Zinc	5.0	U	ug/L

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P347853

Component	Result	Code	Units
Arsenic	0.050	U	ug/L
Cadmium	0.020	U	ug/L
Chromium	0.40	U	ug/L
Copper	0.20	U	ug/L
Lead	0.050	U	ug/L
Nickel	0.20	U	ug/L
Silver	0.010	U	ug/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347951

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347957

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P348265

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P348186

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report

Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347768

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347884

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P347708

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P347690

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P348098

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P347930

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P347874

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P348102

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P348327

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P347853

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Calcium	108		P	85 - 115
Iron	104		P	85 - 115
Magnesium	108		P	85 - 115
Potassium	102		P	85 - 115
Sodium	104		P	85 - 115
Zinc	99.9		P	85 - 115

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P347853

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Arsenic	96.3		P	85 - 115
Cadmium	95.8		P	85 - 115
Chromium	93.6		P	85 - 115
Copper	94.4		P	85 - 115
Lead	92.3		P	85 - 115
Nickel	96.1		P	85 - 115
Silver	92.6		P	85 - 115

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347951

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347957

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	96.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P348265

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	102		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P348186

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347768

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	96.0		P	90 - 110

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347884

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	96.6		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P347708

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.4		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P347690

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	101		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P348098

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P348102

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	99.8		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P348327

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	99.9		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P347853

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003929	Calcium	99.5		P	80 - 120
2003929	Iron	103		P	80 - 120
2003929	Magnesium	101		P	80 - 120
2003929	Potassium	101		P	80 - 120
2003929	Sodium	100		P	80 - 120
2003929	Zinc	99.2		P	80 - 120
2004004	Calcium	106	104	P/P	80 - 120
2004004	Iron	103	101	P/P	80 - 120
2004004	Magnesium	106	104	P/P	80 - 120
2004004	Potassium	105	103	P/P	80 - 120
2004004	Sodium	104	100	P/P	80 - 120
2004004	Zinc	101	99.8	P/P	80 - 120

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P347853

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003929	Arsenic	97.0		P	80 - 120
2003929	Cadmium	100		P	80 - 120
2003929	Chromium	93.6		P	80 - 120
2003929	Copper	91.0		P	80 - 120
2003929	Lead	92.0		P	80 - 120
2003929	Nickel	94.5		P	80 - 120
2003929	Silver	94.4		P	80 - 120
2004004	Arsenic	94.4	95.5	P/P	80 - 120
2004004	Cadmium	95.8	96.2	P/P	80 - 120
2004004	Chromium	94.1	95.7	P/P	80 - 120
2004004	Copper	93.0	93.9	P/P	80 - 120
2004004	Lead	93.2	93.4	P/P	80 - 120
2004004	Nickel	94.9	96.4	P/P	80 - 120
2004004	Silver	92.6	92.4	P/P	80 - 120

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347951

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003710	Chloride	93.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347957

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003710	Sulfate	92.5		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P348265

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003713	Ammonia-N	98.9	101	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347768

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003756	NO2NO3-N	95.5	96.5	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347884

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003756	Total-P	94.7	94.9	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P347708

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003772	O-Phosphate-P	91.0	91.4	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P348102

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003915	Fluoride	104	104	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P348327

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2003829	Organic Carbon	100		P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P347703

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003727	Turbidity	3.98	Sample	P	0 - 20

Reference Method: EPA 200.7 Rev. 4.4
Batch ID: P347853

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2004004	Calcium	1.38	Spike	P	0 - 20
2004004	Iron	1.72	Spike	P	0 - 20
2004004	Magnesium	1.65	Spike	P	0 - 20
2004004	Potassium	1.53	Spike	P	0 - 20
2004004	Sodium	2.43	Spike	P	0 - 20
2004004	Zinc	1.39	Spike	P	0 - 20

Reference Method: EPA 200.8 Rev. 5.4
Batch ID: P347853

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2004004	Arsenic	1.06	Spike	P	0 - 20
2004004	Cadmium	0.367	Spike	P	0 - 20
2004004	Chromium	1.61	Spike	P	0 - 20
2004004	Copper	0.951	Spike	P	0 - 20
2004004	Lead	0.216	Spike	P	0 - 20
2004004	Nickel	1.49	Spike	P	0 - 20
2004004	Silver	0.148	Spike	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P347951

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003710	Chloride	3.67	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P348265

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003713	Ammonia-N	1.53	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P348186

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2004969	Kjeldahl Nitrogen	1.31	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P347768

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003756	NO2NO3-N	0.533	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P347884

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003756	Total-P	0.197	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P347708

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003772	O-Phosphate-P	0.439	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P347690

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003752	Color (true)	0.556	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P348098

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003915	Total Alkalinity	0.730	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P347930

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003764	TDS	1.54	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P348102

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003915	Fluoride	0.0	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P348327

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2003829	Organic Carbon	0.0241	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A84913
Included Lab Sample IDs: 2003764

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	102	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A84917
Included Lab Sample IDs: 2003764

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.5	98.5	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A84918
Included Lab Sample IDs: 2003766

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	92.8	93.0	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A84938
Included Lab Sample IDs: 2003765

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	96.5	96.5	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A85001
Included Lab Sample IDs: 2003764

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	105	102	P/P	90 - 110
Sulfate	106	100	P/P	90 - 110

Reference Method: EPA 200.7 Rev. 4.4
Run ID: A85021
Included Lab Sample IDs: 2003782

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Calcium	104	102	P/P	90 - 110
Iron	102	100	P/P	90 - 110
Magnesium	103	101	P/P	90 - 110
Potassium	100	98.4	P/P	90 - 110
Sodium	102	101	P/P	90 - 110
Zinc	100	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A85050
Included Lab Sample IDs: 2003765

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	96.6	97.2	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 200.8 Rev. 5.4

Run ID: A85067

Included Lab Sample IDs: 2003782

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Arsenic	99.0	97.6	P/P	90 - 110
Cadmium	95.2	97.6	P/P	90 - 110
Chromium	96.9	97.5	P/P	90 - 110
Copper	98.8	97.0	P/P	90 - 110
Lead	95.3	95.8	P/P	90 - 110
Nickel	98.0	96.5	P/P	90 - 110
Silver	96.9	98.0	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A85091

Included Lab Sample IDs: 2003764

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	102	103	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A85091

Included Lab Sample IDs: 2003764

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	101	101	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A85152

Included Lab Sample IDs: 2003765

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	99.9	99.7	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A85175

Included Lab Sample IDs: 2003765

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	99.8	100	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A85182

Included Lab Sample IDs: 2003765

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	101	102	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements. Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery			Precision SMP	MS
			LCS				
EPA 180.1 Rev. 2.0	Turbidity					3.98	
EPA 200.7 Rev. 4.4	Calcium	108	99.5	106	104		1.38
	Iron	104	103	103	101		1.72
	Magnesium	108	101	106	104		1.65
	Potassium	102	101	105	103		1.53
	Sodium	104	100	104	100		2.43
	Zinc	99.9	99.2	101	99.8		1.39
EPA 200.8 Rev. 5.4	Arsenic	96.3	97.0	94.4	95.5		1.06
	Cadmium	95.8	100	95.8	96.2		0.367
	Chromium	93.6	93.6	94.1	95.7		1.61
	Copper	94.4	91.0	93.0	93.9		0.951
	Lead	92.3	92.0	93.2	93.4		0.216
	Nickel	96.1	94.5	94.9	96.4		1.49
	Silver	92.6	94.4	92.6	92.4		0.148
EPA 300.0 Rev. 2.1	Chloride	100	93.7			3.67	
	Sulfate	96.5	92.5				
EPA 350.1 Rev. 2.0	Ammonia-N	102	98.9	101			1.53
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104					1.31
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	96.0	95.5	96.5			0.533
EPA 365.1 Rev. 2.0	Total-P	96.6	94.7	94.9			0.197
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.4	91.0	91.4			0.439
SM 2120 B	Color (true)	101				0.556	
SM 2320 B-97	Total Alkalinity	103				0.730	
SM 2540 C-97	TDS					1.54	
SM 4500 F-C-97	Fluoride	99.8	104	104			0.0
SM 5310 B-00	Organic Carbon	99.9	100				0.0241

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2003764
EPA 200.7 Rev. 4.4 / E31780	Total Recoverable Metals analysis using ICP emission spectroscopy for aqueous samples supporting Clean Water Act Projects	2003782
EPA 200.8 Rev. 5.4 / E31780	Total Recoverable Metals analysis using ICP-MS for aqueous samples supporting Clean Water Act Projects	2003782
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2003764
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2003764
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2003765
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2003765
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2003765
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2003765
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2003766
SM 2120 B / E31780	True Color measured at 450 nm	2003764
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2003764
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2003764
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2003764
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2003764

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2003765

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	06/28/2018			06/28/2018 16:54	William L. Brown IV	2003764
EPA 200.7 Rev. 4.4	06/28/2018	07/02/2018 17:35	Elliott D. Healy	07/03/2018 16:45	Betina Topolski	2003782
EPA 200.8 Rev. 5.4	06/28/2018	07/02/2018 17:35	Elliott D. Healy	07/06/2018 15:02	Robert K Palmer	2003782
EPA 300.0 Rev. 2.1	06/28/2018			07/04/2018 00:10	Lipi Saha	2003764
EPA 350.1 Rev. 2.0	06/28/2018			07/11/2018 15:09	Ping Hua	2003765
EPA 351.2 Rev. 2.0	06/28/2018	07/11/2018 10:45	Adrian Shelby	07/12/2018 12:30	Joseph Suarez	2003765
EPA 353.2 Rev. 2.0	06/28/2018			06/29/2018 11:23	Lauren Bottorf	2003765
EPA 365.1 Rev. 2.0	06/28/2018	07/03/2018 10:50	Sima Feizi	07/06/2018 09:55	Beverly Y. Sanders	2003765
EPA 365.1 Rev. 2.0 dissolved	06/28/2018			06/28/2018 14:55	Megan Treadwell	2003766
SM 2120 B	06/28/2018			06/28/2018 15:59	Tanya Welborn	2003764
SM 2320 B-97	06/28/2018			07/07/2018 08:57	Dale L. Simmons	2003764
SM 2540 C-97	06/28/2018			06/29/2018 16:30	William L. Brown IV	2003764
SM 2540 D-97	06/28/2018			06/29/2018 16:30	William L. Brown IV	2003764
SM 4500 F-C-97	06/28/2018			07/07/2018 08:57	Dale L. Simmons	2003764
SM 5310 B-00	06/28/2018			07/11/2018 19:20	Megan Treadwell	2003765