

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID: 132237

RQ: 248-02-26-17

Project: SMP - Eglu

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 1/22/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.										P / F	L / F
ICV										P / F	L / F
CCV	J. P. H. S.	2/28/18	0719	23.2	8.546	8.51	100	-	1.01	⓪ / F	⓪ / F
CCV	J. P. H. S.	2/28/18	1400	23.2	8.546	8.54	100	-	-	⓪ / F	⓪ / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J. P. H. S.	2/28/18	0721	180124B	2/2019	1000	1000	⓪ / F	⓪ / F
ICV	↓	↓	0722	180124C	2/2019	10000	9696	⓪ / F	⓪ / F
CCV	J. P. H. S.	2/28/18	1402	180124A	2/2019	100	98	⓪ / F	⓪ / F
CCV	J. P. H. S.	2/28/18	1403	180124D	2/2019	50,000	47880	⓪ / F	⓪ / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. P. H. S.	2/28/18	0723	2610663	10/18	7.0	7.0	-30.5	⓪ / F	⓪ / F
Calibr.	↓	↓	0725	4712981	11/19	4.0	4.0	146.7	⓪ / F	⓪ / F
Calibr.	↓	↓	0727	2611A79	5/18	10.0	10.0	-200.7	⓪ / F	⓪ / F
ICV	↓	↓	0729	2611A79	5/18	10.0	10.0	-	⓪ / F	⓪ / F
CCV	J. P. H. S.	2/28/18	1405	4712981	11/19	4.0	4.0	-	⓪ / F	⓪ / F
CCV									P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 170.2, slope from 4 to 7 177.2 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes No If yes, list date of last quarterly depth verification 1/22/18

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air	J. P. H. S.	2/28/18	0733	0.00	0.00	⓪ / F	⓪ / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: