



Meter ID: Smartpro 11
Project: AWA

Method/Parameter		Analyst	MMDDYY	24 HH:MM	Acceptance Criteria			
DEP SOP FT 1400 Temperature		Initials	Date	Time	Previous CV	Bath Type	NIST Thermometer S/N	Meter Temp °C
CAL	ICV	CCV	MAINT					± 0.5 °C
DEP SOP FT 1600 Turbidity								
Meter ID:		Initials	Date	Time	1" or 2" Standard	Previous CV	NTU	Meter NTU
CAL	ICV	CCV	MAINT					Pass Fail
CAL	ICV	CCV	MAINT					± 5%
CAL	ICV	CCV	MAINT					Pass Fail
DEP SOP FT 1500 Dissolved Oxygen								
Initials	Date	Time	Changed membrane?	Saturation mg/L	Meter DO mg/L		% DO	Temp °C
CAL	ICV	CCV	MAINT					± 0.3 mg/l
CAL	ICV	CCV	MAINT					Pass Fail
CAL	ICV	CCV	MAINT					Pass Fail
DEP SOP FT 1200 Specific Conductance								
Initials	Date	Time	µmhos/cm	Expiration	ID	Meter Temp °C		µmhos/cm
CAL	ICV	CCV	MAINT					± 5%
CAL	ICV	CCV	MAINT					Pass Fail
CAL	ICV	CCV	MAINT					Pass Fail
CAL	ICV	CCV	MAINT					Pass Fail
CAL	ICV	CCV	MAINT					Pass Fail
DEP SOP FT 1100 pH								
Initials	Date	Time	SU	Expiration	ID	Meter pH mV		SU
CAL	ICV	CCV	MAINT					± 0.2 SU
CAL	ICV	CCV	MAINT					Pass Fail
CAL	ICV	CCV	MAINT					Pass Fail
CAL	ICV	CCV	MAINT					Pass Fail
CAL	ICV	CCV	MAINT					Pass Fail
DEP SOP FT 1000 ORP								
Initials	Date	Time	mV	Expiration	ID	Meter Temp °C		mV
CAL	ICV	CCV	MAINT					± 20 mV
CAL	ICV	CCV	MAINT					Pass Fail
CAL	ICV	CCV	MAINT					Pass Fail

Notes: