

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

132237

RQ:

208-06-18-20

Project:

SMP

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration

or

Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

4/19/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.										P / F	L / F
ICV										P / F	L / F
CCV	NM	6/18/18	0730	26.03	8.114	8.08	99.8			P / F	L / F
CCV	J. Pass	6/19/18	0705	26.41	8.055	8.06	100	-	-	P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	NM	6/18/18	0732	180124B	2/19	1000	1000	P / F	L / F
ICV	↓	↓	0735	180329	4/19	100	104	P / F	L / F
CCV	J. Pass	6/19/18	0707	180124B	2/19	1000	998	P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	NM	6/18/18	0740	1710655	10/19	7.0	7.0		P / F	L / F
Calibr.	↓	↓	0742	4712981	11/19	4.0	4.0		P / F	L / F
Calibr.	↓	↓	0744	2709903	3/19	10.0	10.0		P / F	L / F
ICV	↓	↓	0745	4712981	11/19	4.0	4.03	137.5	P / F	L / F
CCV	J. Pass	6/19/18	0709	2709903	3/19	10.0	10.0	-	P / F	L / F
CCV									P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes

No If yes, list date of last quarterly depth verification

4/19/18

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

758.3

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

154179

RQ:

Project:

initial setup

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

6/18/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	MIKE KING	6/18/18	1500	24.0	8.418	8.40	99.9		1.015T	☒ P / ☐ F	☒ L / ☐ F
ICV	m	6/18/18	1510	23.7	8.46	8.45	99.8			☒ P / ☐ F	☒ L / ☐ F
CCV										P / F	L / F
CCV										P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	m	6/18	1515	180124C	02/2019	10000	10000	☒ P / ☐ F	☒ L / ☐ F
ICV	m	6/18	1516	180324	04/2019	100	19.1	☒ P / ☐ F	☒ L / ☐ F
CCV								P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	m	6/18	1520	1710655	10/2019	7.0	7.0	75.0	☒ P / ☐ F	☒ L / ☐ F
Calibr.	↓		1521	4712981	11/2019	4.0	4.0	177.2	☒ P / ☐ F	☒ L / ☐ F
Calibr.	↓		1522	2709963	03/2019	10.0	10.0	174.4	☒ P / ☐ F	☒ L / ☐ F
ICV	↓	↓	1525	1710655	10/2019	7.0	7.0	77.5	☒ P / ☐ F	☒ L / ☐ F
CCV									P / F	L / F
CCV									P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;mV pH 4 Range $+180 \pm 50$;mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Is meter equipped with depth sensor: ☒ Yes ☐ No If yes, list date of last quarterly depth verification _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: