

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: 118163

RQ: 2018-08-13-34

Project: Smp LVI

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

5/30/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.										P / F	L / F
ICV										P / F	L / F
CCV	<u>J. Slode</u>	<u>8/21/18</u>	<u>0830</u>	<u>24.75</u>	<u>8.35</u>	<u>8.15</u>	<u>98.2</u>		<u>1.03693</u>	<u>(P)</u> F	<u>(E)</u> F
CCV	<u>F. Butera</u>	<u>8/21/18</u>	<u>17:30</u>	<u>23.7</u>	<u>8.47</u>	<u>8.46</u>	<u>100.0</u>			<u>(P)</u> F	<u>(L)</u> F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	<u>J. Slode</u>	<u>8/21/18</u>	<u>0832</u>	<u>160517-B</u>	<u>May 2019</u>	<u>1000</u>	<u>1000</u>	<u>(P)</u> F	<u>(L)</u> F
ICV	<u>J. Slode</u>	<u>8/21/18</u>	<u>0834</u>	<u>180329</u>	<u>Apr 2019</u>	<u>100</u>	<u>103</u>	<u>(P)</u> F	<u>(L)</u> F
CCV	<u>F. Butera</u>	<u>8/21/18</u>	<u>17:34</u>	<u>180512-B</u>	<u>5/2019</u>	<u>LP40</u>	<u>1009</u>	<u>(P)</u> F	<u>(L)</u> F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	<u>J. Slode</u>	<u>8/21/18</u>	<u>0836</u>	<u>1710655</u>	<u>OCT 2019</u>	<u>7.0</u>	<u>7.00</u>	<u>-8.3</u>	<u>(P)</u> F	<u>(L)</u> F
Calibr.	<u>J. Slode</u>	<u>8/21/18</u>	<u>0838</u>	<u>2803729</u>	<u>Feb 2020</u>	<u>4.0</u>	<u>4.00</u>	<u>166.2</u>	<u>(P)</u> F	<u>(L)</u> F
Calibr.	<u>J. Slode</u>	<u>8/21/18</u>	<u>0840</u>	<u>2709903</u>	<u>Mar 2019</u>	<u>10.0</u>	<u>10.01</u>	<u>-184.1</u>	<u>(P)</u> F	<u>(L)</u> F
ICV	<u>J. Slode</u>	<u>8/21/18</u>	<u>0842</u>	<u>2803729</u>	<u>Feb 2020</u>	<u>4.0</u>	<u>4.07</u>	<u>161.7</u>	<u>(P)</u> F	<u>(L)</u> F
CCV									P / F	L / F
CCV	<u>F. Butera</u>	<u>8/21/18</u>	<u>17:38</u>	<u>17110655</u>	<u>10/2019</u>	<u>7.0</u>	<u>7.05</u>		<u>(P)</u> F	<u>(L)</u> F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 175.8, slope from 4 to 7 171.5 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes (X) No () If yes, list date of last quarterly depth verification

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this page.

Meter ID:

132237

RQ:

2018-08-20-14

Project:

SMP Laurel Hill W

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

6/18/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.										P / F	L / F
ICV										P / F	L / F
CCV	J. P. H. S.	8/21/18	0715	25.1	8.218	8.36	101	—	1.02	P / F	L / F
CCV	J. P. H. S.	8/21/18	1530	25.4	8.203	8.30	101	—	—	P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J. P. H. S.	8/21/18	0717	180517B	5/2019	1000	1000	P / F	L / F
ICV	↓	↓	0719	180329	4/2019	100	104	P / F	L / F
CCV	J. P. H. S.	8/21/18	1532	180329	4/2019	100	103	P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. P. H. S.	8/21/18	0721	1710655	10/2019	7.0	7.0	41.2	P / F	L / F
Calibr.	↓	↓	0723	2803729	2/2020	4.0	4.0	136.8	P / F	L / F
Calibr.	↓	↓	0725	2709903	3/2019	10.0	10.0	210.2	P / F	L / F
ICV	↓	↓	0727	2709903	3/2019	10.0	10.0	—	P / F	L / F
CCV	J. P. H. S.	8/21/18	1534	2803729	2/2020	4.0	4.1	—	P / F	L / F
CCV									P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 169.0 , slope from 4 to 7 178.0 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification 4/19/18

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air	J. P. H. S.	8/21/18	0729	0.00	0.00	P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

154179

RQ:

2018-08-2015

Project:

SMP

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 6/18/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.											P / F	L / F
ICV											P / F	L / F
CCV	M	8/21	0640	24.9	767.1	8.39	8.44	101.5			(P) F	(L) F
CCV	M	8/21	1405	22.0		8.86	8.85	101.3			(P) F	(L) F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	M	8/21	0642	180124C	02/2019	10000	10003	(P) F	(L) F
ICV	↓	↓	0643	180124B	05/2019	1000	1005	(P) F	(L) F
CCV	M	8/21	1410	180517B	04/2019	100	102	(P) F	(L) F
CCV				180324				P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	M	8/21	0645	170655	10/19	7.0	24.8	7.01	-12.3	(P) F	(L) F
Calibr.	↓	↓	0646	2803729	02/20	4.0	24.8	4.02	171.2	(P) F	(L) F
Calibr.	↓	↓	0647	2709903	03/19	10.0	24.8	10.01	-182.7	(P) F	(L) F
ICV	↓	↓	0650	2803729	02/20	4.0		4.07		(P) F	(L) F
CCV	M	8/21	1415	2709903	03/19	10.0		10.05		(P) F	(L) F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 $\pm$ 50; mV pH 4 Range +180 $\pm$ 50; mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 6/26/18

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	M	8/21	0655	0.000	0.000	(P) F	(L) F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: