

Depth Verification Regional Operation Centers

SOP - S&T Sampling Manual and ROC Training Manual.

Report two decimal places for electronic devices. Report one decimal place for manual devices.

Numbers ≤ 4 , are rounded down; numbers ≥ 5 are rounded up.

QUARTERLY VERIFICATION OF ELECTRONIC DEVICES (SONDE, SONAR DEVICE, ETC.)

Meter / Device ID#: 132237 Date of Last Verification: 10/10
Date: 10/16/17 Time: 0903 ETZ / CTZ Verification Location: Cleburn Mite Perdio River
Person Performing Verification: S. Shole
Reference Device: Graduated Bucket / Metal Measuring Tape / Meter Stick / Other Secchi Disk
Depth measurements: Reference Device: 3.9 m ; Device Being Tested: 3.92 m
Result: Pass / Fail (acceptance Criteria 10%)

Meter / Device ID#: _____ Date of Last Verification: _____
Date: _____ Time: _____ ETZ / CTZ Verification Location: _____
Person Performing Verification: _____
Reference Device: Graduated Bucket / Metal Measuring Tape / Meter Stick / Other _____
Depth measurements: Reference Device: _____ m ; Device Being Tested: _____ m
Result: Pass / Fail (acceptance Criteria 10%)

Comments: _____

6 MONTH VERIFICATION OF MANUAL DEVICES (SECCHI DISK, WEIGHTED LINE, ETC.)

Secchi/Weighted Line ID#: Secchi-122 Date of Last Verification: 11/16/16
Date: 10/13/17 Time: 1200 ETZ / CTZ Verification Location: _____ Lab _____
Person Performing Verification: S. Shole
Reference Device: Metal Measuring Tape / Meter Stick / Other _____
Incremental markings of 0.1 m checked: YES / NO Result: Pass / Fail (acceptable criteria 10%)
Total length of line (up to anticipated depth encountered in field) checked: YES / NO
Total Length: indicated by line markings 10.3 m ; measured by reference device 10 m
Result: Pass / Fail (acceptable criteria of 5%) Markings redone: YES / NO

Secchi/Weighted Line ID#: _____ Date of Last Verification: _____
Date: _____ Time: _____ ETZ / CTZ Verification Location: _____ Lab _____
Person Performing Verification: _____
Reference Device: Metal Measuring Tape / Meter Stick / Other _____
Incremental markings of 0.1 m checked: YES / NO Result: Pass / Fail (acceptable criteria 10%)
Total length of line (up to anticipated depth encountered in field) checked: YES / NO
Total Length: indicated by line markings _____ m ; measured by reference device _____ m
Result: Pass / Fail (acceptable criteria of 5%) Markings redone: YES / NO

Comments: _____