

Chemical Analysis Report

NW-DIST-2018-04-25-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **BIOASSESSMENT**
Request ID: **RQ-2018-04-09-35**
Customer: **NW-DIST**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
160 W Government St
Suite 208A
Pensacola, FL 32502-5794
Attn: Michael King

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 11-MAY-2018 11:07

A handwritten signature in black ink, appearing to read "Colin Wright", is displayed on a light-colored rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Swift Creek upstream of 285

Collection Date/Time: 04/24/2018 13:45

Field ID: G3NW0149

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1987050	EPA 180.1 Rev. 2.0	Turbidity	0.95		NTU	P344062	
	EPA 300.0 Rev. 2.1	Chloride	3.9		mg Cl/L	P344281	
		Sulfate	2.4		mg SO4/L	P344285	
	SM 2120 B	Color (true)	17		PCU	P344046	
	SM 2320 B-97	Total Alkalinity	3.0		mg CaCO3/L	P344453	RPD
	SM 2540 C-97	TDS	15	U	mg/L	P344546	
	SM 2540 D-97	TSS	3	I	mg/L	P344676	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P344457	
1987051	EPA 350.1 Rev. 2.0	Ammonia-N	0.004	I	mg N/L	P344430	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.15	I	mg N/L	P344362	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.081		mg N/L	P344122	
	EPA 365.1 Rev. 2.0	Total-P	0.005	I	mg P/L	P344136	
	SM 5310 B-00	Organic Carbon	1.8		mg C/L	P344443	
1987052	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.007	I	mg P/L	P344050	

Ref. Method and Comment:

EPA 300.0 Rev. 2.1: Chloride: Batch matrix spike recoveries for one of two spiked samples are unavailable because of high analyte concentration in the QC sample.

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P344062

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P344281

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P344285

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P344430

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P344362

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P344122

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P344136

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P344050

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P344046

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P344453

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P344546

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P344676

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P344457

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P344443

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P344281

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P344285

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	94.1		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P344430

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P344362

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P344122

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P344136

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P344050

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.0		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P344453

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P344457

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	96.5		P	91 - 105

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
 Batch ID: P344443

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	94.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P344281

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1987142	Chloride	98.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P344285

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1987017	Sulfate	94.0		P	90 - 110
1987142	Sulfate	97.4		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P344430

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1987018	Ammonia-N	100	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P344362

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1987073	Kjeldahl Nitrogen	106	107	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P344122

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1987018	NO2NO3-N	99.5		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P344136

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1987018	Total-P	99.0	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P344050

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1986992	O-Phosphate-P	94.8	95.4	P/P	90 - 110

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P344457

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1986644	Fluoride	94.2	94.7	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P344443

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1986923	Organic Carbon	96.7	95.6	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P344062

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1986990	Turbidity	6.01	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P344281

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1987017	Chloride	0.798	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P344285

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1987017	Sulfate	0.580	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P344430

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1987018	Ammonia-N	0.951	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P344362

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1987073	Kjeldahl Nitrogen	1.17	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P344122

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
--------------------------	-----------	-----------	-------------------	-----------	----------------

Quality Assurance Report Precision

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P344122

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1987018	NO2NO3-N	0.188	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P344136

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1987018	Total-P	1.26	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P344050

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1986992	O-Phosphate-P	0.457	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P344046

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1987017	Color (true)	1.73	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P344453

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1986644	Total Alkalinity	4.08	Sample	F	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P344546

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1987017	TDS	2.25	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P344457

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1986644	Fluoride	0.479	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P344443

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1986923	Organic Carbon	0.754	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A83415

Included Lab Sample IDs: 1987050

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	100	100	P/P	90 - 110
Sulfate	100	101	P/P	90 - 110

Reference Method: SM 2120 B

Run ID: A83456

Included Lab Sample IDs: 1987050

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	97.4	98.9	P/P	85 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A83457

Included Lab Sample IDs: 1987052

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.9	98.9	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A83461

Included Lab Sample IDs: 1987050

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	98.0	98.5	P/P	95 - 105

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A83479

Included Lab Sample IDs: 1987051

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A83587

Included Lab Sample IDs: 1987051

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A83590

Included Lab Sample IDs: 1987051

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	98.3	96.7	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A83591

Included Lab Sample IDs: 1987051

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	102	99.0	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A83593

Included Lab Sample IDs: 1987050

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	99.4	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A83593

Included Lab Sample IDs: 1987050

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	99.0	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A83607

Included Lab Sample IDs: 1987051

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	99.8	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				6.01	
EPA 300.0 Rev. 2.1	Chloride	100	98.0		0.798	
	Sulfate	94.1	94.0	97.4	0.580	
EPA 350.1 Rev. 2.0	Ammonia-N	101	100	101		0.951
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104	106	107		1.17
EPA 353.2 Rev. 2.0	NO2NO3-N	99.5	99.5			0.188
EPA 365.1 Rev. 2.0	Total-P	103	99.0	101		1.26
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.0	94.8	95.4		0.457
SM 2120 B	Color (true)				1.73	
SM 2320 B-97	Total Alkalinity	102			4.08	
SM 2540 C-97	TDS				2.25	
SM 4500 F-C-97	Fluoride	96.5	94.2	94.7		0.479
SM 5310 B-00	Organic Carbon	94.8	96.7	95.6		0.754

Reference Method Descriptions

Method / DoH Cert #

Description

Associated Samples

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1987050
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1987050
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1987050
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1987051
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1987051
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1987051
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1987051
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1987052
SM 2120 B / E31780	True Color measured at 450 nm	1987050
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1987050
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1987050
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1987050
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1987050
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1987051

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	04/25/2018			04/25/2018 16:22	William L. Brown IV	1987050
EPA 300.0 Rev. 2.1	04/25/2018			04/27/2018 21:49	Lipi Saha	1987050
EPA 350.1 Rev. 2.0	04/25/2018			04/30/2018 13:41	Ping Hua	1987051
EPA 351.2 Rev. 2.0	04/25/2018	05/01/2018 09:45	Adrian Shelby	05/02/2018 09:22	Joseph Suarez	1987051
EPA 353.2 Rev. 2.0	04/25/2018			04/26/2018 12:22	Lauren Bottorf	1987051
EPA 365.1 Rev. 2.0	04/25/2018	04/26/2018 12:50	Sima Feizi	05/01/2018 13:05	Beverly Y. Sanders	1987051
EPA 365.1 Rev. 2.0 dissolved	04/25/2018			04/25/2018 14:52	Megan Treadwell	1987052
SM 2120 B	04/25/2018			04/25/2018 14:16	Tanya Welborn	1987050
SM 2320 B-97	04/25/2018			05/02/2018 08:41	Dale L. Simmons	1987050
SM 2540 C-97	04/25/2018			04/27/2018 16:08	Yijie Li	1987050
SM 2540 D-97	04/25/2018			04/27/2018 16:08	Yijie Li	1987050
SM 4500 F-C-97	04/25/2018			05/02/2018 08:41	Dale L. Simmons	1987050
SM 5310 B-00	04/25/2018			04/30/2018 19:40	Megan Treadwell	1987051