



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2018

Data Qualified?

Yes / No

WIN Station Name: Gully Branch @ Gully Branch Rd

Date: 02/14/18
(mm/dd/yyyy)

WIN/ Field ID: G2TLHR0006

WBID: 1242

Latitude: 29.959945

County: Franklin

ORG ID: 21FLTLHR

Longitude: -84.718339

Receiving Body of Water: Crooked River

RQ-2018-01-22-51

Sampling Team Members

J. Burchett
J. Tolbert

WQ Measurements
Sample Collection
Documentation
Preservation
Preservation Check

Sampling Equipment

☒ Sample Container ☐ Van Dorn ☐ Pole
☐ Carboy ☒ Syringe
☐ Dipper ☐ Other

Depth Measured with
Equipment ID

Collection Method

☒ Wading ☐ Canoe/Kayak
☐ From Shore ☐ Electric Motor
☐ Other ☐ Gasoline Motor

Water Level: Dry/ Pooled/ Low / Normal / High / Flooded

Meter ID# E203

Matrix: Fresh / Marine / DI

Photos: Yes / No

Flow: No Flow / Intertidal / Flowing / NA

Tide: Rising/ Falling/ Slack/ NA

Tide Times: High Low

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)
EST	1105	0.24	0.12	0.24	5.92	61.6	3.71	0.04	78	17.25
CEN				(3)						

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4	SA7341020	12/7/17	☒ <2
Metals	☒ W-ICPMS ☒ W-ICP	☒ Ice ☒ HNO3	NA734K00	12/7/18	☒ <2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter	☐ Ice			
Chlorophyll	☒ CHLSUITE-W	☐ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-IC / W-COLOR / W-SO4-IC / W-TDS	☐ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci	☐ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-SUC-AA-R ☐ W-UHERB-AA ☐ W-AHERB-AA	☐ Ice			
Pesticides	☒ W-PSNP-TQ ☒ W-UHERB-AA ☒ W-AHERB-AA ☒ W-CARB-AA ☒ W-PCL-SQ-R ☒ W-PYR-AA-R ☒ W-GLYPH ☒ W-SUC-AA-R	☒ Ice ☐ Other			

Notes:

Kit: C-6 w/ pest

Place Label Here

Signature: [Signature]

Date: 02/14/18

Enter Field Parameters, Verify Station ID In LIMS DMT: JB 2/15/18

QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data In WIN:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID:

Exo3

RQ:

Project:

SMP 2016

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

12/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Burkett	2/14/18	800	19.99	9.09	9.10	100	-	1.04	P/F	L/F
ICV	✓	✓	80	20.4	9.02	9.00	100	-	-	P/F	L/F
CCV	✓	✓	1615	23.7	8.47	8.46	100	-	-	P/F	L/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J. Burkett	2/14/18	802	180116B	06/19	1000	1000	P/F	L/F
ICV	✓	✓	812	270781	09/19	100	100	P/F	L/F
CCV	✓	✓	1617	2710781	09/19	100	98	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Burkett	2/14/18	804	171121A	06/19	7.0	7.0	-232	P/F	L/F
Calibr.	✓	✓	806	180116D	06/19	4.0	4.0	153.0	P/F	L/F
Calibr.	✓	✓	808	171121B	06/19	10.0	10.0	-193.9	P/F	L/F
ICV	✓	✓	814	171121B	06/19	10.0	10.0	-195.6	P/F	L/F
CCV	✓	✓		180116D	06/19	4.0	4.0	151.7	P/F	L/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 170.7, slope from 4 to 7 176.2 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes No If yes, list date of last quarterly depth verification 12/17

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

last revised Jan 3, 2018

503

[illegible]

Date/Time

Date/Time 8:57:11

Group: A	Bottle Type:	P-1L	# of Bottles: 8	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 8	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WO-THI	# of Bottles: 8	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	ECOLL-18QT						
Group: A	Bottle Type:	P-500ML	# of Bottles: 8	Preservative:	HNO3	Enter Number of Bottles Sent to Lab	2
	W-ICP						
	W-ICPMS						
Group: A	Bottle Type:	P-500ML	# of Bottles: 8	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	2
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 8	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	2
	W-PO4-F						
Group: B	Bottle Type:	BG-1L	# of Bottles: 16	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	W-AHERB-AA						
	W-GLYPH						
	W-PYR-AA-R						
	W-SUC-AA-R						
	W-UHERB-AA						
Group: B	Bottle Type:	BG-1L	# of Bottles: 8	Preservative:	CH2ClCOOH Pre-Preserved	Enter Number of Bottles Sent to Lab	2
	W-CARB-AA						
Group: B	Bottle Type:	BG-500ML	# of Bottles: 18	Preservative:	ICE	Enter Number of Bottles Sent to Lab	

Group: B	Bottle Type:	BG-500ML	# of Bottles: 18	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	W-CT-CI-R						
	W-PCL-SQ-R						
Group: B	Bottle Type:	BG-1L	# of Bottles: 18	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	W-PSNP-TQ						