



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

Data Qualified?

Yes / No

January 2018

WIN Station Name: Crooked River @ Post 46

Date: 4/4/18
(mm/dd/yyyy)

WIN/ Field ID: G2TLHR0038

WBID: 1251

Latitude: 29.903

(Decimal Degrees)

County: Franklin

ORG ID: 21FLTLHR

Longitude: 84.6515

(Decimal Degrees)

Receiving Body of Water: Carrabelle River

RQ-2018- 03-26-57

Sampling Team Members						WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment				
Jade Busthett Jessie Tolbert						☒	☒	☒	☒	☒	☒ Sample Container	☐ Van Dorn	☐ Pole		
											☐ Carboy	☒ Syringe			
											☐ Dipper	☐ Other			
											Depth Measured with Equipment ID				
											Collection Method				
											☒ Wading	☐ Canoe/Kayak			
											☐ From Shore	☐ Electric Motor			
											☐ Other	☐ Gasoline Motor			
Water Level: Dry/ Pooled/ Low / <u>Normal</u> / High / Flooded						Meter ID# YSI EXO 3				Matrix: <u>Fresh</u> / Marine / DI					
Photos: Yes / <u>No</u>		Flow: No Flow / Intestinal / <u>Flowing</u> / NA		Tide: Rising/ Falling/ Slack/ <u>NA</u>		Tide Times: High		Low							
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)					
EST	11:05	1.4	0.3	0.8	4.25	49.5	6.56	0.12	248	22.6					
CEN															
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:															
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:															
Parameter Suite	Parameter Methods					Preservation	Lot#	Expiration	pH Check						
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					☒ Ice ☐ H2SO4	7341070	12/7/18	☒ <2						
Metals	☐ W-ICPMS ☐ W-ICP					☐ Ice ☐ HNO3			☐ <2						
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter					☐ Ice									
Chlorophyll	☒ CHL-SUITE-W					☒ Ice									
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CI-IC / W-COLOR / W-SO4-IC / W-TDS					☒ Ice									
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY					☐ Ice									
Macroinvertebrates	☐ MI-FW-QLDC					☐ Formalin									
Periphyton	☐ ALGAE-ID					☐ Ice									
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci					☒ Ice									
Molecular	☐ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD					☐ Ice									
Tracers	☐ W-SUC-AA-R ☐ W-UHERB-AA ☐ W-AHERB-AA					☐ Ice									
Pesticides	☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-CARB-AA					☐ Ice									
	☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R					☐ Other									

Notes:
A-b

Field/WIN ID

Location

G2TLHR0038

CROOKED RIVER @ POST 46

Signature:

Date:

Enter Field Parameters, Verify Station ID In LIMS DMT:

QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data In WIN:

last revised Jan 25, 2018

Field Report Prepared By:

Sampling Agency:

Robert
88341

Relinquished By: <i>Tobias Buckett</i>	Date/Time <i>4/11/18 1532</i>	Shipping Method <i>FD</i>	Number of Coolers Shipped: <i>1</i>	Received By: <i>[Signature]</i>	Date/Time <i>4/11/18 1533</i>
---	----------------------------------	------------------------------	--	------------------------------------	----------------------------------

Group: A	Bottle Type:	P-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WO-THI	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ECOLI-18QT						
Group: A	Bottle Type:	P-500ML	# of Bottles: 5	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 5	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	1
	W-PO4-F						

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: Ex03

RQ:

Project: SMP 2014

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 12/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Burtchett	4/4/18	805	21.5	8.83	8.83	100	-	1.08	P/F	L/F
ICV	↓	↓	815	21.3	8.8	8.87	100	-	-	P/F	L/F
CCV	↓	↓	1537	21.1	8.90	8.91	100	-	-	P/F	L/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J. Burtchett	4/4/18	807	18016B	06/19	1000	1000	P/F	L/F
ICV	↓	↓	818	270781	09/19	100	95	P/F	L/F
CCV	↓	↓	1539	2710781	09/19	100	95	P/F	L/F
CCV			1541	18016B	01/19	1000	996	P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Burtchett	4/4/18	809	171121A	06/19	7.0	7.0	-30.9	P/F	L/F
Calibr.		↓	811	18016D	06/19	4.0	4.0	145.9	P/F	L/F
Calibr.	↓	↓	813	180221E	09/19	10.0	10.0	-202.3	P/F	L/F
ICV	↓	↓	819	180221E	09/19	10.0	10.0	-202.3	P/F	L/F
CCV	↓	↓	1543	18016D	01/19	4.0	4.0	146.0	P/F	L/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 171.4, slope from 4 to 7 1 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification 12/17

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

