



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

Data Qualified?

Yes / No

January 2018

WIN Station Name: Saul CreekDate: 10/31/2016
(mm/dd/yyyy)WIN/ Field ID: G2TLHR0014WBID: 1271Latitude: 29.7896302

(Decimal Degrees)

County: GulfORG ID: 21FLTLHRLongitude: 85.0505733

(Decimal Degrees)

Receiving Body of Water: Apalachicola River

RQ-2018-09-17-57

Sampling Team Members						WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
Jon Woods						X	X			
Jade Butcher						X	X	X	X	

Sampling Equipment				
☒ Sample Container	☒ Van Dorn	☐ Pole		
☐ Carboy	☒ Syringe			
☐ Dipper	☐ Other			
Depth Measured with <u>Sonac</u>				
Equipment ID <u>#1</u>				

Collection Method				
☐ Wading	☐ Canoe/Kayak			
☐ From Shore	☐ Electric Motor			
☐ Other	☒ Gasoline Motor			

Water Level: Dry/ Pooled/ Low / <u>Normal</u> / High / Flooded						Meter ID# <u>154177</u>		Matrix: <u>Fresh</u> / Marine / DI		
Photos: Yes <u>No</u>		Flow: No Flow / Intestinal / <u>Flowing</u> / NA		Tide: Rising / <u>Falling</u> / Slack/ NA		Tide Times: High Low				
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPTH)	Sp COND (µmhos/cm)	Temp. (C°)
EST	1250	6.5	0.3	1.1	2.99	34.3	7.25	0.07	156	21.4
CEN	1253		6.0		0.54	6.1	6.79	0.25	526	20.5
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:										
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:										
Parameter Suite	Parameter Methods					Preservation	Lot#	Expiration	pH Check	
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					☒ Ice ☒ H ₂ SO ₄	<u>619 7090</u>	<u>7/16/19</u>	☒ <2	
Metals	☐ W-ICPMS ☐ W-ICP					☐ Ice ☐ HNO ₃			☐ <2	
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter					☒ Ice				
Chlorophyll	☒ CHLSUITE-W					☒ Ice				
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS					☒ Ice				
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY					☐ Ice				
Macroinvertebrates	☐ MI-FW-QLDC					☐ Formalin				
Periphyton	☐ ALGAE-ID					☐ Ice				
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci					☒ Ice				
Molecular	☐ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD					☐ Ice				
Tracers	☐ W-SUC-AA-R ☐ W-UHERB-AA ☐ W-AHERB-AA					☐ Ice				
Pesticides	☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-CARB-AA					☐ Ice				
	☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R					☐ Other				

Place Preservative Sticker(s) Here
(If Available)

Notes:

A-b

Field/WIN ID →

Location ↓



G2TLHR0014



SAUL CREEK @ MOUTH

Signature:

Date:

Enter Field Parameters / Verify Station ID in LIMS DMT:

QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data in WIN:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 154177

RQ-

Project: SMP 2018

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 9/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Burchett	10/31/18	759	19.9	76.4	9.11	9.13	100.2	-	0.99	P/F	L/F
ICV	↓	↓	809	19.1	76.3	9.26	9.24	99.8	-	-	P/F	L/F
CCV	↓	↓	1517	25.9	759.3	8.13	8.16	100.4	-	-	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J. Burchett	10/31/18	801	180621C	07/19	1200	1200	P/F	L/F
ICV	↓	↓	811	1805581	05/20	100	102	P/F	L/F
CCV	↓	↓	1519	180621A	7/2019	10,000	100087	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Burchett	10/31/18	803	180621C	01/20	7.02	19.0	7.03	-24.2	P/F	L/F
Calibr.	↓	↓	805	1805103	05/19	4.00	19.1	4.00	150.4	P/F	L/F
Calibr.	↓	↓	807	1804030	10/19	10.08	19.3	10.07	-145.9	P/F	L/F
ICV	↓	↓	813	1804030	10/19	10.08	18.9	10.06	-147.6	P/F	L/F
CCV	↓	↓	1521	1805103	05/2019	9.01	22.3	4.03	148.4	P/F	L/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU;

mV pH 7 Range 0 $\pm$ 50;

mV pH 4 Range +180 $\pm$ 50;

mV pH 10 Range -180 $\pm$ 50;

If mV are recorded: slope from 7 to 10 1, slope from 4 to 7 1 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 9/18

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J. Burchett	10/31/18	757	0.00	0.00	P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: RQ- Project:

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.											P / F	L / F
ICV											P / F	L / F
CCV											P / F	L / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. **Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.								P / F	L / F
ICV								P / F	L / F
CCV								P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P / F	L / F
Calibr.						4.				P / F	L / F
Calibr.						10.				P / F	L / F
ICV										P / F	L / F
CCV										P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) **Secchi Disk Line / Sonar** Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Sampling Agency:

Relinquished By:	Date/Time	Shipping Method	Number of Coolers Shipped:	Received By:	Date/Time
Woods	10/31/18: 1515	HD	1	[Signature]	10/31/18 25:19

Group: A	Bottle Type:	P-1L	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WO-THI	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	ECOLI-18QT						
Group: A	Bottle Type:	P-500ML	# of Bottles: 4	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 4	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	4
	W-PO4-F						