



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION
STRATEGIC MONITORING PROGRAM FIELD SHEET

January 2018

Data Qualified?

Yes / No

WIN Station Name: Horseshoe Creek 100m upstream from Intercoastal/Lake Wimico

Date: 1/29/2016
(mm/dd/yyyy)

WIN/ Field ID: G2TLHR0033

WBID: 1272

Latitude: 29.843985

County: Gulf

ORG ID: 21FLTLHR

Longitude: 85.212303

Receiving Body of Water: Intercoastal/Lake Wimico

RQ-2018- 11-26-69

Sampling Team Members				WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment			
Jan Woods Ben Hasler				X	X	X	X	X	☒ Sample Container	☐ Van Dorn	☐ Pole	
									☐ Carboy	☒ Syringe	☐ Other	
									☐ Dipper	☐ Other		
									Depth Measured with <u>SOMC</u>			
									Equipment ID <u>#1</u>			
									Collection Method			
									☐ Wading	☐ Canoe/Kayak		
									☐ From Shore	☐ Electric Motor		
									☐ Other	☒ Gasoline Motor		
Water Level: Dry/ Pooled/ Low / <u>Normal</u> / High / Flooded				Meter ID# <u>153464</u>				Matrix: <u>fresh</u> / Marine / DI				
Photos: Yes / <u>No</u>				Flow: No Flow / Intertidal / <u>Flowing</u> / NA				Tide: Rising/ Falling/ Slack/ <u>NA</u>				
Tide Times: High Low												
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)		
EST	1140	2.9	0.3	1.1	8.80	83.0	6.86	0.11	225	12.6		
CEN	1145		2.4		6.70	62.1	6.36	0.13	276	11.4		
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:												
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:												
Parameter Suite	Parameter Methods				Preservation	Lot#	Expiration	pH Check				
Preserved Nutrients	☒ W-NH3 / ☒ W-NO2NO3 / ☒ W-TKN / ☒ W-TOC / ☒ W-S-A-TP				☒ Ice ☒ H2SO4	<u>6274050</u>		☒ <2				
Metals	☐ W-ICPMS ☐ W-ICP				☐ Ice ☐ HNO3			☐ <2				
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter				☒ Ice							
Chlorophyll	☒ CHLSUITE-W				☒ Ice							
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CL-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice							
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY				☐ Ice							
Macroinvertebrates	☐ MI-FW-QLDC				☐ Formalin							
Periphyton	☐ ALGAE-ID				☐ Ice							
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci				☒ Ice							
Molecular	☐ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD				☐ Ice							
Tracers	☐ W-SUC-AA-R ☐ W-UHERB-AA ☐ W-AHERB-AA				☐ Ice							
Pesticides	☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-CARB-AA				☐ Ice							
	☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R				☐ Other							

Notes:

A-b

Place Label Here

Signature:

Date: 1/29/16

Enter Field Parameters, Verify Station ID In LIMS DMT:

QC Station ID, Field Parameters In LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data In WIN:

Florida Department of Environmental Protection Central Laboratory Submittal Form

Customer: TLH-ROC

Requester: Michael Eckles

Field Report Prepared By: Woods

Project ID: SMP

Collected By: Woods/HardenSampling Agency: 8034

Location	Brothers River										Bottle Group(s)
Field ID											(See pg 2)
WIN/STORET #	G2TLH00013										A
Latitude	dd dms		Longitude		dd dms		Salinity (ppt)		pH		
Location	Sav1 Creek										
Field ID											
WIN/STORET #	G2TLH00014										A
Latitude	dd dms		Longitude		dd dms		Salinity (ppt)		pH		
Location	Double Bayou 800m Lake W.M.C.B										
Field ID											
WIN/STORET #	G2TLH0003										A
Latitude	dd dms		Longitude		dd dms		Salinity (ppt)		pH		
Location	100m upstream. Lake Horseshoe, Creek from intercoastal										
Field ID											
WIN/STORET #	G2TLH00033										A
Latitude	dd dms		Longitude		dd dms		Salinity (ppt)		pH		

Relinquished By: <u>JW</u>	Date/Time: <u>11/29/18 1451</u>	Shipping Method: <u>HD</u>	Number of Coolers Shipped: <u>1</u>	Received By: <u>[Signature]</u>	Date/Time: <u>11/29/18 1452</u>
----------------------------	---------------------------------	----------------------------	-------------------------------------	---------------------------------	---------------------------------

Group: A	Bottle Type:	P-1L	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WO-THI	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	ECOLI-18QT						
Group: A	Bottle Type:	P-500ML	# of Bottles: 4	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	4
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 4	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	4
	W-PO4-F						

6274050

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID:

1534184

RQ:

Project:

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	Ben Harden	11/29/18	0750	19.8	7621	9.15	9.15	100.3		1.034	P/F	D/F
ICV	"	"	0804	20.2	7621	8.98	9.06	99.1			P/F	D/F
CCV	"	"	1451	16.3	7602	9.81	9.89	100.9			P/F	D/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	Ben Harden	11/29/18	0752	180621C	07/19	1000	1000	P/F	D/F
ICV	"	"	0754	1805F81	05/20	100	101.1	P/F	D/F
CCV	"	"	1453	180621C	07/19	1000	996	P/F	D/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	Ben Harden	11/29/18	0756	180621E	01/20	7.03	19.0	17.8	-18.7	P/F	D/F
Calibr.	"	"	0758	180621D	07/19	4.00	18.8	4.00	157.3	P/F	D/F
Calibr.	"	"	0800	180405D	10/19	10.00	18.6	10.08	-174.5	P/F	D/F
ICV	"	"	0802	"	"	10.00	18.4	10.06	-174.0	P/F	D/F
CCV	"	"	1455	180621D	07/19	4.00	18.7	4.14	157.6	P/F	D/F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU;

mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 156.1, slope from 4 to 7 176.0 (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	Ben Harden	11/29/18	0749	0.00	0.00	P/F	D/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: RQ- Project:

- Notes:** (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro- meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.											P / F	L / F
ICV											P / F	L / F
CCV											P / F	L / F
CCV											P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. **Steady-state & Galvanic Sensors:** DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.								P / F	L / F
ICV								P / F	L / F
CCV								P / F	L / F
CCV								P / F	L / F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.						7.				P / F	L / F
Calibr.						4.				P / F	L / F
Calibr.						10.				P / F	L / F
ICV										P / F	L / F
CCV										P / F	L / F
CCV										P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? YES / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification:

If NO, what will be used? (circle one) **Secchi Disk Line / Sonar** Unique ID: ; Date of last verification:

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: