



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

Data Qualified?

Yes / No

January 2018

WIN Station Name: Ellaville Spring

Date: 09/05/2018
(mm/dd/yyyy)

WIN/ Field ID: G1TLHR0027

WBID: 3422Q

Latitude: 30.38437

(Decimal Degrees)

County: Suwannee

ORG ID: 21FLTLHR

Longitude: 83.17249

(Decimal Degrees)

Receiving Body of Water: Suwannee River

RQ-2018-06-1156

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
J. Bartlett M. Eckles					☒	☒	☒	☒	☒

Sampling Equipment		
☒ Sample Container	☐ Van Dorn	☐ Pole
☐ Carboy	☒ Syringe	
☐ Dipper	☐ Other	
Depth Measured with Equipment ID		
Collection Method		
☐ Wading	☐ Canoe/Kayak	
☒ From Shore	☐ Electric Motor	
☐ Other	☐ Gasoline Motor	

Water Level: Dry/ Pooled/ Low / Normal ☒ High / Flooded

Meter ID# 13477

Matrix: ☒ Fresh / ☐ Marine / ☐ DIPhotos: Yes ☐ No ☒ Flow: No Flow / Intestinal ☒ Flowing / NA

Tide: Rising/ Falling/ Slack/ NA

Tide Times: High ☒ Low ☐

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPTH)	Sp COND (µmhos/cm)	Temp. (C°)
EST	12:15	2.7	0.3	1.0	0.12	1.4	7.01	0.20	409	21.9
CEN			2.2		0.14	1.7	7.02	0.26	409	21.9

Biological Samples Collected: ☒ None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☐ W-NH3 / ☐ W-NO2NO3 / ☐ W-TKN / ☐ W-TOC / ☐ W-S-A-TP	☒ Ice ☒ H2SO4	SAV101020	4/1/19	☐ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☒ W-PO4-F ☐ Field Filtered 0.45 µm Filter	☐ Ice			
Chlorophyll	☐ CHL-SUITE-W	☐ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / ☐ W-F / ☐ W-TSS / ☐ TURBIDITY / ☐ W-CL-IC / ☐ W-COLOR / ☐ W-SO4-IC / ☐ W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / ☐ W-F / ☐ W-TSS / ☐ TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QLDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☐ E Coli ☐ F Coli ☐ Enterococci	☐ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-SUC-AA-R ☐ W-UHERB-AA ☐ W-AHERB-AA	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-CARB-AA ☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R	☐ Ice ☐ Other			

Notes:

Kit A only no Bacteria

Field/WIN ID →

Location ↓



G1TLHR0027



ELLAVILLE SPRING

Signature:

Gade R

Date: 09/05/18

Enter Field Parameters, Verify Station ID in LIMS DMT:

JB 9/7/18
(Initials/Date)

QC Station ID, Field Parameters in LIMS DMT:

(Initials/Date)

Scan/Save Field Sheets

JB 9/7/18
(Initials/Date)

Migrate Data to WIN:

(Initials/Date)

Verify Data in WIN:

(Initials/Date)

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there are qualified data on this page.

Meter ID: 154177

RQ-

Project: 225T1809

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp °C	Baro-meter mmHg	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Burchett	9/5/18	730	21.1	759.9	8.90	8.90	100	-	1.00	P/F	S/F
ICV	↓	↓	740	20.3	760.0	9.04	9.04	100	-	-	P/F	F
CCV				25.5	74.9	8.19	8.22	101	-	-	P/F	L/F
CCV											P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15 (Pro DSS 0.75 to 1.50); DO charge N/A. Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J. Burchett	9/5/18	732	2802533	09/19	1000	1000	P/F	L/F
ICV	↓	↓	742	180403A	01/20	100	101	P/F	F
CCV				180326A	04/19	100	102	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Temp °C	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Burchett	9/5/18	734	180221K	01/20	7.0	26.6	7.0	21.9	P/F	L/F
Calibr.	↓	↓	736	1405103	05/19	4.0	26.5	4.0	154.4	P/F	F
Calibr.	↓	↓	738	180221K	09/19	10.0	26.3	10.0	158.4	P/F	F
ICV			744	180221K	09/19	10.0	26.4	10.0	159.6	P/F	F
CCV				180326C	04/19	4.0	28.6	4.0	155.0	P/F	F
CCV										P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 _____, slope from 4 to 7 _____ (both must be between 165 and 180 mV)

Does meter have a depth sensor that will be used to measure total depth & sample depth? (YES) / NO / NA (not surf. water project)

If YES, complete daily Calibr. & ICV below and list date of last quarterly depth verification: 7/1/18

If NO, what will be used? (circle one) Secchi Disk Line / Sonar Unique ID: _____; Date of last verification: _____

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value, meters	Pass / Fail	Lab / Field
Pressure mode in air	J. Burchett	9/5/18	745	0.00	0.0	P/F	F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

Group: A	Bottle Type:	P-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	1
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WO-THI	# of Bottles: 5	Preservative:	ICE	Enter Number of Bottles Sent to Lab	0
	ECOLI-18QT						
Group: A	Bottle Type:	P-500ML	# of Bottles: 5	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	1
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 5	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	1
	W-PO4-F						
Group: B	Bottle Type:	QPCR-P-500ML	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	
	PCR-FILTER						
Group: B	Bottle Type:	BG-500ML	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	
	W-E8321-DI						
	W-E8321-MS						
Group: C	Bottle Type:	P-1L	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	
	ALKALINITY						
	TURBIDITY						
	W-F						
	W-TSS						
Group: C	Bottle Type:	BP-1L	# of Bottles: 4	Preservative:	ICE	Enter Number of Bottles Sent to Lab	
	CHLSUITE-W						