



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

Data Qualified?

Yes / No

January 2018

WIN Station Name: Chipola River @ Magnolia Rd

Date: 01/29/2018
(mm/dd/yyyy)

WIN/ Field ID: G2TLHR0016

WBID: 51E

Latitude: 30.7177819

(Decimal Degrees)

County: Jackson

ORG ID: 21FLTLHR

Longitude: 85.1999727

(Decimal Degrees)

Receiving Body of Water: Dead Lakes

RQ-2018-01-29-48

Sampling Team Members				WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment			
J. Burtchett J. Woods Tolbert				☒	☒	☒	☒	☒	☒ Sample Container	☐ Van Dorn	☐ Pole	
									☐ Carboy	☒ Syringe		
									☐ Dipper	☐ Other		
									Depth Measured with			
									Equipment ID			
									Collection Method			
									☐ Wading	☐ Canoe/Kayak		
									☒ From Shore	☐ Electric Motor		
									☐ Other	☐ Gasoline Motor		
Water Level: Dry/ Pooled/ Low / <u>Normal</u> / High / Flooded				Meter ID# <u>Exo 3</u>				Matrix: <u>Fresh</u> / Marine / DI				
Photos: Yes <u>No</u>				Flow: No Flow / Intestinal / <u>Flowing</u> / NA				Tide: Rising/ Falling/ Slack/ <u>NA</u>				
Tide Times: High Low												
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)		
<u>EST</u>	<u>1145</u>	<u>0.15</u>	<u>0.3</u>	<u>0.3</u>	<u>8.91</u>	<u>91</u>	<u>7.72</u>	<u>0.10</u>	<u>205</u>	<u>16.30</u>		
CEN												
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:												
SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:												
Parameter Suite	Parameter Methods				Preservation	Lot#	Expiration	pH Check				
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP				☒ Ice ☒ H2SO4	<u>5A7275010</u>	<u>6/2/18</u>	☒ <2				
Metals	☐ W-ICPMS ☐ W-ICP				☐ Ice ☐ HNO3			☐ <2				
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter				☒ Ice							
Chlorophyll	☒ CHLSUITE-W				☒ Ice							
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS				☒ Ice							
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY				☐ Ice							
Macroinvertebrates	☐ MI-FW-QLDC				☐ Formalin							
Periphyton	☐ ALGAE-ID				☐ Ice							
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci				☒ Ice							
Molecular	☐ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD				☐ Ice							
Tracers	☐ W-SUC-AA-R ☐ W-UHERB-AA ☐ W-AHERB-AA				☐ Ice							
Pesticides	☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-CARB-AA				☐ Ice							
	☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R				☐ Other							

Notes:

Place Label Here

Signature: Jake BurtchettDate: 01/29/2018

Enter Field Parameters, Verify Station ID In LIMS DMT:

QC Station ID, Field Parameters In LIMS DMT:

(Initials/Date)

(Initials/Date)

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data In WIN:

(Initials/Date)

(Initials/Date)

(Initials/Date)

**Florida Department of Environmental Protection
Central Laboratory Submittal Form**

Customer: TLH-ROC
Project ID: SMP

Requester: Chris Green
Collected By: J. Tolbert, S. Boyett

Field Report Prepared By: J. Burchett
Sampling Agency: SC24

Sampling Agency:

Relinquished By: JBT	Date/Time 1/29/18:1546	Shipping Method FD	Number of Coolers Shipped: 1	Received By: JBT	Date/Time 1/29/18 3:48
-------------------------	---------------------------	-----------------------	---------------------------------	---------------------	---------------------------

Group: A	Bottle Type:	P-1L	# of Bottles: 25	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 25	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WO-THI	# of Bottles: 25	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	ECOLI-18QT						
Group: A	Bottle Type:	P-500ML	# of Bottles: 25	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	4
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 25	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	4
	W-PO4-F						

Page 1 of 2
last revised Jan 3, 2018

J. Butcher
8034

[illegible]

Group: A	Bottle Type:	P-1L	# of Bottles: 25	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 25	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WO-THI	# of Bottles: 25	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	ECOLI-18QT						
Group: A	Bottle Type:	P-500ML	# of Bottles: 25	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	2
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 25	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	2
	W-PO4-F						

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this naae.

Meter ID: Ex03 RQ: 2018-01-29-48 Project: SMP 2018

- Notes: (1) Always wait for meter to stabilize before recording any readings.
 (2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).
 (3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill in: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

12/17

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Burtchett	1/29/18	806	20.30	9.04	9.04	100	-	1.06	P/F	L/F
ICV	✓	✓	816	20.95	8.92	8.94	100	-	-	P/F	L/F
CCV	✓	✓		13.7	10.40	10.39	100	-	-	P/F	L/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J. Burtchett	1/29/18	808	171121D	12/19	1000	1000	P/F	L/F
ICV	✓	✓	815	2710761	9/19	1000	1000	P/F	L/F
CCV	✓	✓		180116A	1/19	1000	1001	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Burtchett	1/29/18	810	171121A	06/19	7.0	7.0	-23.1	P/F	L/F
Calibr.	✓	✓	812	171121B	01/19	4.0	4.0	155.3	P/F	L/F
Calibr.	✓	✓	814	171121B	04/19	10.0	10.0	-193.2	P/F	L/F
ICV	✓	✓	820	171121B	06/19	10.0	10.0	-194.4	P/F	L/F
CCV	✓	✓		180116C	1/19	4.0	4.1	148.7	P/F	L/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 170.1, slope from 4 to 7 178.4 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification 12/17

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P/F	L/F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

