



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

Data Qualified?

Yes / No

January 2018

WIN Station Name: Lake Seminole 200m from boat ramp @ River Rd

Date: 4/18/18
(mm/dd/yyyy)

WIN/ Field ID: G2TLHR0002

WBID: 60

Latitude: 30.743518

(Decimal Degrees)

County: Jackson

ORG ID: 21FLTLHR

Longitude: 84.9204569

(Decimal Degrees)

Receiving Body of Water: Apalachicola River

RQ-2018- 04-09-52

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check
Don Woods					X	X			
Dade Bustchett						X	X	X	

Sampling Equipment		
☒ Sample Container	☐ Van Dorn	☐ Pole
☐ Carboy	☒ Syringe	
☐ Dipper	☐ Other	
Depth Measured with		
Equipment ID		
Collection Method		
☐ Wading	☐ Canoe/Kayak	
☐ From Shore	☐ Electric Motor	
☒ Other Dack	☐ Gasoline Motor	

Water Level: Dry/ Pooled/ Low / Normal / High / Flooded

Meter ID# 603

Matrix: Fresh / Marine / DIPhotos: Yes NoFlow: No Flow / Intestinal / Flowing / NATide: Rising/ Falling/ Slack/ NA

Tide Times: High Low

Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPTH)	Sp COND (umhos/cm)	Temp. (C°)
EST	1040	0.8	0.3	0.6	6.43	97.3	7.91	0.06	163	20.08
CEN				S						

Biological Samples Collected: None HA SCI RPS Algae ID LVS LVI Other:

SBIO ID Numbers: SBIO-ID: HA-ID: RPS-ID: Macrophyte-ID: LIMS#:

Parameter Suite	Parameter Methods	Preservation	Lot#	Expiration	pH Check
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP	☒ Ice ☒ H2SO4	7275010	10/2/16	☒ <2
Metals	☐ W-ICPMS ☐ W-ICP	☐ Ice ☐ HNO3			☐ <2
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 um Filter	☒ Ice			
Chlorophyll	☒ CHLSUITE-W	☒ Ice			
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-CH-IC / W-COLOR / W-SO4-IC / W-TDS	☒ Ice			
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY	☐ Ice			
Macroinvertebrates	☐ MI-FW-QIDC	☐ Formalin			
Periphyton	☐ ALGAE-ID	☐ Ice			
Microbiology	☐ E Coli ☐ F Coli ☐ Enterococci	☐ Ice			
Molecular	☐ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD	☐ Ice			
Tracers	☐ W-SUC-AA-R ☐ W-UHERB-AA ☐ W-AHERB-AA	☐ Ice			
Pesticides	☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-CARB-AA ☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R	☐ Ice ☐ Other			

Notes:
kit:A-b

Field/WIN ID →



G2TLHR0002

Location ↓

Lake Seminole 200m from boat ramp @ River Rd

Signature:

Date:

Enter Field Parameters, Verify Station ID in LIMS DMT: 70 4/20/18

QC Station ID, Field Parameters in LIMS DMT:

Scan/Save Field Sheets

Migrate Data to WIN:

Verify Data in WIN:

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this anaer.

Meter ID: Exc 3

RQ:

Project: SMP 2018/

2151804

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Burkett	4/18/18	7:54	20.5	9.00	9.01	100	-	1.06	☒ F	☒ F
ICV	↓	↓	8:04	19.4	9.20	9.19	100	-	-	☒ F	☒ F
CCV	↓	↓	16:02	26.4	8.06	8.04	100	-	-	☒ F	☒ F
CCV										P / F	L / F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ mhos/cm	Meter Reading μ mhos/cm	Pass / Fail	Lab / Field
Calibr.	J. Burkett	4/18/18	7:56	18011613	01/19	1000	1000	☒ F	☒ F
ICV	↓	↓	8:06	2710781	09/19	100	95	☒ F	☒ F
CCV	↓	↓	16:04	18011613	01/19	1000	995	☒ F	☒ F
CCV			16:08	2710781	09/19	100	96	☒ F	☒ F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Burkett	4/18/18	7:58	17121A	06/19	7.0	7.0	-37.7	☒ F	☒ F
Calibr.	↓	↓	8:00	18011610	01/19	4.0	4.0	140.2	☒ F	☒ F
Calibr.	↓	↓	8:02	180221E	09/19	10.0	10.0	-20.4	☒ F	☒ F
ICV	↓	↓	8:08	180221E	09/19	10.0	10.0	-20.30	☒ F	☒ F
CCV	↓	↓	16:10	18011610	01/19	4.0	4.1	135.4	☒ F	☒ F
CCV									P / F	L / F

pH Acceptance criteria ± 0.2 SU; mV pH 7 Range 0 ± 50 ; mV pH 4 Range $+180 \pm 50$; mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 164.2, slope from 4 to 7 177.9 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? ☒ Yes / No If yes, list date of last quarterly depth verification 12/17

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS:

kit A w/ bacteria

last revised Jan 25, 2018

Customer: TLH-ROC

Requester: Chris Green

Field Report Prepared By: J. Woods

Project ID: SMP

Collected By: JW/SB

Sampling Agency: 8034

Location	Field/WIN ID →		Collection (comp begin or grab)		Composite end		Bottle Group(s)
Field ID	↓		Matrix (type, e.g., Salt, Fresh, etc)	Diss. Oxygen	mg/L	Total Res. Chlorine	(See pg 2)
WIN/STORE	↓		Temp (C)	pH	ft	Sp. Conductance	
Latitude	Longitude		Salinity (ppt)	Comments			
Location	Field ID →		Collection (comp begin or grab)		Composite end		Bottle Group(s)
Field ID	↓		Matrix (type, e.g., Salt, Fresh, etc)	Diss. Oxygen	mg/L	Total Res. Chlorine	
WIN/STORE	↓		Temp (C)	pH	ft	Sp. Conductance	
Latitude	Longitude		Salinity (ppt)	Comments			
Location	Field ID →		Collection (comp begin or grab)		Composite end		Bottle Group(s)
Field ID	↓		Matrix (type, e.g., Salt, Fresh, etc)	Diss. Oxygen	mg/L	Total Res. Chlorine	
WIN/STORE	↓		Temp (C)	pH	ft	Sp. Conductance	
Latitude	Longitude		Salinity (ppt)	Comments			
Location	Field/WIN ID →		Collection (comp begin or grab)		Composite end		Bottle Group(s)
Field ID	↓		Matrix (type, e.g., Salt, Fresh, etc)	Diss. Oxygen	mg/L	Total Res. Chlorine	
WIN/STORE	↓		Temp (C)	pH	ft	Sp. Conductance	
Latitude	Longitude		Salinity (ppt)	Comments			

Relinquished By: JW	Date/Time: 4/18/18 1453	Shipping Method: HD	Number of Coolers Shipped: 1	Received By: J.D.	Date/Time: 4/18/18 2:55
---------------------	-------------------------	---------------------	------------------------------	-------------------	-------------------------

Group: A	Bottle Type:	P-1L	# of Bottles: 25	Preservative:	ICE	Enter Number of Bottles Sent to Lab	4
	ALKALINITY						
	TURBIDITY						
	W-CL-IC						
	W-COLOR						
	W-F						
	W-SO4-IC						
	W-TDS						
	W-TSS						
Group: A	Bottle Type:	BP-1L	# of Bottles: 25	Preservative:	ICE	Enter Number of Bottles Sent to Lab	3
	CHLSUITE-W						
Group: A	Bottle Type:	P-250ML-WO-THI	# of Bottles: 25	Preservative:	ICE	Enter Number of Bottles Sent to Lab	2
	ECOLI-18QT						
Group: A	Bottle Type:	P-500ML	# of Bottles: 25	Preservative:	ICE And H2SO4	Enter Number of Bottles Sent to Lab	4
	W-NH3						
	W-NO2NO3						
	W-S-A-TP						
	W-TKN						
	W-TOC						
Group: A	Bottle Type:	P-125ML	# of Bottles: 25	Preservative:	FILTER-ICE	Enter Number of Bottles Sent to Lab	4
	W-PO4-F						