



FLORIDA DEPARTMENT OF ENVIRONMENTAL PROTECTION STRATEGIC MONITORING PROGRAM FIELD SHEET

Data Qualified?

Yes / No

January 2018

WIN Station Name: Ocklawaha Creek @ SR267

Date: 5-14-18

(mm/dd/yyyy)

WIN/ Field ID: G1TLHR0053

WBID: 811

Latitude: 30.4503

(Decimal Degrees)

County: Gadsden

ORG ID: 21FLTLHR

Longitude: -84.6435

(Decimal Degrees)

Receiving Body of Water: Lake Talquin

RQ-2018- 05-14-41

Sampling Team Members					WQ Measurements	Sample Collection	Documentation	Preservation	Preservation Check	Sampling Equipment				
Chris Green Jon Woods					☒	☒	☒	☒	☒	☐ Sample Container	☐ Van Dorn	☐ Pole		
										☐ Carboy	☒ Syringe			
										☐ Dipper	☐ Other			
Depth Measured with					Equipment ID					Serahi				
										Collection Method				
☒ Wading					☐ Canoe/Kayak									
☐ From Shore					☐ Electric Motor									
☐ Other					☐ Gasoline Motor									
Water Level: Dry/ Pooled/ <u>Low</u> / Normal / High / Flooded					Meter ID# Pro DSS					Matrix: <u>Fresh</u> Marine / DI				
Photos: Yes ☒ No ☐ Flow: No Flow / Intestinal / <u>Flowing</u> / NA					Tide: Rising/ Falling/ Slack/ NA					Tide Times: High Low				
Time Zone	Time	Total Depth (m)	Sample Depth (m)	Secchi Disk (m)	DO (mg/L)	DO (%SAT)	pH	Salinity (PPT)	Sp COND (µmhos/cm)	Temp. (C°)				
EST	14:15	0.2	0.1	0.25	8.24	93.4	5.73	0.01	14.6	21.6				
CEN				0.13										
Biological Samples Collected: <u>None</u> HA SCI RPS Algae ID LVS LVI Other:														
SBIO ID Numbers:		SBIO-ID:		HA-ID:		RPS-ID:		Macrophyte-ID:		LIMS#:				
Parameter Suite	Parameter Methods					Preservation		Lot#	Expiration	pH Check				
Preserved Nutrients	☒ W-NH3 / W-NO2NO3 / W-TKN / W-TOC / W-S-A-TP					☒ Ice ☒ H2SO4		SA 7163030	6/12/18	☒ <2				
Metals	☐ W-ICPMS ☐ W-ICP					☐ Ice ☐ HNO3				☐ <2				
Filtered Nutrients	☒ W-PO4-F ☒ Field Filtered 0.45 µm Filter					☒ Ice								
Chlorophyll	☒ CHLSUITE-W					☒ Ice								
Physical/Aggregate (Fresh)	☒ Alkalinity / W-F / W-TSS / TURBIDITY / W-Cl-IC / W-COLOR / W-SO4-IC / W-TDS					☒ Ice								
Physical/Aggregate (Marine)	☐ Alkalinity / W-F / W-TSS / TURBIDITY					☐ Ice								
Macroinvertebrates	☐ MI-FW-QLDC					☐ Formalin								
Periphyton	☐ ALGAE-ID					☐ Ice								
Microbiology	☒ E Coli ☐ F Coli ☐ Enterococci					☒ Ice								
Molecular	☒ PCR-FILTER ☐ PCR-HF183 ☐ PCR-BacR ☐ PCR-GULL2 ☐ PCR-GFD					☒ Ice								
Tracers	☒ W-SUC-AA-R ☒ W-UHERB-AA ☒ W-AHERB-AA					☒ Ice								
Pesticides	☐ W-PSNP-TQ ☐ W-UHERB-AA ☐ W-AHERB-AA ☐ W-CARB-AA					☐ Ice								
	☐ W-PCL-SQ-R ☐ W-PYR-AA-R ☐ W-GLYPH ☐ W-SUC-AA-R					☐ Other								

Notes:
A-b/t/m

Field/WIN ID

Location



G1TLHR0053



OCKLAWAHA CREEK @ SR 267

Date:

5-14-18

Signature:

Enter Field Parameters, Verify Station ID In LIMS DMT: 06 5-25-18

(Initials/Date)

QC Station ID, Field Parameters In LIMS DMT:

(Initials/Date)

Scan/Save Field Sheets 05-25-18

(Initials/Date)

Migrate Data to WIN:

(Initials/Date)

Verify Data In WIN:

(Initials/Date)

Florida Department of Environmental Protection
Central Laboratory Submittal Form

Customer: TLH-ROC

Requester: Chris Green

Field Report Prepared By: Chris Green

Project ID: SMP

Collected By: Chris Green

Sampling Agency: 8034

Locatio	Field/WIN ID →		Collection (comp begin or grab)		Composite end		Bottle Group(s)	
Field ID	Location ↓		Date 5-14-18 Time 13:40		Date		(See pg 2)	
WIN/STC			Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen 7.59		Total Res. Chlorine	
Latitude	Longitude		Temp (C) 21.9		pH 5.67		Sample Depth 0.1	
	dd dms		Salinity (ppt) 0.01		ft m		Sp. Conductance (umho/cm)	
Location:			Collection (comp begin or grab)		Composite end		Bottle Group(s)	
Field ID	Field/WIN ID →		Date 5-14-18 Time 14:15		Date		(See pg 2)	
WIN/STC	Location ↓		Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen 8.24		Total Res. Chlorine	
Latitude	Longitude		Temp (C) 21.6		pH 5.73		Sample Depth 0.1	
	dd dms		Salinity (ppt) 0.01		ft m		Sp. Conductance (umho/cm) 14.6	
Location			Collection (comp begin or grab)		Composite end		Bottle Group(s)	
Field ID			Date		Date		Time	
WIN/STORET #			Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen		Total Res. Chlorine	
Latitude	Longitude		Temp (C)		pH		Sample Depth	
	dd dms		Salinity (ppt)		ft m		Sp. Conductance (umho/cm)	
Location			Collection (comp begin or grab)		Composite end		Bottle Group(s)	
Field ID			Date		Date		Time	
WIN/STORET #			Matrix (type, e.g., Salt, Fresh, etc)		Diss. Oxygen		Total Res. Chlorine	
Latitude	Longitude		Temp (C)		pH		Sample Depth	
	dd dms		Salinity (ppt)		ft m		Sp. Conductance (umho/cm)	

Relinquished By: Chris Green	Date/Time 5-14-18 15:10	Shipping Method Hand	Number of Coolers Shipped:	Received By:	Date/Time
------------------------------	-------------------------	----------------------	----------------------------	--------------	-----------

CALIBRATION AND VERIFICATION LOG (FDEP SOP FT 1000-FT 1500, FD 1000-FD 4000)

Boldly "X" this box if there is qualified data on this name.

Meter ID: ProDSS

RQ-

Project: SMP2018

Notes: (1) Always wait for meter to stabilize before recording any readings.

(2) Report all digits displayed. Do not round before reporting measurements. (See special instructions for depth).

(3) For Calibrations, record calibrated meter reading. Do not record initial meter reading before calibration.

Circle/Fill In: Lab Calibration or Calibration/Verification on Site

Temperature (Quarterly) FT 1400

Date of Last Temperature Verification 3/18

DO DEP SOP FT 1500	Name	Date	Time CT-ET	Temp	D.O. Chart mg/L	Meter D.O. mg/L	% DO	Probe Charge	Probe Gain	Pass / Fail	Lab / Field
Calibr.	J. Burchett	5/14/18	755	22.7	8.63	6.63	100	-	1.09	P/F	L/F
ICV	↓	↓	805	22.8	8.61	8.62	100	-	-	P/F	L/F
CCV	Chris Green	8/11	15:30	23.2	8.54	8.57	100.3	-	-	P/F	L/F
CCV										P/F	L/F

DO Acceptance criteria from Table ± 0.3 mg/L.

Rapid-Pulse Sensors: DO Gain Range 0.7 to 1.4; DO Charge Range 25-75.

Optical: DO gain range 0.85 to 1.15; DO charge N/A.

Steady-state & Galvanic Sensors: DO Gain & Charge N/A.

Spec. Cond. FT 1200	Name	Date	Time CT-ET	Lot #	Expir. Date	Standard μ hos/cm	Meter Reading μ hos/cm	Pass / Fail	Lab / Field
Calibr.	J. Burchett	5/14/18	757	180403A	04/19	1000	1000	P/F	L/F
ICV	↓	↓	807	2802833	01/20	100	101	P/F	L/F
CCV	Chris Green	8/11	15:35	2802833	01/20	101	99	P/F	L/F
CCV								P/F	L/F

Conductivity Acceptance criteria $\pm 5\%$

pH DEP SOP FT 1100	Name	Date	Time CT-ET	Lot #	Expir. Date	pH Buffer SU	Meter reading SU	mV	Pass / Fail	Lab / Field
Calibr.	J. Burchett	5/14/18	759	171124	06/19	7.0	7.0	14.7	P/F	L/F
Calibr.	↓	↓	801	180116D	01/19	4.0	4.0	148.0	P/F	L/F
Calibr.	↓	↓	803	180221E	09/19	10.0	10.0	-167.2	P/F	L/F
ICV	↓	↓	809	180221E	09/19	10.0	10.0	-171.0	P/F	L/F
CCV	Chris Green	8/11	15:38	180116D	01/19	4.0	4.0	-	P/F	L/F
CCV									P/F	L/F

pH Acceptance criteria ± 0.2 SU;

mV pH 7 Range 0 ± 50 ;

mV pH 4 Range $+180 \pm 50$;

mV pH 10 Range -180 ± 50 ;

If mV are recorded: slope from 7 to 10 152.5, slope from 4 to 7 162.7 (both must be between 165 and 180 mV)

Is meter equipped with depth sensor? Yes / No If yes, list date of last quarterly depth verification 3/18

Depth Sensor (Daily Calibration & ICV)	Name	Date	Time CT-ET	Calibrated Value (0.00 or Offset), meters	ICV Value meters	Pass / Fail	Lab / Field
Pressure mode in air						P / F	L / F

Report two decimal places. Round numbers ≤ 4 down, ≥ 5 up. ICV acceptance criteria $\pm 5\%$ or ± 0.05 m, whichever is greater.

COMMENTS: