



a xylem brand

SOLUTION EXPIRATION DATES

CONDUCTIVITY

Unopened -- Glass Quarts _____ 1 Year
Unopened -- Plastic Pints _____ 18 Months
Opened _____ 1 Month

ZOBELL

Unmixed _____ 2 Years
Mixed _____ 6 Months

pH

Unopened _____ 2 Years
Opened _____ 6 Months

CONFIDENCE SOLUTION

Unopened _____ 1 Years
Opened _____ 3 Months

TURBIDITY

Unopened _____ 1 Year
Opened _____ 6 Months

AMMONIUM


Unopened _____ 18 Months
Opened _____ 6 Months

NITRATE

Unopened _____ 1 Year
Opened _____ 6 Months

O² PROBE ELECTROLYTE SOLUTION

Unmixed Crystals _____ No State Expiration
Mixed Solution _____ 1 Year



NOTE:

Expiration times for unopened / unmixed solutions are from date of manufacture, not date of purchase.

Expiration dates are clearly labeled on the bottle.

YSI Inc. | 1725 Brannum Ln | Yellow Springs, OH 45387
Tel +1.937.767.7241 | 800.897.4151 | environmental@ysi.com

© 2014 Xylem, Inc. All rights reserved.

xylem
Let's Solve Water