

Chemical Analysis Report

TLH-ROC-2018-04-11-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **kit A w/ bacteria**
Request ID: **RQ-2018-03-26-57**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 01-MAY-2018 10:20

A handwritten signature in black ink, appearing to read "Colin Wright", is written over a light gray rectangular background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: WOODS CREEK AT OSTEEN RD

Collection Date/Time: 04/11/2018 11:15

Field ID: G1TLHR0067

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1983677	EPA 180.1 Rev. 2.0	Turbidity	1.4		NTU	P343259	
	EPA 300.0 Rev. 2.1	Chloride	6.2		mg Cl/L	P343434	
		Sulfate	2.1		mg SO4/L	P343437	
	SM 2120 B	Color (true)	220		PCU	P343211	
	SM 2320 B-97	Total Alkalinity	134		mg CaCO3/L	P343950	
	SM 2540 C-97	TDS	180	A	mg/L	P343621	
	SM 2540 D-97	TSS	3	IA	mg/L	P343873	
	SM 4500 F-C-97	Fluoride	0.14		mg F/L	P343853	
1983678	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P343686	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.47		mg N/L	P343350	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.056		mg N/L	P343568	
	EPA 365.1 Rev. 2.0	Total-P	0.16		mg P/L	P343480	
	SM 5310 B-00	Organic Carbon	18		mg C/L	P343323	
1983679	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.14		mg P/L	P343224	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P343259

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P343434

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P343437

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P343686

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P343350

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P343568

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P343480

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P343224

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P343211

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P343950

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P343621

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P343873

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P343853

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P343323

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P343434

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	98.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P343437

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	98.0		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P343686

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	95.9		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P343350

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P343568

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P343480

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P343224

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	99.9		P	90 - 110

Reference Method: SM 2320 B-97
Batch ID: P343950

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	101		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P343853

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	95.1		P	91 - 105

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
Batch ID: P343323

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.2		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P343434

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1983244	Chloride	101		P	90 - 110
1983293	Chloride	96.3		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P343437

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1983244	Sulfate	98.3		P	90 - 110
1983293	Sulfate	93.7		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P343686

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1983414	Ammonia-N	98.5	95.1	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P343350

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1983903	Kjeldahl Nitrogen	97.4	97.8	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P343568

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1983678	NO2NO3-N	98.2	98.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P343480

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1983415	Total-P	101	103	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P343853

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1983677	Fluoride	96.7	97.4	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P343323

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1981878	Organic Carbon	98.2		P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P343259

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983557	Turbidity	1.73	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P343434

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983244	Chloride	0.573	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P343437

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983244	Sulfate	0.423	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P343686

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983414	Ammonia-N	2.03	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P343350

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983903	Kjeldahl Nitrogen	0.259	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P343568

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983678	NO2NO3-N	0.0	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P343480

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983415	Total-P	1.80	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P343224

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983679	O-Phosphate-P	0.248	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P343211

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983425	Color (true)	1.07	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P343950

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1984545	Total Alkalinity	0.310	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P343621

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983677	TDS	2.77	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P343853

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1983677	Fluoride	0.473	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P343323

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1981878	Organic Carbon	1.65	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A83166
Included Lab Sample IDs: 1983677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	102	103	P/P	90 - 115

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A83169
Included Lab Sample IDs: 1983679

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.6	99.2	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A83184
Included Lab Sample IDs: 1983677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.5	99.0	P/P	95 - 105

Reference Method: SM 5310 B-00
Run ID: A83208
Included Lab Sample IDs: 1983678

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.6	97.0	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A83234
Included Lab Sample IDs: 1983677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	100	99.1	P/P	90 - 110
Sulfate	102	99.3	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A83278
Included Lab Sample IDs: 1983678

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	100	105	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A83287
Included Lab Sample IDs: 1983678

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	101	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A83314
Included Lab Sample IDs: 1983678

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	102	104	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A83326

Included Lab Sample IDs: 1983678

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A83392

Included Lab Sample IDs: 1983677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	99.0	98.6	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A83420

Included Lab Sample IDs: 1983677

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	98.6	102	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				1.73	
EPA 300.0 Rev. 2.1	Chloride	98.9	101 96.3		0.573	
	Sulfate	98.0	98.3 93.7		0.423	
EPA 350.1 Rev. 2.0	Ammonia-N	95.9	98.5 95.1			2.03
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103	97.4 97.8			0.259
EPA 353.2 Rev. 2.0	NO2NO3-N	101	98.2 98.2			0.0
EPA 365.1 Rev. 2.0	Total-P	101	101 103			1.80
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	99.9				0.248
SM 2120 B	Color (true)				1.07	
SM 2320 B-97	Total Alkalinity	101			0.310	
SM 2540 C-97	TDS				2.77	
SM 4500 F-C-97	Fluoride	95.1	96.7 97.4			0.473
SM 5310 B-00	Organic Carbon	98.2	98.2			1.65

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1983677
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1983677
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1983677
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1983678

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1983678
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1983678
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1983678
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1983679
SM 2120 B / E31780	True Color measured at 450 nm	1983677
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1983677
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1983677
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices	1983677
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1983677
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1983678

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	04/11/2018			04/11/2018 17:57	William L. Brown IV	1983677
EPA 300.0 Rev. 2.1	04/11/2018			04/14/2018 02:16	Lipi Saha	1983677
EPA 350.1 Rev. 2.0	04/11/2018			04/17/2018 10:38	Ping Hua	1983678
EPA 351.2 Rev. 2.0	04/11/2018	04/13/2018 10:00	Adrian Shelby	04/17/2018 10:39	Joseph Suarez	1983678
EPA 353.2 Rev. 2.0	04/11/2018			04/17/2018 13:21	Lauren Bottorf	1983678
EPA 365.1 Rev. 2.0	04/11/2018	04/16/2018 13:16	Sima Feizi	04/18/2018 10:29	Beverly Y. Sanders	1983678
EPA 365.1 Rev. 2.0 dissolved	04/11/2018			04/11/2018 16:26	Megan Treadwell	1983679
SM 2120 B	04/11/2018			04/11/2018 16:08	Tanya Welborn	1983677
SM 2320 B-97	04/11/2018			04/24/2018 00:44	Dale L. Simmons	1983677
SM 2540 C-97	04/11/2018			04/13/2018 15:23	Yijie Li	1983677
SM 2540 D-97	04/11/2018			04/13/2018 15:23	Yijie Li	1983677
SM 4500 F-C-97	04/11/2018			04/20/2018 20:43	Lauren Bottorf	1983677
SM 5310 B-00	04/11/2018			04/13/2018 02:11	Megan Treadwell	1983678