

Chemical Analysis Report

TLH-ROC-2018-05-24-03

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **kit A w/ bacteria**
Request ID: **RQ-2018-05-14-31**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 12-JUN-2018 17:07



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: G1TLHR0067

Collection Date/Time: 05/24/2018 10:15

Field ID: G1TLHR0067

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
1995430	EPA 180.1 Rev. 2.0	Turbidity	2.0		NTU	P345908	
	EPA 300.0 Rev. 2.1	Chloride	6.2		mg Cl/L	P346221	
		Sulfate	2.1		mg SO4/L	P346225	
	SM 2120 B	Color (true)	250		PCU	P345907	
	SM 2320 B-97	Total Alkalinity	126		mg CaCO3/L	P346092	RPD
	SM 2540 C-97	TDS	180		mg/L	P346204	
	SM 2540 D-97	TSS	4	I	mg/L	P346206	
	SM 4500 F-C-97	Fluoride	0.14		mg F/L	P346097	
1995431	EPA 350.1 Rev. 2.0	Ammonia-N	0.008		mg N/L	P346341	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.76		mg N/L	P346275	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.076		mg N/L	P346060	
	EPA 365.1 Rev. 2.0	Total-P	0.19		mg P/L	P346101	
	SM 5310 B-00	Organic Carbon	21		mg C/L	P346490	
1995432	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.15		mg P/L	P345988	

Ref. Method and Comment:

EPA 365.1 Rev. 2.0 dissolved: O-Phosphate-P: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0

Batch ID: P345908

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P346221

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1

Batch ID: P346225

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0

Batch ID: P346341

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0

Batch ID: P346275

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P346060

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P346101

Component	Result	Code	Units
Total-P	0.005	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P345988

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P345907

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P346092

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P346204

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P346206

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P346097

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P346490

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P346221

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P346225

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P346341

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	103		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P346275

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	91.8		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P346060

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	102		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P346101

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P345988

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.6		P	90 - 110

Reference Method: SM 2320 B-97
 Batch ID: P346092

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	102		P	98 - 107

Reference Method: SM 4500 F-C-97
 Batch ID: P346097

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	95.8		P	91 - 105

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: SM 5310 B-00
 Batch ID: P346490

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.1		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P346221

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1994926	Chloride	97.9		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P346225

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1994926	Sulfate	99.2		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P346341

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1995453	Ammonia-N	96.3	97.6	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P346275

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1995453	Kjeldahl Nitrogen	104	108	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P346060

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1995431	NO2NO3-N	101	99.2	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P346101

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1995453	Total-P	106	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97
 Batch ID: P346097

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1995691	Fluoride	96.8	96.1	P/P	91 - 105

Reference Method: SM 5310 B-00
 Batch ID: P346490

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
---------------	-----------	---------	---------	-----------	----------------

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P346490

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
1995351	Organic Carbon	96.9		P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P345908

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995131	Turbidity	9.16	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P346221

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1994926	Chloride	1.51	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P346225

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1994926	Sulfate	2.69	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P346341

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995453	Ammonia-N	1.05	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P346275

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995453	Kjeldahl Nitrogen	2.91	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P346060

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995431	NO2NO3-N	1.09	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P346101

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995453	Total-P	3.92	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P345988

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995432	O-Phosphate-P	1.03	Spike	P	0 - 20

Reference Method: SM 2120 B
Batch ID: P345907

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995452	Color (true)	7.58	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P346092

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995691	Total Alkalinity	2.32	Sample	F	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P346204

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995450	TDS	2.01	Sample	P	0 - 10

Reference Method: SM 2540 D-97
Batch ID: P346206

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995450	TSS	2.53	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P346097

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995691	Fluoride	0.470	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P346490

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
1995351	Organic Carbon	0.0	Spike	P	0 - 20

Quality Assurance Report Precision

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A84144
Included Lab Sample IDs: 1995430

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	106	101	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A84145
Included Lab Sample IDs: 1995430

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	99.0	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A84169
Included Lab Sample IDs: 1995432

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	98.8	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Run ID: A84207
Included Lab Sample IDs: 1995431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	102	102	P/P	90 - 110

Reference Method: SM 2320 B-97
Run ID: A84220
Included Lab Sample IDs: 1995430

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	104	101	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Run ID: A84220
Included Lab Sample IDs: 1995430

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	98.1	98.6	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A84230
Included Lab Sample IDs: 1995430

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	103	P/P	90 - 110
Sulfate	99.9	101	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A84336

Included Lab Sample IDs: 1995431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	103	101	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A84352

Included Lab Sample IDs: 1995431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	103	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A84367

Included Lab Sample IDs: 1995431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	107	106	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A84394

Included Lab Sample IDs: 1995431

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	99.5	100	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery	LCS	Precision SMP	MS
EPA 180.1 Rev. 2.0	Turbidity				9.16	
EPA 300.0 Rev. 2.1	Chloride	102	97.9		1.51	
	Sulfate	100	99.2		2.69	
EPA 350.1 Rev. 2.0	Ammonia-N	103	96.3	97.6		1.05
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	91.8	104	108		2.91
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	102	101	99.2		1.09
EPA 365.1 Rev. 2.0	Total-P	101	106	101		3.92
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.6				1.03
SM 2120 B	Color (true)				7.58	
SM 2320 B-97	Total Alkalinity	102			2.32	
SM 2540 C-97	TDS				2.01	
SM 2540 D-97	TSS				2.53	
SM 4500 F-C-97	Fluoride	95.8	96.8	96.1		0.470
SM 5310 B-00	Organic Carbon	98.1	96.9			0.0

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	1995430
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	1995430
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	1995430
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	1995431
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	1995431
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	1995431
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	1995431
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	1995432
SM 2120 B / E31780	True Color measured at 450 nm	1995430
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	1995430
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	1995430
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	1995430
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	1995430
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	1995431

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	05/24/2018			05/24/2018 16:55	William L. Brown IV	1995430
EPA 300.0 Rev. 2.1	05/24/2018			06/01/2018 04:38	Lipi Saha	1995430
EPA 350.1 Rev. 2.0	05/24/2018			06/04/2018 15:06	Julia Storbeck	1995431
EPA 351.2 Rev. 2.0	05/24/2018	06/01/2018 14:20	Adrian Shelby	06/05/2018 14:35	Joseph Suarez	1995431
EPA 353.2 Rev. 2.0	05/24/2018			05/29/2018 13:17	Lauren Bottorf	1995431
EPA 365.1 Rev. 2.0	05/24/2018	05/30/2018 13:15	Sima Feizi	06/05/2018 09:18	Beverly Y. Sanders	1995431
EPA 365.1 Rev. 2.0 dissolved	05/24/2018			05/25/2018 14:09	Megan Treadwell	1995432
SM 2120 B	05/24/2018			05/24/2018 16:09	Tanya Welborn	1995430
SM 2320 B-97	05/24/2018			05/29/2018 22:11	Dale L. Simmons	1995430
SM 2540 C-97	05/24/2018			05/25/2018 13:00	William L. Brown IV	1995430
SM 2540 D-97	05/24/2018			05/25/2018 13:00	William L. Brown IV	1995430
SM 4500 F-C-97	05/24/2018			05/29/2018 22:11	Dale L. Simmons	1995430
SM 5310 B-00	05/24/2018			06/05/2018 03:50	Megan Treadwell	1995431