

Chemical Analysis Report

TLH-ROC-2018-08-16-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **A-b/t/m**
Request ID: **RQ-2018-05-14-41**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Liang Lin, Environmental Administrator

Date Certified: 06-SEP-2018 11:15

A handwritten signature in black ink, appearing to read "Liang Lin", on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical groups are included in this report: Nutrients and Pesticides.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Moore Branch @ Cody Church Rd.

Collection Date/Time: 08/16/2018 09:50

Field ID: G1TLHR0041

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2017775	EPA 180.1 Rev. 2.0	Turbidity	2.9	A	NTU	P350291	
	EPA 300.0 Rev. 2.1	Chloride	4.5		mg Cl/L	P350643	
		Sulfate	0.72		mg SO4/L	P350645	
	SM 2120 B	Color (true)	150		PCU	P350295	
	SM 2320 B-97	Total Alkalinity	0.65	U	mg CaCO3/L	P350424	
	SM 2540 C-97	TDS	21	I	mg/L	P350588	
	SM 2540 D-97	TSS	4	I	mg/L	P350578	
	SM 4500 F-C-97	Fluoride	0.035	U	mg F/L	P350427	
	EPA 350.1 Rev. 2.0	Ammonia-N	0.013		mg N/L	P350629	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.64		mg N/L	P350458	
2017776	EPA 353.2 Rev. 2.0	NO2NO3-N	0.009	I	mg N/L	P350527	
	EPA 365.1 Rev. 2.0	Total-P	0.074		mg P/L	P350510	
	SM 5310 B-00	Organic Carbon	13		mg C/L	P350548	
	EPA 365.1 Rev. 2.0	O-Phosphate-P	0.032		mg P/L	P350265	
	dissolved						
2017805	EPA 8321B	2,4-D	0.0020	U	ug/L	P350331	
		Bentazon	8.0E-04	U	ug/L	P350331	
		MCP	0.0020	UJ	ug/L	P350331	RPD
		Triclopyr**	0.0040	UJ	ug/L	P350331	RPD
		Imidacloprid	0.0020	U	ug/L	P350331	
		Diuron	0.23		ug/L	P350331	
		Pyraclostrobin**	0.0020	U	ug/L	P350331	
		Primidone**	0.0040	U	ug/L	P350331	
		Linuron	0.0040	U	ug/L	P350331	
		Acetaminophen**	0.0080	U	ug/L	P350331	
		Fenuron	0.016	U	ug/L	P350331	
		Imazapyr**	0.0040	U	ug/L	P350331	
		Carbamazepine**	8.0E-04	U	ug/L	P350331	
		Fluridone**	4.0E-04	U	ug/L	P350331	
		Hydrocodone**	8.0E-04	U	ug/L	P350331	
		Sucralose**	0.010	U	ug/L	P350234	

Ref. Method and Comment:

EPA 8321B: Insufficient sample to perform matrix spikes. Refer to the narrative for an explanation of QC Codes.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P350291

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350643

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350645

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350629

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350458

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350527

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350510

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P350265

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: EPA 8321B
Batch ID: P350234

Component	Result	Code	Units
Sucralose	0.010	U	ug/L

Quality Assurance Report Method Blank Results

Reference Method: EPA 8321B
Batch ID: P350331

Component	Result	Code	Units
2,4-D	0.0020	U	ug/L
Acetaminophen	0.0080	U	ug/L
Bentazon	8.0E-04	U	ug/L
Carbamazepine	8.0E-04	U	ug/L
Diuron	0.0020	U	ug/L
Fenuron	0.016	U	ug/L
Fluridone	4.0E-04	U	ug/L
Hydrocodone	8.0E-04	U	ug/L
Imazapyr	0.0040	U	ug/L
Imidacloprid	0.0020	U	ug/L
Linuron	0.0040	U	ug/L
MCPP	0.0020	U	ug/L
Primidone	0.0040	U	ug/L
Pyraclostrobin	0.0020	U	ug/L
Triclopyr	0.0040	U	ug/L

Reference Method: SM 2120 B
Batch ID: P350295

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Reference Method: SM 2320 B-97
Batch ID: P350424

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO3/L

Reference Method: SM 2540 C-97
Batch ID: P350588

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P350578

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P350427

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P350548

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350643

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	102		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350645

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	101		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350629

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	99.1		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350458

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	103		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350527

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350510

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	97.9		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P350265

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	100		P	90 - 110

Reference Method: EPA 8321B
Batch ID: P350234

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sucralose	107		P	40 - 150

Reference Method: EPA 8321B
Batch ID: P350331

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2,4-D	101	75.1	P/P	40 - 150
Acetaminophen	131	113	P/P	30 - 150
Bentazon	110	99.5	P/P	30 - 150
Carbamazepine	92.3	90.0	P/P	40 - 150

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 8321B
Batch ID: P350331

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Diuron	91.6	95.4	P/P	40 - 150
Fenuron	117	107	P/P	40 - 150
Fluridone	79.5	76.8	P/P	40 - 150
Hydrocodone	105	101	P/P	40 - 150
Imazapyr	102	98.2	P/P	40 - 150
Imidacloprid	113	106	P/P	40 - 160
Linuron	76.1	74.7	P/P	40 - 150
MCPP	109	72.1	P/P	40 - 150
Primidone	109	108	P/P	40 - 150
Pyraclostrobin	78.9	74.0	P/P	40 - 150
Triclopyr	100	64.7	P/P	40 - 150

Reference Method: SM 2120 B
Batch ID: P350295

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	96.1		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P350424

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	103		P	98 - 107

Reference Method: SM 4500 F-C-97
Batch ID: P350427

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.7		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P350548

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	98.1		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350643

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017552	Chloride	93.0		P	90 - 110
2017638	Chloride	94.4		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P350645

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017552	Sulfate	95.7		P	90 - 110
2017638	Sulfate	97.1		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P350629

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017776	Ammonia-N	98.5	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P350458

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017697	Kjeldahl Nitrogen	104	105	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P350527

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017776	NO2NO3-N	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P350510

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017697	Total-P	96.8	94.8	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P350265

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017702	O-Phosphate-P	99.0	99.2	P/P	90 - 110

Reference Method: EPA 8321B
Batch ID: P350234

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017129	Sucralose	109	105	P/P	40 - 150

Reference Method: SM 4500 F-C-97
Batch ID: P350427

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017691	Fluoride	99.2	97.9	P/P	91 - 105

Quality Assurance Report Matrix Spike Accuracy

Reference Method: SM 5310 B-00
Batch ID: P350548

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2017596	Organic Carbon	108		P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
 Batch ID: P350291

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017775	Turbidity	3.47	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P350643

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017552	Chloride	1.91	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P350645

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017552	Sulfate	1.47	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P350629

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017776	Ammonia-N	2.14	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P350458

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017697	Kjeldahl Nitrogen	0.780	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P350527

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017776	NO2NO3-N	0.0	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P350510

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017697	Total-P	1.91	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P350265

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017702	O-Phosphate-P	0.202	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: EPA 8321B
Batch ID: P350234

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017129	Sucralose	3.80	Spike	P	0 - 30

Reference Method: EPA 8321B
Batch ID: P350331

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
LFB	2,4-D	29.6	LCS	P	0 - 30
LFB	Acetaminophen	15.4	LCS	P	0 - 30
LFB	Bentazon	10.3	LCS	P	0 - 30
LFB	Carbamazepine	2.51	LCS	P	0 - 30
LFB	Diuron	4.14	LCS	P	0 - 30
LFB	Fenuron	8.57	LCS	P	0 - 30
LFB	Fluridone	3.51	LCS	P	0 - 30
LFB	Hydrocodone	3.72	LCS	P	0 - 30
LFB	Imazapyr	4.06	LCS	P	0 - 30
LFB	Imidacloprid	6.58	LCS	P	0 - 30
LFB	Linuron	1.81	LCS	P	0 - 30
LFB	MCP	40.5	LCS	F	0 - 30
LFB	Primidone	1.36	LCS	P	0 - 30
LFB	Pyraclostrobin	6.43	LCS	P	0 - 30
LFB	Triclopyr	42.9	LCS	F	0 - 30

Reference Method: SM 2120 B
Batch ID: P350295

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017515	Color (true)	3.41	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P350424

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017691	Total Alkalinity	0.0281	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P350588

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017808	TDS	1.44	Sample	P	0 - 10

Reference Method: SM 2540 D-97
Batch ID: P350578

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017785	TSS	2.15	Sample	P	0 - 10

Quality Assurance Report Precision

Reference Method: SM 4500 F-C-97
Batch ID: P350427

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017691	Fluoride	0.986	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P350548

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2017596	Organic Carbon	0.448	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Surrogates

Lab Sample ID: 2017805
Field Sample ID: G1TLHR0041

Reference Method	Surrogate	% Rec.	Pass/Fail	Control Limits
EPA 8321B	2,4-D-d3	128	P	30 - 160
EPA 8321B	Acetaminophen-d4	143	P	30 - 160
EPA 8321B	Carbamazepine-d10	66.5	P	30 - 160
EPA 8321B	Diuron-d6	66.5	P	30 - 160
EPA 8321B	Primidone-d5	103	P	30 - 160
EPA 8321B	Sucralose-d6	98.6	P	40 - 160

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A85936

Included Lab Sample IDs: 2017777

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.2	101	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A85945

Included Lab Sample IDs: 2017775

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	102	100	P/P	95 - 105

Reference Method: SM 2120 B

Run ID: A85946

Included Lab Sample IDs: 2017775

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	96.9	98.5	P/P	90 - 115

Reference Method: EPA 8321B

Run ID: A85983

Included Lab Sample IDs: 2017805

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Sucralose	92.5	96.8	P/P	60 - 160

Reference Method: SM 2320 B-97

Run ID: A85999

Included Lab Sample IDs: 2017775

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	104	102	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A85999

Included Lab Sample IDs: 2017775

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	100	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A86029

Included Lab Sample IDs: 2017776

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	94.3	97.1	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A86032

Included Lab Sample IDs: 2017776

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: SM 5310 B-00
Run ID: A86036
Included Lab Sample IDs: 2017776

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.6	97.8	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A86049
Included Lab Sample IDs: 2017775

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	101	102	P/P	90 - 110
Sulfate	101	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A86052
Included Lab Sample IDs: 2017776

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	99.1	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A86067
Included Lab Sample IDs: 2017776

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	101	P/P	90 - 110

Reference Method: EPA 8321B
Run ID: A86112
Included Lab Sample IDs: 2017805

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
2,4-D	100	106	P/P	50 - 150
Bentazon	139	150	P/P	50 - 160
MCP	111	112	P/P	50 - 150
Triclopyr	93.2	97.8	P/P	50 - 160

Reference Method: EPA 8321B
Run ID: A86195
Included Lab Sample IDs: 2017805

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Acetaminophen	117	106	P/P	60 - 150
Carbamazepine	99.5	100	P/P	60 - 150
Diuron	120	119	P/P	60 - 150
Fenuron	109	108	P/P	60 - 150
Fluridone	110	111	P/P	60 - 160
Hydrocodone	113	110	P/P	60 - 150
Imazapyr	99.8	95.9	P/P	50 - 150
Imidacloprid	111	109	P/P	60 - 150
Linuron	107	104	P/P	60 - 150
Primidone	112	100	P/P	60 - 150
Pyraclostrobin	94.5	96.6	P/P	60 - 150

Quality Assurance Report

Calibration Verification

* Pass/Fail determinations are made for each bracketing calibration verification check.
Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.
Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery		MS % Recovery		LCS	Precision	MS
							SMP	
EPA 180.1 Rev. 2.0	Turbidity						3.47	
EPA 300.0 Rev. 2.1	Chloride	102		93.0	94.4		1.91	
	Sulfate	101		95.7	97.1		1.47	
EPA 350.1 Rev. 2.0	Ammonia-N	99.1		98.5	101			2.14
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	103		104	105			0.780
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	101		100	100			0.0
EPA 365.1 Rev. 2.0	Total-P	97.9		96.8	94.8			1.91
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	100		99.0	99.2			0.202
EPA 8321B	2,4-D	101	75.1			29.6		
	Acetaminophen	131	113			15.4		
	Bentazon	110	99.5			10.3		
	Carbamazepine	92.3	90.0			2.51		
	Diuron	91.6	95.4			4.14		
	Fenuron	117	107			8.57		
	Fluridone	79.5	76.8			3.51		
	Hydrocodone	105	101			3.72		
	Imazapyr	102	98.2			4.06		
	Imidacloprid	113	106			6.58		
	Linuron	76.1	74.7			1.81		
	MCP	109	72.1			40.5		
	Primidone	109	108			1.36		
	Pyraclostrobin	78.9	74.0			6.43		
	Sucralose	107		109	105			3.80
	Triclopyr	100	64.7			42.9		
SM 2120 B	Color (true)	96.1					3.41	
SM 2320 B-97	Total Alkalinity	103					0.0281	
SM 2540 C-97	TDS						1.44	
SM 2540 D-97	TSS						2.15	
SM 4500 F-C-97	Fluoride	97.7		99.2	97.9			0.986
SM 5310 B-00	Organic Carbon	98.1		108				0.448

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2017775
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2017775
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2017775
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2017776
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2017776
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2017776
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2017776
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2017777
EPA 8321B / E31780	Pesticides, herbicides, and other chemicals in water matrices by HPLC/MS/MS	2017805
SM 2120 B / E31780	True Color measured at 450 nm	2017775
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2017775
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2017775
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2017775

Reference Method Descriptions

Method / DoH Cert #	Description	<u>Associated Samples</u>
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2017775
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2017776

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	08/16/2018			08/16/2018 15:39	William L. Brown IV	2017775
EPA 300.0 Rev. 2.1	08/16/2018			08/22/2018 19:44	Lipi Saha	2017775
EPA 350.1 Rev. 2.0	08/16/2018			08/22/2018 13:23	Ping Hua	2017776
EPA 351.2 Rev. 2.0	08/16/2018	08/20/2018 11:00	Adrian Shelby	08/21/2018 11:52	Joseph Suarez	2017776
EPA 353.2 Rev. 2.0	08/16/2018			08/21/2018 13:20	Lauren Bottorf	2017776
EPA 365.1 Rev. 2.0	08/16/2018	08/21/2018 13:00	Sima Feizi	08/22/2018 09:48	Beverly Y. Sanders	2017776
EPA 365.1 Rev. 2.0 dissolved	08/16/2018			08/16/2018 15:54	Julia Storbeck	2017777
EPA 8321B	08/16/2018	08/17/2018 09:00	Mohammad Ghaffari	08/18/2018 05:45	Mohammad Ghaffari	2017805
	08/16/2018	08/21/2018 15:00	Hoor Shaik	08/25/2018 07:42	Juyoung Kim	2017805
	08/16/2018	08/21/2018 15:00	Hoor Shaik	08/25/2018 17:58	Juyoung Kim	2017805
SM 2120 B	08/16/2018			08/16/2018 15:37	Tanya Welborn	2017775
SM 2320 B-97	08/16/2018			08/18/2018 05:23	Dale L. Simmons	2017775
SM 2540 C-97	08/16/2018			08/17/2018 16:00	William L. Brown IV	2017775
SM 2540 D-97	08/16/2018			08/17/2018 17:25	William L. Brown IV	2017775
SM 4500 F-C-97	08/16/2018			08/18/2018 05:23	Dale L. Simmons	2017775
SM 5310 B-00	08/16/2018			08/20/2018 23:51	Megan Treadwell	2017776