

Chemical Analysis Report

TLH-ROC-2018-09-24-01

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **G2 Run 5**
Request ID: **RQ-2018-09-10-65**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Colin Wright, Program Administrator

Date Certified: 19-OCT-2018 12:28

A handwritten signature in black ink, appearing to read "Colin Wright", is shown on a light-colored background.

Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: UNNAMED BRANCH @ HARDAWAY RD 2

Collection Date/Time: 09/24/2018 12:00

Field ID: G2TLH0041

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2027898	EPA 180.1 Rev. 2.0	Turbidity	4.1	A	NTU	P352234	
	EPA 300.0 Rev. 2.1	Chloride	3.8		mg Cl/L	P352521	MS
		Sulfate	0.65		mg SO4/L	P352523	
	SM 2120 B	Color (true)	32		PCU	P352238	
	SM 2320 B-97	Total Alkalinity	3.0		mg CaCO3/L	P352643	
	SM 2540 C-97	TDS	28		mg/L	P352446	
	SM 2540 D-97	TSS	3	I	mg/L	P352517	
	SM 4500 F-C-97	Fluoride	0.057	I	mg F/L	P352647	
2027899	EPA 350.1 Rev. 2.0	Ammonia-N	0.015		mg N/L	P352622	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.36		mg N/L	P352397	
	EPA 353.2 Rev. 2.0	NO2NO3-N	0.21		mg N/L	P352266	
	EPA 365.1 Rev. 2.0	Total-P	0.12		mg P/L	P352323	
	SM 5310 B-00	Organic Carbon	3.4		mg C/L	P352349	
2027900	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.059		mg P/L	P352222	

Ref. Method and Comment:

SM 5310 B-00: Batch matrix spike recoveries are unavailable because of high analyte concentration in the QC sample.

Quality Assurance Report

Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P352234

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P352521

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P352523

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P352622

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P352397

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P352266

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P352323

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352222

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P352238

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P352643

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P352446

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P352517

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P352647

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P352349

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P352521

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	97.7		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
 Batch ID: P352523

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	97.8		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
 Batch ID: P352622

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	101		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
 Batch ID: P352397

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	104		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
 Batch ID: P352266

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	101		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
 Batch ID: P352323

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	96.8		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
 Batch ID: P352222

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	97.4		P	90 - 110

Reference Method: SM 2120 B
 Batch ID: P352238

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	94.8		P	85 - 115

Reference Method: SM 2320 B-97
 Batch ID: P352643

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	104		P	98 - 107

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P352647

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	98.6		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P352349

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	94.5		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P352521

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2027629	Chloride	90.4		P	90 - 110
2027725	Chloride	88.0		F	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P352523

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2027629	Sulfate	93.5		P	90 - 110
2027725	Sulfate	92.9		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P352622

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2027684	Ammonia-N	99.1	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P352397

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2027899	Kjeldahl Nitrogen	97.9	102	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P352266

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2027678	NO2NO3-N	103	102	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P352323

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2027639	Total-P	94.7	93.7	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352222

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2027988	O-Phosphate-P	94.4	95.4	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P352647

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2027635	Fluoride	97.6	98.2	P/P	91 - 105

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P352234

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027898	Turbidity	3.92	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P352521

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027629	Chloride	0.449	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P352523

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027629	Sulfate	0.953	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P352622

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027684	Ammonia-N	1.12	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P352397

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027899	Kjeldahl Nitrogen	2.77	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P352266

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027678	NO2NO3-N	0.487	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P352323

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027639	Total-P	1.01	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P352222

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027988	O-Phosphate-P	1.05	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
Batch ID: P352238

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027917	Color (true)	2.66	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P352643

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027635	Total Alkalinity	0.752	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P352446

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027725	TDS	8.19	Sample	P	0 - 10

Reference Method: SM 2540 D-97
Batch ID: P352517

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027725	TSS	5.71	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P352647

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027635	Fluoride	0.527	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P352349

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2027590	Organic Carbon	0.320	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: EPA 365.1 Rev. 2.0 dissolved

Run ID: A86711

Included Lab Sample IDs: 2027900

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	99.4	98.2	P/P	90 - 110

Reference Method: EPA 180.1 Rev. 2.0

Run ID: A86716

Included Lab Sample IDs: 2027898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	104	P/P	95 - 105

Reference Method: SM 2120 B

Run ID: A86717

Included Lab Sample IDs: 2027898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	97.7	98.4	P/P	90 - 115

Reference Method: EPA 300.0 Rev. 2.1

Run ID: A86718

Included Lab Sample IDs: 2027898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	98.7	99.1	P/P	90 - 110
Sulfate	98.2	98.9	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A86724

Included Lab Sample IDs: 2027899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	99.5	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0

Run ID: A86739

Included Lab Sample IDs: 2027899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	101	103	P/P	90 - 110

Reference Method: SM 5310 B-00

Run ID: A86745

Included Lab Sample IDs: 2027899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	93.8	95.2	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0

Run ID: A86769

Included Lab Sample IDs: 2027899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	107	103	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 350.1 Rev. 2.0

Run ID: A86854

Included Lab Sample IDs: 2027899

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	102	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A86860

Included Lab Sample IDs: 2027898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	104	99.8	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A86860

Included Lab Sample IDs: 2027898

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	99.0	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					3.92	
EPA 300.0 Rev. 2.1	Chloride	97.7	90.4	88.0		0.449	
	Sulfate	97.8	93.5	92.9		0.953	
EPA 350.1 Rev. 2.0	Ammonia-N	101	99.1	101			1.12
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	104	97.9	102			2.77
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	101	103	102			0.487
EPA 365.1 Rev. 2.0	Total-P	96.8	94.7	93.7			1.01
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	97.4	94.4	95.4			1.05
SM 2120 B	Color (true)	94.8				2.66	
SM 2320 B-97	Total Alkalinity	104				0.752	
SM 2540 C-97	TDS					8.19	
SM 2540 D-97	TSS					5.71	
SM 4500 F-C-97	Fluoride	98.6	97.6	98.2			0.527
SM 5310 B-00	Organic Carbon	94.5					0.320

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2027898
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2027898
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2027898
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2027899
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2027899
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2027899
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2027899
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2027900
SM 2120 B / E31780	True Color measured at 450 nm	2027898
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2027898
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2027898
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2027898
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2027898
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2027899

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	09/24/2018			09/25/2018 16:42	Megan Treadwell	2027898
EPA 300.0 Rev. 2.1	09/24/2018			09/29/2018 00:31	Lipi Saha	2027898
EPA 350.1 Rev. 2.0	09/24/2018			10/01/2018 11:34	Ping Hua	2027899
EPA 351.2 Rev. 2.0	09/24/2018	09/27/2018 15:15	Adrian Shelby	09/28/2018 10:44	Joseph Suarez	2027899
EPA 353.2 Rev. 2.0	09/24/2018			09/26/2018 09:24	Beverly Y. Sanders	2027899
EPA 365.1 Rev. 2.0	09/24/2018	09/26/2018 11:30	Adrian Shelby	09/27/2018 09:29	Beverly Y. Sanders	2027899
EPA 365.1 Rev. 2.0 dissolved	09/24/2018			09/25/2018 16:40	Julia Storbeck	2027900
SM 2120 B	09/24/2018			09/25/2018 16:03	DeAsia Armster	2027898
SM 2320 B-97	09/24/2018			10/02/2018 10:26	Dale L. Simmons	2027898
SM 2540 C-97	09/24/2018			09/25/2018 15:05	Yijie Li	2027898
SM 2540 D-97	09/24/2018			09/25/2018 15:05	Yijie Li	2027898
SM 4500 F-C-97	09/24/2018			10/02/2018 10:26	Dale L. Simmons	2027898
SM 5310 B-00	09/24/2018			09/27/2018 03:08	Megan Treadwell	2027899