

Chemical Analysis Report

TLH-ROC-2018-11-20-04

Florida Department of Environmental Protection
Central Laboratory
2600 Blair Stone Road
Tallahassee, FL 32399-2400
DOH Accreditation E31780

Event Description: **G1 Run 4**
Request ID: **RQ-2018-10-08-17**
Customer: **TLH-ROC**
Project ID: **SMP**

Send Reports to:
FL Dept. of Environmental Protection
2600 Blair Stone Rd
Tallahassee, FL 32399
Attn: Michael Eckles

For additional information please contact
Colin Wright, Ph.D.
Liang-Tsair Lin, Ph.D.
Kerry Tate, Ph.D.
Dr. rer. nat. Bettina Steinbock
Thekkekalathil Chandrasekhar, Ph.D, QA Officer
Phone (850) 245-8085

Certified by: Dr. rer. nat. Bettina Steinbock, Environmental Administrator

Date Certified: 11-DEC-2018 16:38



Case Narrative

Unless otherwise noted, all samples included in this report were received in accordance with protocols referenced in Chapter 62-160, Florida Administrative Code (F.A.C.). Results published in this report pertain only to the samples as submitted to, and received by the laboratory. All times in this report are adjusted to the applicable Eastern Time Zone (EST or EDT).

Results for the following analytical group are included in this report: Nutrients.

Scientific notation may be used in reporting very large or small values. Values reported using scientific notation will take the form of the following example: 1.3E+03, which is equivalent to 1.3×10^3 or 1300.

Unless otherwise noted, analytical values for soil and sediment samples are reported on a dry weight basis, and analytical values for waste and tissue samples are reported on a wet weight basis.

Results for TNI accredited tests met requirements established by The NELAC Institute. A double asterisk (**) is used to indicate an analyte/matrix/method for which the laboratory is not TNI accredited or where accreditation for that field of testing is not applicable.

Any significant anomalies or deviations from established protocols are documented in Non-Conformance Reports, which, where appropriate, are included within this analytical report. Additional comments related to specific analytical tests may be included as remarks following the analytical results for each sample. Such comments and remarks are for informational purposes only and are not intended to convey judgement about the usability of the reported data.

A quality control report on the performance of the test method for the submitted samples is included. Uncertainty associated with the analytical results contained in this report can be estimated from the reported quality assurance results and from published quality control acceptance limits for each analytical test. Matrix quality control results (matrix spike recoveries and matrix sample precision) pertain only to the matrix sample tested and do not necessarily reflect test method performance for other samples.

Typical matrix quality control (QC) measurements may include matrix spike recovery, matrix spike duplicate recovery, matrix spike precision and matrix sample precision. Not all matrix QC results may be available or reportable; where they are not an explanation is provided. Typical reasons for unavailable QC results include, but are not limited to, a) insufficient matrix sample to perform some or all QC measurements; b) analyte concentration in the sample replicated was too low for a meaningful measurement of precision and c) analyte concentration in the matrix sample spiked was too high (relative to the amount of analyte spiked) for a meaningful measurement of recovery. Where matrix QC results are unavailable, other method performance metrics (e.g., LCS recovery, LCS precision, surrogate recovery) may be used to assess performance of the method. Comments explaining any missing QC measurements are not intended to convey any adverse conclusions about the quality of the reported data.

Precision is reported as relative percent difference unless otherwise noted.

Quality Control codes as defined below may be used in this report to indicate results that are associated with one or more quality control elements which did not fall within established test method criteria. Such results may be qualified as estimates using a J qualifier as required by 62-160 F.A.C. Explanations are included in the report for any results that were reported as estimates for other reasons.

QC Codes used in this report may include:

- LCS – Recovery for the batch Laboratory Control Sample (LCS) was outside existing control limits;
- MS – Recovery for the batch matrix spike (MS) was outside existing control limits;
- CCV – Recovery for a continuing calibration verification (CCV) standard was outside existing control limits;
- SUR – Recovery of a surrogate (SUR) for associated analytes was outside existing control limits;
- RPD – The precision, measured as relative percent difference (RPD), of batch replicate measurements was outside existing control limits;
- RSD – The precision, measured as relative standard deviation (RSD), of batch replicate measurements was outside existing control limits;
- SMP – Sample - used precision derived from replicate analyses of a sample;

The following data qualifiers are used, where applicable, in this report as specified in 62-160 F.A.C.

- A - Value reported is the mean of two or more determinations.
- B - Results based on colony counts outside the acceptable range.
- I - The reported value is between the laboratory method detection limit and the laboratory practical quantitation limit.
- J - Estimated value and/or the analysis did not meet established quality control criteria.
- K - Actual value is known to be less than value given.
- L - Actual value is known to be greater than value given.
- N - Presumptive evidence of presence of material.
- O - Sampled, but analysis lost or not performed.
- Q - Sample held beyond normal holding time.
- T - Value reported is less than the criterion of detection.
- U - Material was analyzed for but not detected. The reported value is the method detection limit for the sample analyzed.
- V - Analyte was detected in both sample and method blank.
- X - Too few individuals to calculate SCI value.
- Y - The laboratory analysis was from an unpreserved or improperly preserved sample. The data may not be accurate.
- Z - Colonies were too numerous to count (TNTC).

Quality control information from overflow laboratories may not be included in this report. Please refer to the associated report from the overflow laboratory for additional information.

Sample Location: Unnamed Branch @ Power Line Bridge

Collection Date/Time: 11/20/2018 13:55

Field ID: G2TLHR0040

Matrix: W-SURF-FRH

Sample ID	Ref. Method	Component	Result	Code	Units	Batch ID	QC Codes
2043243	EPA 180.1 Rev. 2.0	Turbidity	2.9		NTU	P355039	
	EPA 300.0 Rev. 2.1	Chloride	4.2		mg Cl/L	P355365	
		Sulfate	0.59	I	mg SO4/L	P355368	
	SM 2120 B	Color (true)	17		PCU	P355036	
	SM 2320 B-97	Total Alkalinity	1.9	I	mg CaCO3/L	P355468	
	SM 2540 C-97	TDS	37		mg/L	P355460	
	SM 2540 D-97	TSS	4	I	mg/L	P355462	
	SM 4500 F-C-97	Fluoride	0.071	I	mg F/L	P355473	
2043244	EPA 350.1 Rev. 2.0	Ammonia-N	0.034		mg N/L	P355303	
	EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	0.29		mg N/L	P355127	
	EPA 353.2 Rev. 2.0	NO2NO3-N	2.0		mg N/L	P355328	
	EPA 365.1 Rev. 2.0	Total-P	0.098		mg P/L	P355083	
	SM 5310 B-00	Organic Carbon	2.0		mg C/L	P355325	
2043245	EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	0.058		mg P/L	P355042	

Ref. Method and Comment:

SM 2540 D-97: Batch relative percent difference is unavailable because of low analyte concentration in the QC sample.

Quality Assurance Report Method Blank Results

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P355039

Component	Result	Code	Units
Turbidity	0.10	U	NTU

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355365

Component	Result	Code	Units
Chloride	0.20	U	mg Cl/L

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355368

Component	Result	Code	Units
Sulfate	0.20	U	mg SO4/L

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P355303

Component	Result	Code	Units
Ammonia-N	0.002	U	mg N/L

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P355127

Component	Result	Code	Units
Kjeldahl Nitrogen	0.080	U	mg N/L

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P355328

Component	Result	Code	Units
NO2NO3-N	0.004	U	mg N/L

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P355083

Component	Result	Code	Units
Total-P	0.002	U	mg P/L

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P355042

Component	Result	Code	Units
O-Phosphate-P	0.004	U	mg P/L

Reference Method: SM 2120 B
Batch ID: P355036

Component	Result	Code	Units
Color (true)	2.5	U	PCU

Quality Assurance Report

Method Blank Results

Reference Method: SM 2320 B-97
Batch ID: P355468

Component	Result	Code	Units
Total Alkalinity	0.65	U	mg CaCO ₃ /L

Reference Method: SM 2540 C-97
Batch ID: P355460

Component	Result	Code	Units
TDS	15	U	mg/L

Reference Method: SM 2540 D-97
Batch ID: P355462

Component	Result	Code	Units
TSS	2	U	mg/L

Reference Method: SM 4500 F-C-97
Batch ID: P355473

Component	Result	Code	Units
Fluoride	0.035	U	mg F/L

Reference Method: SM 5310 B-00
Batch ID: P355325

Component	Result	Code	Units
Organic Carbon	0.50	U	mg C/L

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355365

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Chloride	100		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355368

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Sulfate	100		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P355303

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Ammonia-N	96.0		P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P355127

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Kjeldahl Nitrogen	100		P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P355328

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
NO2NO3-N	99.0		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P355083

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total-P	103		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P355042

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
O-Phosphate-P	98.6		P	90 - 110

Reference Method: SM 2120 B
Batch ID: P355036

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Color (true)	98.1		P	85 - 115

Reference Method: SM 2320 B-97
Batch ID: P355468

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Total Alkalinity	99.9		P	98 - 107

Quality Assurance Report

Laboratory Control Sample Accuracy

Reference Method: SM 4500 F-C-97
Batch ID: P355473

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Fluoride	97.9		P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P355325

Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
Organic Carbon	94.8		P	85 - 115

Quality Assurance Report Matrix Spike Accuracy

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355365

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2040903	Chloride	100		P	90 - 110
2043165	Chloride	99.0		P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355368

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2040903	Sulfate	100		P	90 - 110
2043165	Sulfate	99.3		P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P355303

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2043244	Ammonia-N	93.1	101	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P355127

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2043244	Kjeldahl Nitrogen	90.2	100	P/P	90 - 110

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P355328

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2043238	NO2NO3-N	99.2		P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P355083

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2043197	Total-P	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P355042

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2043245	O-Phosphate-P	97.6	99.3	P/P	90 - 110

Reference Method: SM 4500 F-C-97
Batch ID: P355473

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2043270	Fluoride	98.6	98.0	P/P	91 - 105

Reference Method: SM 5310 B-00
Batch ID: P355325

Spiked Sample	Component	% Rec.1	% Rec.2	Pass/Fail	Control Limits
2043003	Organic Carbon	98.6	95.6	P/P	85 - 115

Quality Assurance Report Precision

Reference Method: EPA 180.1 Rev. 2.0
Batch ID: P355039

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043280	Turbidity	2.32	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355365

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043165	Chloride	0.111	Sample	P	0 - 20

Reference Method: EPA 300.0 Rev. 2.1
Batch ID: P355368

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043165	Sulfate	1.00	Sample	P	0 - 20

Reference Method: EPA 350.1 Rev. 2.0
Batch ID: P355303

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043244	Ammonia-N	4.95	Spike	P	0 - 20

Reference Method: EPA 351.2 Rev. 2.0
Batch ID: P355127

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043244	Kjeldahl Nitrogen	8.51	Spike	P	0 - 20

Reference Method: EPA 353.2 Rev. 2.0
Batch ID: P355328

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043238	NO2NO3-N	0.564	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0
Batch ID: P355083

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043197	Total-P	0.0243	Spike	P	0 - 20

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Batch ID: P355042

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043245	O-Phosphate-P	1.09	Spike	P	0 - 20

Quality Assurance Report Precision

Reference Method: SM 2120 B
Batch ID: P355036

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043270	Color (true)	1.49	Sample	P	0 - 20

Reference Method: SM 2320 B-97
Batch ID: P355468

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043270	Total Alkalinity	1.03	Sample	P	0 - 1.6

Reference Method: SM 2540 C-97
Batch ID: P355460

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043165	TDS	3.22	Sample	P	0 - 10

Reference Method: SM 4500 F-C-97
Batch ID: P355473

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043270	Fluoride	0.530	Spike	P	0 - 3.5

Reference Method: SM 5310 B-00
Batch ID: P355325

Replicated Lab Sample	Component	% RSD/RPD	Sample/Spike/LCS*	Pass/Fail	Control Limits
2043003	Organic Carbon	2.03	Spike	P	0 - 20

* Sample, spike and/or laboratory control sample precision (LCS) is reported.
Replicate spike precision may be reported when sample results are below quantifiable levels.

Quality Assurance Report Calibration Verification

Reference Method: SM 2120 B
Run ID: A87929
Included Lab Sample IDs: 2043243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Color (true)	98.8	98.6	P/P	90 - 115

Reference Method: EPA 180.1 Rev. 2.0
Run ID: A87930
Included Lab Sample IDs: 2043243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Turbidity	100	100	P/P	95 - 105

Reference Method: EPA 365.1 Rev. 2.0 dissolved
Run ID: A87931
Included Lab Sample IDs: 2043245

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
O-Phosphate-P	100	100	P/P	90 - 110

Reference Method: EPA 365.1 Rev. 2.0
Run ID: A87969
Included Lab Sample IDs: 2043244

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total-P	100	99.6	P/P	90 - 110

Reference Method: EPA 300.0 Rev. 2.1
Run ID: A87993
Included Lab Sample IDs: 2043243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Chloride	102	102	P/P	90 - 110
Sulfate	102	102	P/P	90 - 110

Reference Method: EPA 351.2 Rev. 2.0
Run ID: A88009
Included Lab Sample IDs: 2043244

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Kjeldahl Nitrogen	106	106	P/P	90 - 110

Reference Method: EPA 350.1 Rev. 2.0
Run ID: A88037
Included Lab Sample IDs: 2043244

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Ammonia-N	101	96.0	P/P	90 - 110

Reference Method: SM 5310 B-00
Run ID: A88042
Included Lab Sample IDs: 2043244

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Organic Carbon	96.6	95.1	P/P	90 - 110

Quality Assurance Report Calibration Verification

Reference Method: EPA 353.2 Rev. 2.0

Run ID: A88043

Included Lab Sample IDs: 2043244

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
NO2NO3-N	100	100	P/P	90 - 110

Reference Method: SM 2320 B-97

Run ID: A88097

Included Lab Sample IDs: 2043243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Total Alkalinity	100	98.5	P/P	90 - 110

Reference Method: SM 4500 F-C-97

Run ID: A88097

Included Lab Sample IDs: 2043243

Component	% Rec.1	% Rec.2	Pass/Fail*	Control Limits
Fluoride	100	99.5	P/P	90 - 110

* Pass/Fail determinations are made for each bracketing calibration verification check.

Control limits for initial calibration checks may be different from those for continuing checks, depending on method requirements.

Where they are different, both control limits are provided.

Quality Assurance Report Summary

Ref. Method	Analyte	LCS % Recovery	MS % Recovery		LCS	Precision	MS
						SMP	
EPA 180.1 Rev. 2.0	Turbidity					2.32	
EPA 300.0 Rev. 2.1	Chloride	100	99.0	100		0.111	
	Sulfate	100	99.3	100		1.00	
EPA 350.1 Rev. 2.0	Ammonia-N	96.0	93.1	101			4.95
EPA 351.2 Rev. 2.0	Kjeldahl Nitrogen	100	90.2	100			8.51
EPA 353.2 Rev. 2.0	NO ₂ NO ₃ -N	99.0	99.2				0.564
EPA 365.1 Rev. 2.0	Total-P	103	100	100			0.0243
EPA 365.1 Rev. 2.0 dissolved	O-Phosphate-P	98.6	97.6	99.3			1.09
SM 2120 B	Color (true)	98.1				1.49	
SM 2320 B-97	Total Alkalinity	99.9				1.03	
SM 2540 C-97	TDS					3.22	
SM 4500 F-C-97	Fluoride	97.9	98.6	98.0			0.530
SM 5310 B-00	Organic Carbon	94.8	98.6	95.6			2.03

Reference Method Descriptions

Method / DoH Cert #	Description	Associated Samples
EPA 180.1 Rev. 2.0 / E31780	Turbidity in aqueous matrices	2043243
EPA 300.0 Rev. 2.1 / E31780	Chloride in aqueous matrices	2043243
EPA 300.0 Rev. 2.1 / E31780	Sulfate in aqueous matrices	2043243
EPA 350.1 Rev. 2.0 / E31780	Ammonia in aqueous matrices as mg N/L	2043244
EPA 351.2 Rev. 2.0 / E31780	Total Kjeldahl Nitrogen in aqueous matrices	2043244
EPA 353.2 Rev. 2.0 / E31780	Nitrite/Nitrate in aqueous matrices as mg N/L	2043244
EPA 365.1 Rev. 2.0 / E31780	Total Phosphorus in aqueous matrices as mg P/L	2043244
EPA 365.1 Rev. 2.0 dissolved / E31780	Ortho-phosphate, dissolved, in filtered, aqueous matrices as mg P/L	2043245
SM 2120 B / E31780	True Color measured at 450 nm	2043243
SM 2320 B-97 / E31780	Alkalinity in aqueous matrices as mg CaCO ₃ /L	2043243
SM 2540 C-97 / E31780	Total Dissolved Solids in aqueous matrices	2043243
SM 2540 D-97 / E31780	Total Suspended Solids in aqueous matrices using a maximum of 250 mL of sample	2043243
SM 4500 F-C-97 / E31780	Fluoride in aqueous matrices, without distillation (not valid for NPDES monitoring)	2043243
SM 5310 B-00 / E31780	Nonpurgeable Organic Carbon in aqueous matrices	2043244

Preparation and Analysis Log

Ref. Method	Received Date	Prep Date/Time	Prepared By	Analysis Date/Time	Analyzed By	Associated Samples
EPA 180.1 Rev. 2.0	11/20/2018			11/21/2018 15:37	Megan Treadwell	2043243
EPA 300.0 Rev. 2.1	11/20/2018			11/30/2018 06:35	Lipi Saha	2043243
EPA 350.1 Rev. 2.0	11/20/2018			11/29/2018 10:18	Ping Hua	2043244
EPA 351.2 Rev. 2.0	11/20/2018	11/27/2018 12:30	Adrian Shelby	11/28/2018 14:32	Joseph Suarez	2043244
EPA 353.2 Rev. 2.0	11/20/2018			11/29/2018 11:42	Lauren Bottorf	2043244
EPA 365.1 Rev. 2.0	11/20/2018	11/26/2018 12:15	Sima Feizi	11/27/2018 11:15	Beverly Y. Sanders	2043244
EPA 365.1 Rev. 2.0 dissolved	11/20/2018			11/21/2018 15:12	Jhamieka S. Greenwood	2043245
SM 2120 B	11/20/2018			11/21/2018 15:58	Chelsea Hernandez	2043243
SM 2320 B-97	11/20/2018			11/30/2018 05:47	Dale L. Simmons	2043243
SM 2540 C-97	11/20/2018			11/26/2018 16:02	Yijie Li	2043243
SM 2540 D-97	11/20/2018			11/26/2018 16:02	Yijie Li	2043243
SM 4500 F-C-97	11/20/2018			11/30/2018 05:47	Dale L. Simmons	2043243
SM 5310 B-00	11/20/2018			11/28/2018 19:22	Megan Treadwell	2043244